

Name: National Institute of Measurement and Testing Technology

Address: No.10, Yushuang Road, Chengdu, Sichuan, China

Registration No. CNAS L0893

Accreditation Criteria: ISO/IEC 17025:2017 and relevant requirements of CNAS

Effective Date: 2023-01-18 Expiry Date: 2024-11-10

SCHEDULE 5 ACCREDITED CALIBRATION AND MEASUREMENT CAPABILITY SCOPE

Note: The instruments with * represents onsite calibration can be performed.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
I Geometry measuring instrument							
1	Steel tape	Length	Verification Regulation of Steel Measuring Tapes JJG 4	(0~100)m	$U=((0.02+0.02L)\text{mm (L:m)})$		
2	Class 3 line ruler of metal	Length	Verification Regulation of Standard Metallic Scale(Grade III) JJG 71	(0~1000)mm	$U=(5+5L)\text{ }\mu\text{ m(L:m)}$		
3	Standard steel tape	Length	Verification Regulation of Standard Steel Tapes JJG 741	(0~20)m	$U=(5+5L)\text{ }\mu\text{ m(L:m)}$		
4	*Toolmaker' s microscope	Length	Verification Regulation of Toolmaker's Microscope JJG 56	(1000×200)mm	$U=(0.3+L/300)\text{ }\mu\text{ m (L:mm)}$		
5	Reading Microscope	Length	Verification Regulation of Reading Microscope and Measuring Microscope JJG 571	(0~8)mm	$U=1\text{ }\mu\text{ m}$		



No. CNAS L0893

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
6	Measuring Microscope	Length	Verification Regulation of Reading Microscope and Measuring Microscope JJG 571	(0~50)mm	$U=(1+L/50) \mu m (L:mm)$		
7	Linear comparator	Length	Verification Regulation of Linear comparator JJG 72	(0~200)mm	$U=(0.2+2L) \mu m (L:mm)$		
8	Steel rulers	Length	Verification Regulation of Steel Rule JJG 1	(0~2000)mm	$U=(0.02+0.02L)mm (L:m)$		
9	Projector Ruler	Length	Projector Ruler NIMTT(CM) 024	(0~200)mm	$U=(0.5+2.5L) \mu m (L:mm)$		
10	Cable Length Meter	Length	Verification Regulation of Cable Length Meter JJG 987	(1~1000)m	$U_{rel}=0.05\%$		
11	*Projector	Length	Calibration Specification for Projectors JJF 1093	200×100mm	$U=(1+L/200) \mu m (L:mm)$		
12	633nm Frequency Stabilized Lasers	Wavelength	633nm Frequency Stabilized Lasers JJG 353	633nm	$U_{rel}=5 \times 10^{-10}$		
13	Micrometer for outside dimension	Length	Calibration Specification for Micrometers with Measuring Range from 500mm to 3000mm JJF 1088	(500~3000)mm	$U=(1+L/200) \mu m (L:mm)$		
14	Height Caliper	Length	Verification Regulation of Height Caliper JJG 31	(0~2000)mm	$U=(0.01+0.01L)mm (L:m)$		
15	Micrometer for inside dimension	Length	Calibration Specification for Micrometers for Measuring Inside Dimension JJF 1411	(5~200)mm	$U=(1+L/200) \mu m (L:mm)$		
16	Micrometer	Length	Verification Regulation of Micrometer JJG 21	(0~500)mm	$U=(0.5+L/200) \mu m (L:mm)$		
17	Current Calipers	Length	Verification Regulation of Current Calipers JJG 30	(0~2000) mm	$U=(0.01+0.01L)mm (L:m)$		



No. CNAS L0893

第 2 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
18	Instruction sheet	Length	Verification Regulation of Dial Gauges(dial and digital) JJG 34	(0~10)mm/Dividing : 0.001mm	$U=1.6 \mu m$		
				(0~50)mm reading 0.01mm	$U=4 \mu m$		
19	Dial indicator gauges	Length	Verification Regulation of Tester for Dial Indicator Gauges JJG 201	(0~50)mm	$U=(0.3+L/40) \mu m L:mm$		
20	Portable Vertical Deviation Measuring Instrument	Length	Verification Regulation of Vertical Metal Tank Diametrical Deviation Measuring Instrument JJG 988	(0~300)mm	$U=0.02mm$		
21	Wide Range Dauges Reading in 0.01mm	Length	Verification Regulation of Wide Range Dauges Reading in 0.01mm JJG 379	(0~100)mm	$U=(0.005+0.1L)mm L:m$		
22	Thickness Gauges	Length	Calibration Specification for Thickness Gauges JJF 1255				
23	Depth Micrometers	Length	Depth Micrometers JJG 24	(0~300)mm	$U=1.5 \mu m$		
24	Micrometers with Dial Comparator and Indication Snap Gauge	Length	Verification Regulation of Micrometers with Dial Comparator and Indication Snap Gauge JJG 26	Indication Snap Gauge(0~200)mm	$U=0.5 \mu m$		
				Micrometers with Dial Comparator(0~100)mm	$U=1 \mu m$		
25	Internal Micrometers	Length	Verification Regulation of Internal Micrometers JJG 22	(50~6000)mm	$U=(2+L/250) \mu m(L:mm)$		
26	Dial Test Indicator	Length	Verification Regulation of Dial Test Indicator JJG 35	(0~1)mm	$U=1 \mu m$		
27	Depth Dial Gauge	Length	Verification Regulation of Depth Dial Gauge JJG 830	(0~300)mm	$U=2 \mu m$		



No. CNAS L0893

第 3 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
28	Bore Dial Indicators	Length	Calibration Specification for Bore Dial Indicators JJF 1102	Bore dial indicators for 0.001(10~400)mm	$U=1.6 \mu m$		
				Bore dial indicators for 0.01(2~450)mm	$U=3 \mu m$		
29	Micro-alignment Telescopes	Length	Calibration Specification of Micro-alignment Telescopes JJF 1077	(-1.2~+1.2)mm	$U=3 \mu m$		
30	Gauge Blocks	central length	Gauge Blocks JJG 146	(0.5~100)mm	$U=0.02 \mu m+0.2 \times 10^{-6} L(k=2.7)$		
		central length		(>100~1000)mm	$U=0.05 \mu m+0.5 \times 10^{-6} L(k=2.7)$		
31	*Gear Degree dais	Angle	Verification Regulation of Precise Angle Dividing Table JJG 472	$0^{\circ} \sim 360^{\circ}$	$U=0.06''$		
32	*Measure angle instrument	Angle	Verification Regulation of Goniometers JJG 97	$0^{\circ} \sim 360^{\circ}$	$U=0.5''$		
33	*Small Angle Tester	Angle	Verification Regulation of Small Angle Testers JJG 300	$(0 \sim 40)'$	$U=0.3''$		
34	Angle gauge blocks	Angle	Verification Regulation of Angle Gauge Blocks JJG 70	$10^{\circ} \sim 100^{\circ}$	$U=1''$		
35	Angular polygon	Angle	Verification Regulation of Angular Polygon JJG 283	$0^{\circ} \sim 360^{\circ}$	$U=0.2''$		
36	Optical clinometer	Angle	Calibration Specification for Clinometers JJF 1915	$0^{\circ} \sim 240^{\circ}$	$U=6''$		
37	Electronic level	Angle	Verification Regulation of Electronic levels and Coincidence levels JJG 103	(-10mm/m~+10mm/m)	$U=0.001mm/m$		
38	Dividing head	Angle	Verification Regulation of Optical Digital Dividing Head JJG 57	$0^{\circ} \sim 360^{\circ}$	$U=1''$		



No. CNAS L0893

第 4 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
39	Optics-electricity angle coded senso	Angle	Calibration Specification for Photoelectric Shaft Encoders JJF 1115	0° ~ 360°	$U=0.25''$		
40	General bevel protractors	Angle	Calibration Specification for General Bevel Protractors JJF 1959	0° ~ 360°	$U=1'$		
41	Autocollimator	Angle	Verification Regulation of Autocollimator JJG 202	(0~10)'	$U=0.1''$		
42	Angle gauges	perpendicularity	Verification Regulation of Squares JJG 7	H:(63~1000)mm	$U=(1+H/500) \mu m$		
43	Coincidence level	Angle	Verification Regulation of Electronic Levels and Coincidence Levels JJG 103	(0~20)mm/m	$U=0.005mm/m$		
44	Frame levels and shaft levels	Angle	Calibration Specification for Frame Levels and Shaft Levels JJF 1084	(0.01~20)mm/m	$U_{rel}=6\%$		
45	Right-angle Edge	Straightness	Verification Regulation of Squares JJG 7	H:(50~200)mm	$U=1 \mu m$		
46	*Square Testers	perpendicularity	Calibration Specification for Square Testers JJF 1140	H:(63~1000)mm	$U=0.8 \mu m$		
47	*Optical Comparators for Angle Measurement	Angle	Calibration Specification for Optical Comparators for Angle Measurement JJF 1078	(0~60)'	$U=2''$		
48	*Calibrators for the Levels	Angle	Verification Regulation of Calibrators for the Levels JJG 191	(0.001~1.5)mm/m	$U_{rel}=2\%$		
49	Square Gauge	Perpendicularity	Verification Regulation of Square Gauge JJG 1046	(100~500)mm	$U=0.3 \mu m$		
50	*Optical rotary table	Angle	Calibration Specification for Optical & Digital Dividing Table JJF 1114	(0~360)°	$U=1''$		

No. CNAS L0893

第 5 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.




№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
51	Milling Straight Edges	Flatness	Verification Regulation of Milling Straight Edges JJG 740	(200~500)mm	$U=0.05 \mu\text{m}$		
52	Straight Edge	Straightness	Verification Regulation of Straight Edge JJG 63	75mm	$U=0.4 \mu\text{m}$		
				(> 75~225)mm	$U=0.7 \mu\text{m}$		
				(> 225~500)mm	$U=1.3 \mu\text{m}$		
53	Taper gauges	Angle	Taper gauges JJG 177	(0~45)°	$U=2''$		
		Straightness		L:(0~200)mm	$U=1 \mu\text{m}$		
54	Sine bars	Angle	Sine bars JJG 37	(0~30)°	$U=2''$		
		Flatness		(0~200)mm	$U=1 \mu\text{m}$		
55	Theodolites	Angle	Optical Theodolites JJG 414	Horizontal angle:(0~360)° Vertical angle: $\pm 31^\circ$	$U=0.2''$		
56	Level	Angle	Verification Regulation of Levels JJG 425	$\pm 25''$ (2m~∞)	$U=2.0''$		
57	Plumb Instruments	Angle	Calibration Specification for Plumb Instruments JJF 1081	$\pm 1'$	$U=1.2''$		
58	Electronic Theodolite	Angle	Verification Regulation of Electronic Tachometer Total Station JJG 100	Horizontal angle:(0~360)° Vertical angle: $\pm 31^\circ$	$U=0.2''$		
59	Height Measuring Instrument with Digital Display	Length	Calibration Specification for Height Measuring Instrument with Digital Display JJF 1254	(0~1000)mm	$U=(1+L/1000) \mu\text{m}(L:\text{mm})$		



No. CNAS L0893

第 6 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
60	*Theodolite Verification Device	Angle	Verification Regulation of Theodolite Verification Devices JJG 949	(0~360)°	$U=0.1''$		
61	Gear Helix Master	Length	Gear Helix Master JJG 408	rb (20~50) mm	$U=2.0 \mu m$		
				rb ($\geq 50 \sim 100$) mm	$U=2.3 \mu m$		
				rb ($\geq 100 \sim 150$) mm	$U=2.5 \mu m$		
				rb ($> 150 \sim 300$) mm	$U=2.8 \mu m$		
62	Screw Templates	Length	Screw Templates JJG 60	(0.4~6)mm	$U=(3+L/100) \mu m (L: mm)$		
63	Instruments of Thread Inspection of Casing, Tubing, Line Pipe and New Rotary Shouldered Connection	Length	Calibration Specification for Instruments of Thread Inspection of Casing, Tubing, Line Pipe and New Rotary Shouldered Connection JJF 1063	briquette (0.1~300) mm	$U=(1+L/100) \mu m (L: mm)$		
				dial (0~50) mm	$U=0.5 \mu m$		
64	Bearing Inside and Outside Diameter Testers	Length	Bearing Inside and Outside Diameter Testers JJG 471	(0.1~100)mm	$U=0.7 \mu m$		
65	*The universer involute gear teste	Length	Calibration Specification for Gear involute Measuring instrument JJF 1124	rb(0.05~3)m	$U=2 \mu m$		
66	*Lead Tester of gear	Length	Calibration Specification for Gear Helix Measuring Instruments JJF 1122	rb(0.05~3)m	$U=2 \mu m$		
67	*Gear Pitch Measuring Instruments	Length	Calibration Specification for Gear Pitch Measuring Instruments JJF 1209	(0~100)mm	$U=1 \mu m$		



No. CNAS L0893

第 7 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
68	Micrometers with Prismatically Arranged Measuring Faces	Length	Verification Regulation of Micrometers with Prismatically Arranged Measuring Faces JJG 182	(1~100)mm	$U=(1+L/100) \mu m$ (L: mm)		
69	*Measuring Instrument for Axial Clearance of Ball Bearings	Length	Verification Regulation of Measuring Instrument for Axial Clearance of Ball Bearings JJG 626	(9~250)mm	$U=2 \mu m$		
70	Common Normal Micrometer	Length	Verification Regulation of Common Normal Micrometer JJG 82	(0~150)mm	$U=2 \mu m$		
71	Screw Thread Micrometers	Length	Verification Regulation of Screw Thread Micrometers JJG 25	(0~200)mm	$U=2 \mu m$		
72	*Concentricity Tester	Length	Calibration Specification for Concentricity Tester JJF 1109	L: (50~1000) mm	$U=2 \mu m$		
73	*Base Circle Pitch Comparator	Length	Calibration Specification for Base circle pitch comparator JJF 1123	gear modulus m(2~20)	$U=1 \mu m$		
74	*OCTG Thread Working Gauges	pitch diameter	C.S for OCTG Thread Working Gauges JJF 1108, Threading, Gauging, and Inspection of Casing, Tubing, and Line Pipe Threads API API	(0~500) mm	$U= (3.3+L/400) \mu m$ (L:mm)		
		taper		(0~5) mm	$U=1 \mu m$		
		pitch		(0.5~10) mm	$U= (3.3+L/400) \mu m$ (L:mm)		
		Standoff		(5~20) mm	$U=8 \mu m$		
		Angle		(0~45)°	$U=3'$		
75	Multiple score Specimen	Roughness	C.S. for multiple score Specimen NIMTT(CM) 002	(0.01~100) μm	$U_{rel}=5\%$		



No. CNAS L0893

第 8 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
76	Roughness Comparison Specimens	Roughness	C.S.for Roughness Comparison Specimens JJF 1099	Ra:(0.012~25) μm	$U_{\text{rel}}=6\%$		
77	Smooth limit gauge	Length	V.R.of Smooth limit gauge JJG 343	Outside Size:(0.9~500)mm Inside Size:(20~500)mm	$U=0.35 \mu\text{m}+3 \times 10^{-6}L$ $U=1.0 \mu\text{m}+3 \times 10^{-6}L$		
78	Cylindrical Measuring Pin	Length	C.S for Cylindrical Measuring Pin JJF 1207	Pin gauge:(0.1~25)mm Three stitches:(0.118~6.585)mm	$U=0.4 \mu\text{m}$ $U=0.4 \mu\text{m}$		
79	Gear InvoluteMasters	Length	Verification Regulation of Gear Involute Masters JJG 332	rb (20~50) mm rb (>50~100) mm rb (>100~150) mm rb (>150~300) mm	$U=2.0 \mu\text{m}$ $U=2.3 \mu\text{m}$ $U=2.5 \mu\text{m}$ $U=2.8 \mu\text{m}$		
80	Step gauges	Length	C. S. for Step Gauges JJF 1258	(10~1000)mm	$U=0.4 \mu\text{m}+1.2 \times 10^{-6}L$		
81	Cylindrical Thread Gauges	Length	C.S for Cylindrical Thread Gauges JJF 1345	Thread Ring Gauge:(1.2~500)mm Thread Plug Gauge:(1~500)mm	$U=2.8 \mu\text{m}+2 \times 10^{-6}L$ $U=2.8 \mu\text{m}+2 \times 10^{-6}L$		
		Angle		Dental Angle:(29~60)°	$U=3'$		
82	Test sieves	Length	Calibration Specification for Test Sieves JJF 1175	(0.02~4)mm (>4~125)mm	$U=2 \mu\text{m}$ $U=0.03\text{mm}$		
83	Spline Gauges	Length	Calibration Specification for Straight Cylindrical Involute Spline Gauges JJF 1557	gear modulus m (0.3~1.5)	$U=2.1 \mu\text{m}$		



No. CNAS L0893

第 9 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				gear modulus m ($> 1.5 \sim 3.0$)	$U=2.4 \mu m$		
				gear modulus m ($> 3.0 \sim 5.0$)	$U=2.6 \mu m$		
				gear modulus m ($> 5.0 \sim 8.0$)	$U=2.8 \mu m$		
84	*System of coordinate Teste	Length	Calibration Specification for Coordinate Measuring Machine JJF 1064	($0 \sim 6$)m	$U=(0.2+L/1000) \mu m (L:mm)$		
				$6 \sim 20$)m	$U=(1+0.2L) \mu m (L:m)$		
85	*Flatness Interferometer with Parallel Light Brightening	Plane degree	Calibration Specification for Flat Equal Thickness Interferometers JJF 1100	$D:150 mm$	$U=0.01 \mu m$		
86	Optical flat	Flatness	V.R.of Optical flat JJG 28	Plane Optical flats: $D(30 \sim 200)mm$	$U=0.007 \mu m$		
				Plane Optical flats: $H(15 \sim 91) mm$	$U=0.014 \mu m$		
				Long Optical flats:: $210mm、310mm$	$U=0.007 \mu m$		
87	Magnetic and Eddy Current Measuring Instrument for Coating Thickness	Length	V.R.of Magnetic and Eddy Current Measuring Instrument for Coating Thickness JJG 818	Thickness measuring($0.01 \sim 0.05$)mm	$U=0.1 \mu m$		
				Thickness measuring($0.05 \sim 10$) mm	$U=(0.6\% \times H/2) \mu m (H: \mu m)$		
				Standard thickness sheet: $H \leq 50 \mu m$	$U=0.2 \mu m$		
				Standard thickness sheet: $H > 50 \mu m$	$U=0.4\%H(H:mm)$		



No. CNAS L0893

第 10 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
88	Ultrasonic Thickness Instruments	Length	Calibration Specification for Ultrasonic Thickness Instruments JJF 1126	(0.1~10) mm	$U=0.02 \text{ mm}$		
				(10~200) mm	$U=(0.003+H/600)\text{mm}(H:\text{mm})$		
89	Fineness of Grind Gage	Length	V.R.of Fineness of Grind Gage JJG 905	(0~150) μm	$U=(0.5+L/500) \mu\text{m}(L: \mu\text{m})$		
90	Roundness Flick Calibration Standard	Length	Calibration Specification for Roundness Flick Calibration Standard JJF 1485	(0.2~50) μm	$U_{\text{rel}}=2\%$		
91	*Extensometer	Length	V.R.of Extensometer JJG 762	(0~0.3)mm	$U=0.5 \mu\text{m}$		
				(0.3~25)mm	$U_{\text{rel}}=0.15\%$		
92	Coulomb Thickness Measurement Instrument	Length	Calibration Specification for Electrolytic (Coulometric) Coating Instruments Thickness JJF 1707	(0.5~50) μm	$U_{\text{rel}}=3\%$		
93	Coordinate Measuring Spheres	Length	C.S. for Coordinate Measuring Spheres JJF 1422	Roundness: ϕ (5~100)mm	$U=0.02 \mu\text{m}+5 \times 10^{-8}D$		
				Diameter: ϕ (5~100)mm	$U=0.2 \mu\text{m}+5 \times 10^{-7}D$		
94	Single score Speumen	Length	C. S.for single score Speumen NIMTT(CM) 001	(0.1~0.5) μm	$U_{\text{rel}}=5\%$		
				(>0.5~1.5) μm	$U_{\text{rel}}=3\%$		
				(>1.5~50) μm	$U_{\text{rel}}=2\%$		
95	Standard Ring Gauge	Length	V.R.of Standard ring gauge JJG 894	$D:(1\sim400)\text{mm}$	$U=0.35 \mu\text{m}+3 \times 10^{-6}D$		
96	Radius Gauge	Length	V.R.of Radius Gauge JJG 58	$R:(1\sim25)\text{mm}$	$U=2 \mu\text{m}+10^{-4}L$		
97	*Length measuring machine	Length	C.S. for Length Measuring Machine JJF 1066	Decimeter scale:(0~6000)mm	$U=0.3 \mu\text{m}+4 \times 10^{-6}L$		



No. CNAS L0893

第 11 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				Millimeter scale:(0~100)mm	$U=0.1 \mu\text{m}+4 \times 10^{-6}L$		
				Micron scale::(-100~+100) μm	$U=0.09 \mu\text{m}$		
98	Light-section Microscope	Roughness	C. S.for Light-Section Microscopes JJF 1092	(1.0~80) μm	$U_{\text{rel}}=5\%$		
99	Interferal microscope	Roughness	V.R.of Interference Microscopes JJG 77	(0.01~1.0) μm	$U_{\text{rel}}=5\%$		
100	Contact (Stylus) Instruments of Surface Roughness Measurement by the Profile Method	Roughness	C.S.for Contact (Stylus) Instruments of Surface Roughness Measurement by the Profile Method JJF 1105	$R_a:(0.025 \sim 25)\mu\text{m}$	$U_{\text{rel}}=4\%$		
101	*Measurement Standard instrument of roundness and cylindricity	Roundness	V.R.of Measurement Standard Instrument of Roundness and Cylindricity JJG 429	(50~500)mm	$U_{\text{rel}}=2\%$		
102	standard sphere and standard hemisphere	Roundness	V. R. of standard sphere and standard hemisphere GJB 8627	Standard Hemisphere: Φ (25~50)mm	$U=0.02 \mu\text{m}+5 \times 10^{-8}L$		
103	Torsional comparator	Length	V.R.of Microcator JJG 118	(-100~+100) μm	$U=0.2 \mu\text{m}$		
104	*Spherometer	Length	Calibration Specification for Spherometers JJF 1831	(-15~+15)mm	$U=0.4 \mu\text{m}$		
105	*Pneumatic Measuring Instrument for Micrometers	Length	V.R.of Pneumatic Measuring Instrument for Micrometers JJG 356	(0~60) μm	$U=0.4 \mu\text{m}$		



No. CNAS L0893

第 12 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
106	Wedge-feet Calibrator for Micrometers	Length	V.R.of Wedge-feet Calibrator for Micrometers JJG 525	(0~2)mm	$U=0.1 \mu m+1 \times 10^{-4}L$		
107	*Flatness interferometer with Isoclinic Circle Fring	Length	V.R.of Flatness Interferometer with Isoclinic Circle Fringe JJG 661	(0~500)mm	$U=0.009 \mu m$		
108	*Capacitance Comparators	Length	C.S.for Capacitance Comparators JJF 1944	(0~2)mm	$U=0.1 \mu m+1 \times 10^{-4}L$		
109	Dial Snap Gauges	Length	C.S.for Dial Snap Gauges JJF 1253	(5~100)mm	$U=5 \mu m+1 \times 10^{-4}L$		
110	*Laser Diameter Measuring Gauges	Length	C.S.for Laser Diameter Measuring Gauges JJF 1250	(0.1~30)mm	$U=0.3 \mu m$		
111	Measuring Instrument for Laser Parallelism of Micrometers	Length	C.S. for Measuring Instrument for Laser Parallelism of Micrometers JJF 1252	(150~2000)mm	$U=0.7 mm$		
112	Precision Bore Diameter Measuring Instrument	Length	C. S. for Measuring Equipment of Diameter of Micro and Small JJF 1806	(1~20)mm	$U=0.2 \mu m$		
113	Inductive Micrometers	Length	C.S. for Inductive Micrometers JJF 1331	$(-1000 \sim -0.01) \mu m$; $(+0.01 \sim +1000) \mu m$	$U_{rel}=0.1\%$		
114	Reinforced Concrete Covermeter and Floorslab Thickness Tester	Length	C.S.for reinforced concrete covermeter and floorslab thickness tester JJF 1224	(10~200)mm	$U=0.7 mm$		
				Reinforcing Bar Shield block:(6~200)mm	$U=0.1 mm$		
				Slab thickness block:(50~200)mm	$U=0.1 mm$		
115	Feeler Gauges	Length	V.R.of Feeler Gauges JJG 62	(0.02~3.00)mm	$U=2 \mu m$		



No. CNAS L0893

第 13 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
116	*Measuring System of Coordinate Position	Length	C.S. for Measuring System of Coordinate Position JJF 1251	(0~20)m	$U=0.1 \mu m+1 \times 10^{-7}L$		
117	*Contact type inter-ferometer	Length	V.R.of Contact-type interferometer JJG 101	(-10~+10) μm	$U=0.01 \mu m$		
118	*Optimerer	Length	V.R.of Optimerer JJG 45	(-100~+100) μm	$U=0.02 \mu m$		
119	*Horizontal metro-scope	Length	C.S. for Length Measuring Instrument JJF 1189	(0~500)mm	$U=0.2 \mu m+2 \times 10^{-6}L$		
120	Calibration Specification for Grating Micrometers	Length indication error	C.S. for Grating Micrometers JJF1682	(0~100)mm	$U=0.1 \mu m+2 \times 10^{-6}L$		
121	Calibrator of Extensometers	Length	C. S. for Calibrator of Extensometers JJF1096	(0~0.3)mm	$U=0.2 \mu m$		
				(>0.3~50)mm	$U_{rel}=0.04\%$		
122	Callipers for Welding Inspection	Length	V. R.of Callipers for Welding Inspection JJG704	(0~60)mm	$U=0.02 mm$		
123	*Surface Plates	Flatness	V. R.of Surface Plates JJG 117	(160×160~5000×3000)mm	$U=0.8 \mu m+1 \times 10^{-6}L$ (L: Span of deck)		
124	*Straight Edges	Straightness	C.S.for Straight Edges JJF 1097	(300~6300)mm	$U=0.4 \mu m+3 \times 10^{-7}L$		
125	*Articulated Arm Coordinate Measuring Machine	length	C.S. for Articulated Arm Coordinate Measuring Machine JJF 1408	(30~5000)mm	$U=23 \mu m$		
126	Standard Grid Plate	Length	V. R. of Standard glass grid plate JJG 832	(0~400)mm	$U=1 \mu m$		



No. CNAS L0893

第 14 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
127	*Orifice Plate Comprehensive Measurement System	Length	C. S. for Orifice Plate Comprehensive Measurement System NIMTT(CM) 033	(12.5~300)mm	$U_{rel}=2\%$		
128	*Optical 3D measuring systems - Optical systems based on area scanning	Geometric Quantity	C. S. for Optical 3D measuring systems Optical systems based on area scanning VDI/VDE 2634 Blatt 2	(16~1500)mm	$U=5\text{ }\mu\text{m}$		
129	*Clearance Detector of Rolling Stock Gauge for Standard Gauge Railway	length	V. R. of Clearance Detector of Rolling Stock Gauge for Standard Gauge Railway JJG (tiedao) 180	(0~5)m	$U=0.6\text{ mm}$		
130	Laser Interferometers	Angle	V. R. of Laser Interferometers JJG 739-2005 JJG 739	(0~360)°	$U=0.3''$		
		Length		Wavelength:633nm	$U_{rel}=2.5\times 10^{-8}$		
131	*self-calibration measuring instrument with automobile model cutting system	Length	C.S.for self-calibration measuring instrument with automobile model cutting system NIMTT(CM) 138	(0~40)m	$U=3\text{ }\mu\text{m}+1.6\times 10^{-6}L$		
132	Moulds	Roughness	C.S. for Moulds JJF 1307	$Ra:(1.6\sim 3.2)\text{ }\mu\text{m}$	$U_{rel}=6\%$		
		Length		(100~600)mm	$U=0.30\text{ mm}$		
133	Standard stick of Measuring spheres	Length	V.R.for Standard stick of Measuring spheres JJG (jun gong) 177-2019	(0~2000)mm	$U=3.0\text{ }\mu\text{m}$		
134	*Equator	Length	C. S. for Equator NIMTT(CM) 147	(0~1000)mm	$U=1.2\text{ }\mu\text{m}+1.5\times 10^{-6}L$		



No. CNAS L0893

第 15 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
135	*Scanning Probe Microscopes	Length	C.S. for Scanning Probe Microscopes JJF 1351	(0~200) μ m	$U=5$ nm		
136	3D Optial Area Sanning Measurement Instrments	Length	C.S. for 3D Optial Area Sanning Measurement Instrments JJF(jun gong) 116	(16~1500)mm	$U=5$ μ m		
137	Pi Tapes	Length	C.S. for Pi Tapes JJF 1423	(0.009~16)m	$U=0.02$ mm+ $1 \times 10^{-5}D$		
138	Wedge-Shape Filler Gauges	Length	C.S. for Wedge-Shape Filler Gauges JJF 1548	(1~60)mm	$U=11$ μ m		
139	Calibration Specification for Linear Displacement Sensors	length	C.S. for Linear Displacement Sensors JJF 1305	(0~3000)mm	$U=0.08\%$ FS	Calibration only for linear displacement sensors with direct digital output length	
140	Calibration Specification for Calibration Devices of Four-wheel Aligner	Angle	C.S. for Calibration Devices of Four-wheel Aligner JJF 1489	(-30~+30) $^{\circ}$	$U=28''$		
141	Crack width measuring instrument	length	C.S. for Concrete Width and Depth Measuring Instruments JJF 1334	breadth(0.01~10)mm	$U=3$ μ m		
				depth(20~500)mm	$U_{rel}=0.2\%$		
				Width standard plate (0.01~6) mm	$U=3$ μ m		
				Depth standard block (20~500) mm	$U=1$ mm		



No. CNAS L0893

第 16 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
142	Image measuring instrument	length	C.S. for Imaging Probe Measuring Machines JJF 1318	(0~1000)mm	$U=0.4 \mu m+0.4 \times 10^{-6}L$		
143	Tread depth ruler	length	C.S. for Tire Tread Depth Gauges JJF 1477	(0~100)mm	$U=0.01 \text{ mm}$		
144	Borehole inclinometer	Angle	C.S. for Borehole Clinometers JJF 1550	$(-30 \sim +30)^{\circ}$	$U=0.008\%FS$		
145	Block comparator	length	C.S. for Gauge Block JJF 1304	$(-100 \sim +100) \mu m$	$U=0.017 \mu m$		
146	*White Light Interference Profiler	Length	C. S. for White Light Interference Profiler JJF(jungong) 201	(0~300)mm	$U=8 \text{ nm}$		
147	Step Profiler	Length	C.S. for Step Profiler JJF(jungong) 129	$0.1 \mu m \sim 5 \text{ mm}$	$U=3 \text{ nm}$		
148	Pipe Robot	length	C. S. for Pipe Robot Mobile Positioning NIMTT(CM) 144	(0~100)m	$U=0.59 \text{ mm}+9 \times 10^{-2}L$	Calibration only for Mobile Positioning	
		angle		$(0 \sim 90)^{\circ}$	$U=0.51^{\circ}$		
149	three-dimensional spherical center distance calibrator	Length	C.S. for three-dimensional spherical center distance calibrator NIMTT(CM) 137	Center distance:(0~1)m	$U=1.2 \mu m$		
				Diameter:(0~100)m	$U=0.5 \mu m$		
				Roundness:(0~100)m	$U=0.7 \mu m$		
150	Laser Marker	angle	C. S. for Laser Marker JJF (jin) 01	$(-8.5 \sim +8.5)^{\circ}$	$U=6.6''$		
		Length		(0~10)mm	$U=0.3 \text{ mm}$		
151	*digital laser plane interferometer	Length	C. S. for digital laser plane interferometer GJB 8704	$D:150 \text{ mm}$	$U=0.06 \mu m$		
152	Ultrasonic test block	Length	C.S. for Blocks used in Ultrasonic Testing JJF 1487	(0~1000)mm	$U=0.020 \text{ mm}$		
		Roughness		$Ra:(0.025 \sim 10) \mu m$	$U_{rel}=6\%$		



No. CNAS L0893

第 17 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
153	Master Gears	Length	V. R. of Master Gears JJG 1008	(3~100)mm (> 100~280) mm	$U=2.5 \mu m$		
154	Taper Thread Gauges	Length	Calibration Specification for Taper Thread Gauges JJF (J1) 185	Thread Ring Gauge:(5~500) mm Thread Plug Gauge:(1~500) mm	$U=2.8 \mu m+2 \times 10^{-6}L$ $U=2.8 \mu m+2 \times 10^{-6}L$		
		Angle		(0~45)°	$U=3'$		
155	Total station	angle	V. R. of Electronic TachometerTotal Station JJG100,V. R. of Industrial Measurement Total Stations JJG1152	Horizontal :(0~360)° Horizontal :(-60~+60)°	$U=0.3''$ $U=0.3''$	Length: Just calibrate Point-to-point Length error	
		Length		(0~14)m	$U=0.2mm$		
156	Two-dimensional Optics Calibration Target	length	C.S.for Two-dimensional Optics Calibration Target NIMTT(CM) 140	(20~400)mm	$U=1.0 \mu m+2 \times 10^{-5}L$		
157	Calibration Specification for Paint Film Scriber	Length	C. S. for Paint Film Scriber JJF (ji) 137	Knife spacing:(1~3)mm	$U=2 \mu m$		
		Angle		(20~30)°	$U=0.1^\circ$		
158	Metallographic microscope	length	Calibration Specification for Metallurgical Microscopes JJF 1914	Objective Lens:(1~100)X	$U_{rel}=1.2\%$		
159	Biological microscope	length	C.S. for Biological Microscopes JJF 1402	Objective Lens:(1~100)X	$U_{rel}=1.2\%$		
160	*Thread Gauge Scanners	length	Calibration Specification for Thread Gauge Scanners JJF1950	(0~500)mm	$U=(0.8 \sim 3.0) \mu m$		
		angle		(0~80)°	$U=6'' \sim 4'$		
	Le	length	V. R. of Le Chatelier,Tester for Determining Expansion of	(0~200) mm	$U=5 \mu m$		



No. CNAS L0893

第 18 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	for Determining Expansion of Le Chatelier Needles	quality	Le Chatelier Needles JJG (jiao tong) 093	(0~300)g	$U=0.03 \text{ g}$		
162	Surface Profile Gauges	Length	C. S. for Surface Profile Gauges JJF 1476	(0~6.5) mm	$U=1.7 \mu\text{m}$		
163	Material Measures of Length Measuring Instrument with Optical Principle	Length	C. S. for Material Measures of Length Measuring Instrument with Optical Principle JJF 1941	(0~500)mm	$U=0.5 \mu\text{m}$		
		Angle		$0^\circ \sim 360^\circ$	$U=0.4''$		
		Straightness		(0~500)mm	$U=(0.02+0.005L) \mu\text{m}(L:\text{mm})$		
		Surface roughness		$Ra(0.01 \sim 10) \mu\text{m}$	$U_{\text{rel}}=5.0\%$		
		circel run-out		(0~40)mm	$U=(0.02+0.005L) \mu\text{m}(L:\text{mm})$		
		Flatness		$(70 \times 70) \text{ mm} \sim (400 \times 500) \text{ mm}$	$U=(0.02+0.007L) \mu\text{m}(L:\text{mm})$		
		parallelism		(0~600)mm	$U=(0.02+0.009L) \mu\text{m}(L:\text{mm})$		
		verticality		$H:(0 \sim 70) \text{ mm}$	$U=(0.02+0.010L) \mu\text{m}(L:\text{mm})$		
		offset		(0~200)mm	$U=(0.02+0.011L) \mu\text{m}(L:\text{mm})$		
164	Step Wedge Used in Radiographic Testing	length	C. S. for Step Wedge Used in Radiographic Testing JJF(ji xie)1016	(0~500) mm	$U=0.004 \text{ mm}$		
		Surface roughness		$Ra(0.01 \sim 10) \mu\text{m}$	$U_{\text{rel}}=6.0\%$		
165	Micropattern Standards	Length	C. S. for micropattern standards JJF 1917	line spacing: (0~20) mm	$U=0.2 \mu\text{m}$		
				grid spacing: (0~50) mm	$U=0.5 \mu\text{m}$		
				diameter: (0~50) mm	$U=0.5 \mu\text{m}$		



No. CNAS L0893

第 19 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Angle		(0~360) °	$U=0.01^{\circ}$		
166	measuring instruments for contact wire geometric parameter	length	V.R. of measuring instruments for contact wire geometric parameter JJG(tie dao)150	gauge: (1410~1470) mm	$U=0.2\text{mm}$	Calibration only portable	
				superelevation: (-185~185) mm	$U=0.2\text{ mm}$		
				stagger value: (-600~600) mm	$U=1.1\text{ mm}$		
				Height of contact line: (5100~6500) mm	$U=1.1\text{ mm}$		
				Mast gauge: (2400~6500) mm	$U=0.9\text{ mm}$		
167	*calibrator for measuring instruments for contact wire geometric parameter	length	calibrator for measuring instruments for contact wire geometric parameter JJG(tie dao)210	gauge: (1410~1470) mm	$U=0.1\text{ mm}$		
				superelevation: (-185~185) mm	$U=0.1\text{ mm}$		
				stagger value: (-610~610) mm	$U=0.5\text{ mm}$		
				Height of contact line: (5100~6500) mm	$U=0.5\text{ mm}$		
				mast gauge: (2400~6500) mm	$U=0.5\text{ mm}$		
168	Liquid Level Transducer	length	C. S. for Liquid Level Transducer NIMTT(CM) 166	(0.01~3000) mm	$U_{\text{rel}}=0.07\%$		
169	Industrial electronic endoscope	Illuminance	C. S. for Industrial Electronic Endoscope JJF (liao) 437	(50~3000)lx	$U_{\text{rel}}=1.3\%$		
		length		(0~30)mm	$U=0.08\text{ mm}$		
170	Straightness measuring instrument	straightness accuracy	C. S. of straightness measuring instrument JJF 1890	(0~1000) mm	$U=1.0\text{ }\mu\text{m}$		

No. CNAS L0893

第 20 页 共 181 页



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
171	Conduit Detection Robots	Length	C. S. of Conduit Detection Robots JJF (e) 60	(0~100) m	$U=0.2 \text{ mm}+2 \times 10^{-4}L$		
172	Digital Levels	Angle	V. R. of Digital Levels JJG (ce hui) 2101	(-25~+25) "	$U=2.8''$		
173	*Optical 3D Measuring Systems Based on Structured Light Scanning	Length	C. S. for Optical 3D Measuring Systems Based on Structured Light Scanning JJF 1951	(0~1.5) m	$U=5 \mu\text{m}$		
174	*Scanning Electronic Microscopes (SEM)	Length	C. S. for Scanning Electronic Microscope (SEM) JJF 1916	(100nm~10 μm)	$U=3 \text{ nm}$		
175	Standard Ball Bars	Length	C. S. for Standard Ball Bars JJF 1859	sphere diameter: (0~300) mm	$U=0.5 \mu\text{m}$		
				spherical shape error: (0~10) mm	$U=0.7 \mu\text{m}$		
				from its: (0~1.5) m	$U=2 \mu\text{m}+0.2L$		
176	Ground Penetration Radar	speed	Calibration Specification for Ground Penetration Radar JJF (qian) 58	Radar wave velocity: (280~320) mm/ns	$U=0.8\%$		
		Length		Measuring thickness: (0~0.5) m	$U=1.0\%$		
177	Step Specimen	Length	V. R. of Single Groove Specimen and Step Specimen JJG (军工) 175	$H: (0.01 \sim 0.02) \mu\text{m}$	$U=1.3 \text{ nm}+2.4\%H$		
				$H: (>0.02 \sim 0.1) \mu\text{m}$	$U=1.8 \text{ nm}+2.0\%H$		
				$H: (>0.1 \sim 2) \mu\text{m}$	$U=2.5 \text{ nm}+0.8\%H$		
				$H: (>2 \sim 10) \mu\text{m}$	$U=3 \text{ nm}+0.5\%H$		
				$H: (>10 \sim 25) \mu\text{m}$	$U=3 \text{ nm}+0.3\%H$		



No. CNAS L0893

第 21 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				$H: (>25\sim100) \mu\text{m}$	$U=3 \text{ nm}+0.2\%H$		
178	*Measuring contact angle with image method	Angle	C. S. for Instrument for measuring contact angle with image method JJF (su) 219	$6^\circ \sim 180^\circ$	$U=0.06^\circ$		
179	Standard Angle slice	Angle	C. S. for Instrument for measuring contact angle with image method JJF (su) 219	$6^\circ \sim 180^\circ$	$U=0.03^\circ$		
180	Ringelmann smoke telescope	Length	C. S. for Ringelmann smoke telescope JJF (jin) 144	$(0\sim5)$ ringelmann number	$U=0.04$ ringelmann number		
181	Coating thickness standard sheet	Length	C. S. For X-Ray Fluorescence Coating Thickness Instruments JJF 1306	$(0.01\sim80) \mu\text{m}$	$U_{\text{rel}}=7.0\%$		
182	X-Ray Fluorescence Coating Thickness Instruments	Length	Calibration Specification For X-Ray Fluorescence Coating Thickness Instruments JJF 1306	$(0.01\sim80) \mu\text{m}$	$U_{\text{rel}}=6.5\%$		
183	Resolution Target	Length	V.R.of Resolution Target JJG 827	$(0.0005\sim400)\text{mm}$	$U=1 \mu\text{m}$		
184	Check the sample	Length	C. S. for Check the sample JJF (chuan) 110	$(0\sim200) \text{mm}$	$U=0.008 \text{ mm}$		
185	Laser Tunnel Section Detector	Length	C. S. for Laser Tunnel Section Detector JJF (e) 50	$(0\sim50) \text{m}$	$U=1 \text{ mm}$		
		Angle		$(30\sim330)^\circ$	$U=0.08^\circ$		
186	Bridge Deflection Instrument	Length	C. S. for Bridge Deflection Instrument JJF(min)1077	$(10\sim400) \text{mm}$	$U_{\text{rel}}=0.4\%$		
187	*Contact (Stylus) Surface Contour Tester	Length	C. S. for Contact (Stylus) Surface Contour Tester JJF (min) 1043	$(0\sim100) \text{mm}$	$U=0.8 \mu\text{m}$		
		angle		$0^\circ \sim 360^\circ$	$U=41''$		



在线扫码获取验证

No. CNAS L0893

第 22 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
188	*Steel tape inspection platform based on image measurement	Length	Specification for calibration of Steel tape inspection platform based on image measurement JJF (chuan) 184	(0~20) m	$U=0.07$ mm		
189	*Auto Tester For Steel Rule	Length	C. S. for Auto Tester For Steel Rule JJF (zhe) 1129	(0~1000)mm	$U=8$ μm		
190	Industrial Endoscope	Angle	Calibration specification for Industrial Endoscope JJF(Ji) 69	$0.5^{\circ} \sim 360^{\circ}$	$U_{rel}=2\%$		
				(0.1~200) klx	$U_{rel}=5\%$		
191	*	Length	Calibration Specification for Industrial X-ray CT Coordinate Measuring Machine JJF (军工) 284	(0~116) mm	$U=4.0$ μm		
192	Spirit level	Angle	Calibration Specification for Level Rules JJF1085	(0.5~10) mm/m	$U_{rel}=6\%$		
193	Slit lamp Microscopes	Length	Calibration specification for Slit lamp Microscopes JJF (浙) 1169	(0~10) mm	$U_{rel}=0.07$		
194	Stereomicroscope	Length	Calibration Specification for Stereomicroscope JJF(闽)1063	(0.1~10) mm	$U_{rel}=(0.4+0.1 \gamma)\%$		
195	Laser Centering Instrument	Length	Laser Centering Instrument NIMTT(CM) 168	(0.1~200) mm	$U_{rel}=1\%$		
		Angle		$(0 \sim \pm 360)^{\circ}$	$U=6''$		
196	Gyrotheodolites	angle	C. S. for Gyrotheodolites JJF 1350	$(0 \sim 360)^{\circ}$	$U=1.5''$		
197	Baseline and Basenet	Length	C. S. for Baseline and Basenet JJF 1214	(0~100)km	$U=0.10\text{mm}+0.34 \times 10^{-6}D$		
198	Spherical	Length	Calibration Specification for Large-Scale Laser Scanners	(0~24)m	$U=0.03\text{mm}$		

No. CNAS L0893

第 23 页 共 181 页



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	Scanner	Length	Base on Spherical Coordinate JJF (Military) 283	($>24\sim1018$)m	$U=0.5\text{mm}+0.8\times10^{-6}D$		
199	Global Positioning System(GPS)Receiver	measurement error	Calibration Specification for GPS Receiver of Surveying Model JJF1118	($0\sim1.2$)km	$U=1.0\text{mm}+1.0\times10^{-6}D$ (short base line)		
		Consistency of antenna phase center		($>1.2\sim20$) km	$U=3.0\text{mm}+1.0\times10^{-6}D$ (Long base line)		
				($0\sim8.5$)m	$U=0.8\text{ mm}$		
200	GNSS Receivers Working at Navigation	positioning error	Calibration Specification for GNSS Receivers Working at Navigation JJF 1942	4m $\sim$ 42km	$U=2.72\text{m}$		
		Speed error		($0\sim15$) km/s	$U=0.04\text{m/s}$		
		acceleration error		($0\sim1000$)m/s ²	$U=0.05\text{m/s}^2$		
201	GNSS Vehicle Travelling Data Recorders	Positioning deviation	Calibration Specification for GNSS Vehicle Travelling Data Recorders JJF 1921	($0\sim1000$) km	$U=1.0\text{m}$		
		Speed deviation		($0\sim15$) km/s	$U=0.02\text{m/s}$		
		mileage		($0\sim1000$)km	$U=6.2\text{m}$		
		Time recording error		($0\sim24$)h	$U=0.1\text{ s}$		
		sensitivity		($-150\sim-100$)dBm	$U=0.3\text{dB}$		
202	Ground Penetrating Radar	Length	Highway Structure Defect and Layer Thickness Ground Penetration Radar JJG (交通) 124	($50\sim100$) mm	$U=0.74\text{mm}$		
		Length		($>100\sim715$) mm	$U_{\text{rel}}=1.1\%$		
		wave velocity		($180\sim300$) mm/ns	$U_{\text{rel}}=0.2\%$		



No. CNAS L0893

第 24 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
203	Deflection /Camber instrument	Length	Bridge Deflection Measuring Instrument JJG (交通) 143	(0~500)mm	$U=0.05\text{mm}$		
204	Verification Regulation of Measure for Gauge for Measuring Distance Between Inside Rim Faces of Railway Wheels	Geometry	Verification regulation of internal distance gauge of railway locomotive and rolling stock wheel JJG 1159	(1345~1365) mm	$U=0.08\text{ mm}$		
205	Rules for Measuring Center Height of Coupler for Railway Locomotive and Vehicle	Geometry	Rules for Measuring Center Height of Coupler for Railway Locomotive and Vehicle JJG 1150	(650~950) mm	$U=0.11\text{mm}$		
206	Verification Regulation of Gauges for Measuring Distance Between Inside Rim Faces of Railway Wheels	Geometry	Calibrators of Wheel- Checkers for Railway Locomotives and Vehicles JJG 1153	(1345~1365) mm	$U=0.06\text{mm}$		
207	Calibrators of Wheel-Checkers for Railway Locomotives and Vehicles	Geometry	Calibrators of Wheel- Checkers for Railway Locomotives and Vehicles JJG 1155	(3~70) mm	$U=0.02\text{ mm}$		




在线扫码获取验证

No. CNAS L0893

第 25 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
208	Verification regulation of height measuring device of railway frog structure	Geometry	Verification regulation of height measuring device of railway frog structure JJG 1183	(-15~+15) mm	U=0.05 mm		
209	Calibrators for Railway Switch Offset Rule	Geometry	Calibrators for Railway Switch Offset Rule JJG 1109	(100~1800) mm	U=0.12 mm		
210	Railway Switch Offset Rules	Geometry	Railway Switch Offset Rules JJG 1108	(100~1800) mm	U=0.18 mm		
211	Inspecting Instruments for Railway Track	Geometry	Inspecting Instruments for Railway Track JJG 1090	1410~1470) mm	U=0.10mm		
212	Calibrator for Inspecting Instruments for Railway Track	Geometry	Calibrator for Inspecting Instruments for Railway Track JJG 1091	1410~1470) mm	U=0.04 mm		
213	Measuring Scale for Wheel-Diameter	Geometry	Means of Measuring Instrument for Wheel-Diameter of Railway Locomotives and Vehicles-Part1:Means of Ruler for Wheel-Diameter JJG 1081.1	(760~1270) mm	U=0.06 mm		
214	Measuring Instrument for Wheel-Diameter of Railway	Geometry	Measuring Instrument for Wheel-Diameter of Railway Locomotives and Vehicles-Part2:Measuring Tools for Wheel-Diameter JJG 1081.2	(760~1270) mm	U=0.04mm		
215	Wheel-Checker for Railway Locomotives and Vehicles	Geometry	Wheel-Checker for Railway Locomotives and Vehicles JJG 1080	(3~95) mm	U=0.03mm		



在线扫码获取验证

No. CNAS L0893

第 26 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
216	Measuring Ruler for Difference of Wheels Set Position and Brake Discs Set Position of Railway Vehicles	Geometry	Measuring Ruler for Difference of Wheels Set Position and Brake Discs Set Position of Railway Vehicles JJG 1110	(0~500) mm	U=0.07mm		
217	Inspection of rail straightness measuring instrument	Geometry	JJG(铁道)167	(0~3000) mm	U=0.02mm		
218	Verification Regulation of Wheel Contour Gauge for Railway Locomotive and Vehicle	Geometry	Verification Regulation of Wheel Contour Gauge for Railway Locomotive and Vehicle JJG (铁道) 175	(-69~100) mm	U=5 μ m		
219	Wear Tools for Rail	Geometry	Wear Tools for Rail JJG 1127	(0~25) mm	U=0.15mm		
II Thermology measuring instrument							
1	*Temperature and Humidity Standard Chambers	Temperature	Calibration Specification for Temperature and Humidity Standard Chambers JJF1564	uniformity:(5~50)℃	U=0.05℃		
				fluctuation:(5~50)℃	U=0.02℃		
		Humidity		uniformity:10%RH~90%RH	U=0.5%RH		
				fluctuation:10%RH~90%RH	U=0.1%RH		
2	*recorders for industrial-process measurement	Temperature	Verification regulation of the recorders for industrial-process measurement JJG 74	RTD: (-200~850)℃	U=(0.10~0.15)℃		
				K、N、J:(-200~0)℃	U=(0.18~0.12)℃		



No. CNAS L0893

第 27 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				K、N、J:(0~1200)℃	$U=(0.12\sim0.19)^\circ\text{C}$		
				S:(0~1600)℃	$U=(0.7\sim0.4)^\circ\text{C}$		
				B:(800~1800)℃	$U=(0.5\sim0.4)^\circ\text{C}$		
				T:(-200~400)℃	$U=(0.27\sim0.11)^\circ\text{C}$		
				E:(-200~1000)℃	$U=(0.21\sim0.12)^\circ\text{C}$		
3	*Digital temperature indicators and/controllers	Temperature	Verification regulation of digital temperature indicators and controller JJG 617	RTD:(-200~850)℃	$U=(0.10\sim0.15)^\circ\text{C}$		
				K、N、J:(-200~0)℃	$U=(0.18\sim0.12)^\circ\text{C}$		
				K、N、J:(0~1200)℃	$U=(0.12\sim0.19)^\circ\text{C}$		
				S:(0~1600)℃	$U=(0.7\sim0.4)^\circ\text{C}$		
				B:(800~1800)℃	$U=(0.5\sim0.4)^\circ\text{C}$		
				T:(-200~400)℃	$U=(0.27\sim0.11)^\circ\text{C}$		
				E:(-200~1000)℃	$U=(0.21\sim0.12)^\circ\text{C}$		
4	*Temperature transmitter	Temperature	Calibration specification of the temperature transmitter JJF 1183	RTD:(-200~850)℃	$U=(0.08\sim0.20)^\circ\text{C}$		
				K:(-200~1200)℃	$U=0.5^\circ\text{C}$		
				E:(-200~1000)℃	$U=0.4^\circ\text{C}$		
				S:(0~1600)℃	$U=(0.9\sim0.6)^\circ\text{C}$		
				Match RTD:(-60~300)℃	$U=(0.1\sim0.2)^\circ\text{C}$		
				Match S:(300~1100)℃	$U=1.3^\circ\text{C}$		
				Match K,N:(-40~1200)℃	$U=(0.5\sim1.5)^\circ\text{C}$		



No. CNAS L0893

第 28 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
5	Temperature itinerant detecting instrument	Temperature	Calibration specification for temperature itinerant detecting instrument JJF 1171	(-60~100)°C	U=0.08°C		
				(100~300)°C	U=0.11°C		
6	Standard platinum-30% rhodium/platinum- 6%rhodium thermocouple	Temperature	Verification regulation of the standard platinum-30% rhodium/platinum-6% rhodium thermocouple JJG 167	1100°C	U=2.0°C		
				1200°C	U=1.9°C		
				1300°C	U=1.9°C		
				1400°C	U=1.9°C		
				1500°C	U=1.9°C		
7	Standard platinum-10% rhodium/platinum thermocouple	Temperature	verification regulation of the standard platinum-10% rhodium/platinum thermocouple JJG 75	419.527°C	U=0.24°C		
				660.323°C	U=0.26°C		
				1084.62°C	U=0.29°C		
8	The working noble metal thermocouples	Temperature	Verification regulation of the working noble metal thermocouples JJG 141	S: 419.527°C	U=0.48°C		
				S: 660.323°C	U=0.49°C		
				S: 1084.62°C	U=0.55°C		
				B:1100°C	U=2.0°C		
				B:1300°C	U=2.0°C		
				B:1500°C	U=2.0°C		
9	Standard platinum resistance thermometer	Temperature	Verification regulation of standard platinum resistance thermometer JJG 160	(83.8058 ~273.16)K	U=3.4mK (83.8058K)		
				• (83.8058~273.16)K	U=1.6mK (234.3156K)		
				(83.8058~273.16)K	U=1.0mK (273.16K)		



No. CNAS L0893

第 29 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				(273.16 ~ 933.473)K	$U=2.3\text{mK}$ (505.078K)		
				(273.16~933.473)K	$U=3.5\text{mK}$ (692.677K)		
				(273.16~933.473)K	$U=4.5\text{mK}$ (933.473K)		
10	Industry platinum copper resistance thermometers	Temperature	Verification Regulation of industry platinum and copper resistance thermometers JJG 229	(-80~+300) °C	$U=(0.02\sim0.05)$ °C		
				-196°C	$U=0.03$ °C		
11	Standard clinical thermometer	Temperature	Verification regulation of standard clinical thermometer JJG 881	(35~45)°C	$U=0.018$ °C		
12	*On-line Temperature Measuring System	Temperature	Calibration Specification for On-line Temperature Measuring System JJF (Chuan) 143	(-30~300)°C	$U=0.1$ °C		
				(300~650)°C	$U=1.0$ °C		
13	Temperature Calibration Devices for Polymerase Chain Reaction Analyzers	Temperature	Calibration Specification of Temperature Calibration Devices for Polymerase Chain Reaction Analyzer JJF 1821	(0~110)°C	$U=0.04$ °C		
14	Clinical Electronic Thermometers	Temperature	Verification Regulation of Clinical Electronic Thermometers JJG 1162	(35.0~41.3) °C	$U=0.03$ °C		
15	*Blood culture device	Temperature	Calibration Specification for Blood Culture device JJF(Chuan) 171	(20~50)°C	$U=0.2$ °C		
16	*hot aging test chambers	Temperature	Calibration Specification of Hot Aging Test Chambers JJF (Chuan) 162	(30~100)°C	$U=0.3$ °C		
				(100~300)°C	$U=0.5$ °C		



No. CNAS L0893

第 30 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				(300~500)°C	$U=1.3^{\circ}\text{C}$		
		air change rate		(5~200) times/h	$U_{\text{rel}}=5\%$		
17	*Temperature Controller for Transformers	Temperature	Calibration Specification of Temperature Controller for Transformers JJF (Chuan) 151	(-20~160)°C	$U=0.4^{\circ}\text{C}$		
18	*Passive Medical Cold Boxes	Temperature	Calibration Specification for Temperature Parameter of Passive Medical Cold Boxes JJF 1676	(-20~20)°C	$U=0.16^{\circ}\text{C}$		
19	*Environmental testing equipment	Temperature	Calibration Specification for Environmental Testing Equipment for Temperature and Humidity Parameters JJF 1101	(-80~0)°C	$U=0.2^{\circ}\text{C}$		
				(0~100)°C	$U=0.1^{\circ}\text{C}$		
				(100~300)°C	$U=0.3^{\circ}\text{C}$		
		Humidity		10%RH~95%RH	$U=1.5\%\text{RH}$		
20	Electric Ventilation Psychrometer	Temperature	Verification Regulation of Meteorological Ventilation Psychrometer JJG 993	(5~50)°C	$U=0.07^{\circ}\text{C}$		
		Humidity		10%RH~90%RH	$U=0.8\%\text{RH}$		
21	Meteorological Ventilation Psychrometer	air speed	Verification Regulation of Electric Ventilation Psychrometer JJG 204	(2~5)m/s	$U=0.6\text{m/s}$		
		Time		(4~8)min	$U=0.2\text{s}$		
22	Mechanical Thermo - hygrometers	Humidity	Verification Regulation of Mechanical Thermo - hygrometers JJG 205	30%RH~95%RH	$U=1\%\text{RH}$		
		Temperature		(5~50)°C	$U=0.07^{\circ}\text{C}$		
23	Precision Dew-point Hygrometers	Temperature	Verification Regulation of Precision Dew-point Hygrometers JJG 499	Dew-point: (-80~+20)°C	$U=(0.2\sim0.4)^{\circ}\text{C}$		



在线扫码获取验证

No. CNAS L0893

第 31 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
24	Resistance and Capacitance Dew Point Hygrometer	Temperature	Calibration Specification for Resistance and Capacitance Dew Point Hygrometer JJF 1272	Dew-point: $(-80 \sim +20) ^\circ\text{C}$	$U=(0.2 \sim 0.4) ^\circ\text{C}$		
25	Electrolytic hygrometers	Moisture	Verification Regulation of Electrolytic Hygrometers JJG 500	$(0.1 \sim 1000) \mu\text{L/L}$	$U_{\text{rel}}=4.4\% \sim 0.5\%$		
26	Secondary Standard Divided Flow Humidity Generator	Humidity	Verification Regulation of Secondary Standard Divided Flow Humidity Generator JJG 826	$10\%\text{RH} \sim 95\%\text{RH}$	$U=0.8\%\text{RH}$		
27	Temperature and Humidity Generator	Humidity	Calibration Specification for Temperature and humidity generator NIMTT(CM) 121	$10\%\text{RH} \sim 95\%\text{RH}$	$U=0.8\%\text{RH}$		
		temperature		$(5 \sim 50) ^\circ\text{C}$	$U=0.1 ^\circ\text{C}$		
28	Index detector	Temperature	Calibration Specification for WBGT index thermometer calibration specification JJF 1407	$(5 \sim 50) ^\circ\text{C}$	$U=0.07 ^\circ\text{C}$		
29	Base metal thermocouples	Temperature	Calibration Specification for Base Metal Thermocouples JJF 1637	K、N: $(-40 \sim 1200) ^\circ\text{C}$	$U=(0.26 \sim 1.2) ^\circ\text{C}$		
				E: $(-40 \sim 900) ^\circ\text{C}$	$U=(0.26 \sim 0.80) ^\circ\text{C}$		
				J: $(-40 \sim 750) ^\circ\text{C}$	$U=(0.26 \sim 0.70) ^\circ\text{C}$		
30	Standard Mercury-in-Glass thermometers	Temperature	Verification regulation of standard mercury-in-glass thermometers JJG 161	$(-60 \sim 300) ^\circ\text{C}$	$U=(0.03 \sim 0.05) ^\circ\text{C}$		
31	Liquid-in-Glass thermometers for working	Temperature	Verification regulation of liquid-in-class thermometer for working JJG 130	High precision: $(0 \sim 150) ^\circ\text{C}$	$U=(0.012 \sim 0.024) ^\circ\text{C}$		
				Working: $(-60 \sim 300) ^\circ\text{C}$	$U=(0.05 \sim 0.09) ^\circ\text{C}$		



No. CNAS L0893

第 32 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
32	*Bimetallic thermometers	Temperature	Calibration Specification for Bimetallic Thermometers JJF 1908	$(-60\sim 300)^{\circ}\text{C}$	$U=0.3^{\circ}\text{C}$		
33	Filled system thermometers	Temperature	Calibration Specification for Filled System Thermometers JJF 1909	$(-60\sim 300)^{\circ}\text{C}$	$U=0.4^{\circ}\text{C}$		
34	*Thermocouple calibration furnaces	Temperature	Testing Specification of temperature uniformity in thermocouple calibration furnaces JJF 1184	$(300\sim 1500)^{\circ}\text{C}$	$U=0.3^{\circ}\text{C}$		
35	Temperature data collector	Temperature	Calibration specification for temperature data collector NIMTT(CM) 018	RTD: $(-200\sim 850)^{\circ}\text{C}$	$U=(0.01\sim 0.03)^{\circ}\text{C}$		
				K,N,J: $(-200\sim 1200)^{\circ}\text{C}$	$U=(0.12\sim 0.19)^{\circ}\text{C}$		
				S: $(0\sim 1600)^{\circ}\text{C}$	$U=(0.7\sim 0.4)^{\circ}\text{C}$		
				E: $(-200\sim 1000)^{\circ}\text{C}$	$U=(0.21\sim 0.11)^{\circ}\text{C}$		
				B: $(800\sim 1800)^{\circ}\text{C}$	$U=(0.5\sim 0.4)^{\circ}\text{C}$		
36	*Thermostatic bath	Temperature	Measurement and Test Norm of Thermostatic Bath's Metrological Characteristics JJF 1030	Uniformity: $(-100\sim 40)^{\circ}\text{C}$	$U=(0.006\sim 0.002)^{\circ}\text{C}$		
				Uniformity: $(40\sim 300)^{\circ}\text{C}$	$U=(0.002\sim 0.005)^{\circ}\text{C}$		
				Volatility: $(-100\sim 300)^{\circ}\text{C}$	$U=0.003^{\circ}\text{C}$		
37	*Auto-measuring system of thermocouples and resistance thermometers	Temperature	Calibration Specification for Auto-measuring System of Thermocouples and Resistance Thermometers JJF 1098	Industry platinum resistance: 0°C	$U=0.05^{\circ}\text{C}$		
				Industry platinum resistance: 100°C	$U=0.07^{\circ}\text{C}$		
				Base metal thermocouples: $(300\sim 1200)^{\circ}\text{C}$	$U=(0.9\sim 1.1)^{\circ}\text{C}$		



No. CNAS L0893

第 33 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				Working noble metal thermocouples 419.527°C	$U=0.70^{\circ}\text{C}$		
				Working noble metal thermocouples 660.323°C	$U=0.72^{\circ}\text{C}$		
				Working noble metal thermocouples 1084.62°C	$U=0.80^{\circ}\text{C}$		
				Standard thermocouples 419.527°C	$U=0.50^{\circ}\text{C}$		
				Standard thermocouples 660.323°C	$U=0.60^{\circ}\text{C}$		
				Standard thermocouples 1084.62°C	$U=0.60^{\circ}\text{C}$		
38	*Temperature block calibrators	Temperature	Calibration guideline of the temperature block calibrators JJF 1257	$(-80\sim 660)^{\circ}\text{C}$	$U=(0.03\sim 0.36)^{\circ}\text{C}$		
				$(660\sim 1300)^{\circ}\text{C}$	$U=(0.5\sim 1.2)^{\circ}\text{C}$		
39	*Fixed-Point Devices for Standard Platinum Resistance Thermometer	Temperature	Calibration specification of fixed-point devices for standard platinum resistance thermometer JJF 1178	83.8058K	$U=2.0\text{mK}$		
				234.3156K	$U=1.2\text{mK}$		
				273.16K	$U=0.35\text{mK}$		
				302.9146K	$U=1.2\text{mK}$		
				505.078K	$U=1.7\text{mK}$		
				692.677K	$U=2.0\text{mK}$		
40	*The Triple Point of Water Cell (Cell comparison)	Temperature	Calibration Specification of The Triple Point of Water Cell (Cell comparison) NIMTT(CM) 017	273.16K	$U=0.30\text{mK}$		



No. CNAS L0893

第 34 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
41	*Box-type Resistance Furnace	Temperature	Calibration Specification for Box-type Resistance Furnace JJF 1376	(300~1100)°C	$U=(1.2\sim1.5)^{\circ}\text{C}$		
42	Temperature Data Acquisition Instruments	Temperature	Calibration Specification of Temperature Data Acquisition Instruments JJF 1366	(-80~300) °C	$U=0.03^{\circ}\text{C}$		
43	*Thermometers of Clinic Autoclave	Temperature	Calibration Specification for Thermometers of Clinic Autoclave JJF 1308	(25~140)°C	$U=0.4^{\circ}\text{C}$		
		Pressure		(101~400)kPa	$U=2.8\text{kPa}$		
44	*Digital Thermometer	Temperature	Calibration Specification for Digital Thermometer JJF (Chuan) 139	(-196~660.323)°C	$U=(0.010\sim0.018)^{\circ}\text{C}$		
				(660.323~1200) °C	$U=1.2^{\circ}\text{C}$		
45	Sheathed Thermocouples	Temperature	Calibration specification for sheathed thermocouples JJF 1262	K、N: (-40~1100)°C	$U=(0.26\sim1.2)^{\circ}\text{C}$		
				E: (-40~900)°C	$U=(0.26\sim0.80)^{\circ}\text{C}$		
				J: (-40~750)°C	$U=(0.26\sim0.70)^{\circ}\text{C}$		
46	Device of verification and measurement of temperature meter	Temperature	Calibration specification of temperature indicators and simulators by electrical simulation and measurement JJF 1309	RTD: (-200~850)°C	$U=(0.01\sim0.03)^{\circ}\text{C}$		
				K、N、J: (-200~1200) °C	$U=(0.08\sim0.05)^{\circ}\text{C}$		
				S: (0~1600)°C	$U=(0.09\sim0.06)^{\circ}\text{C}$		
				E: (-200~1000)°C	$U=(0.08\sim0.05)^{\circ}\text{C}$		
				T: (-200~400)°C	$U=0.05^{\circ}\text{C}$		
				B: (800~1800)°C	$U=(0.08\sim0.06)^{\circ}\text{C}$		



No. CNAS L0893

第 35 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
47	Temperature Switches	Temperature	Calibration Specification for Temperature Parameters of Temperature Switches JJF1632	$(-30\sim 300)^\circ\text{C}$	$U=0.4^\circ\text{C}$		
48	*Microwave Digestion Instrument	Temperature	Calibration Specification for Temperature Parameters of Microwave Digestion Instrument JJF(Chuan)142	$(0\sim 170)^\circ\text{C}$	$U=0.15^\circ\text{C}\sim 0.25^\circ\text{C}$		
49	*Temperature Indicators	Temperature	Calibration Specification for Temperature Indicators JJF1664	$(-200\sim 1800)^\circ\text{C}$	$TCU=0.38^\circ\text{C}$, $RTDU=0.16^\circ\text{C}$, $DCU=0.12^\circ\text{C}$		
50	Digital Temperature-hygrometer	Humidity	C. S. for Digital Temperature-hygrometers JJF 1076	10%RH~95%RH	$U=0.6\%\text{RH}$		
		Temperature		$(5\sim 50)^\circ\text{C}$	$U=0.1^\circ\text{C}$		
51	Heat flow meter	Heat flux	Calibration Specification of Contact Heat Flow Meters of Absolute method NIMTT(CM) 157	$(2\sim 200)\text{ W/m}^2$	$U_{\text{rel}}=2.8\%$		
52	*Thermal imagers	Temperature	Calibration Specification for Thermal Imagers JJF 1187	$(-20\sim 2000)^\circ\text{C}$	$U=(0.3\sim 4.6)^\circ\text{C}$		
53	*Infrared Thermometers for Measurement of Human Temperature	Temperature	Calibration Specification of Infrared Thermometers for Measurement of Human Temperature JJF 1107	$(22.0\sim 42.0)^\circ\text{C}$	$U=0.11^\circ\text{C}$		
54	Infrared Ear Thermometer	Temperature	Verification Regulation of Infrared Ear Thermometer JJG 1164	$(35.0\sim 42.0)^\circ\text{C}$	$U=0.04^\circ\text{C}$		
55	*Blackbody radiators in $-50^\circ\text{C}\sim +90^\circ\text{C}$	Temperature	Calibration Specification of Blackbody Radiators in $-50^\circ\text{C}\sim +90^\circ\text{C}$ JJF 1080	$(-50\sim +90)^\circ\text{C}$	$U=(0.15\sim 0.20)^\circ\text{C}$		
		Emissivity		0.5~1.0	$U_{\text{rel}}=0.6\%$		



No. CNAS L0893

第 36 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
56	*Blackbody radiators at the 500K~1000K	Temperature	Verification Regulation of Blackbody Radiators at the 500K~1000K JJG 309	(500~1000)K	$U=(0.4\sim1.3)K$		
		Emissivity		0.5~1.0	$U_{rel}=0.6\%$		
57	*Blackbody radiators	Temperature	Calibration specification of Blackbody Radiators NIMTT(CM) 112	(-50~3000) °C	$U=(0.07\sim5.9) ^\circ C$		
		Emissivity		0.5~1.0 (Range:-50°C~1600°C)	$U_{rel}=0.6\%$		
58	*Blackbody radiation sources of radiation thermometry from -10°C to 200°C	Temperature	Calibration Specification for Blackbody Radiation Sources of Radiation Thermometry from -10°C to 200°C JJF 1552	(-10~200) °C	$U=(0.07\sim0.4) ^\circ C$		
59	*Blackbody radiators for radiation thermometry calibration from -50°C~1000°C	Temperature	Verification Regulation of Reference Blackbody Radiator for Radiation Thermometry from -50°C to 1000°C JJG(JUNGONG)162	(-50~1000)°C	$U=(0.07\sim1.5) ^\circ C$		
60	*Radiation temperature of 800 °C~3000°C blackbody radiators	Temperature	Verification Regulation of Reference Blackbody Radiators for Radiation Thermometry from 800°C to 3000°C JJG(JUNGONG)161	(800~3000) °C	$U=(1.2\sim5.9) ^\circ C$		
61	*Radiation thermometers	Temperature	Verification Regulation of Radiation Thermometers JJG 856	(-50~3000)°C	$U=(0.04\sim6.7)^\circ C$		
62	Standard tungsten ribbon lamps	Temperature	Verification Regulation of Standard Tungsten Ribbon Lamps JJG 110	(800~2000)°C	$U=(1.5\sim2.8)^\circ C$		



在线扫码获取验证

No. CNAS L0893

第 37 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
63	Standard photoelectric pyrometers	Temperature	Verification Regulation of Standard Photoelectric Pyrometers JJG 1032	(800~3200)°C	$U=(0.9\sim5.9)^{\circ}\text{C}$		
III Mechanics measuring instrument							
1	Elevator Overspeed Governor	Speed	Calibration Specification for Elevator Overspeed Governor Testers JJF 1374	(0.1~10.0) m/s	$U_{\text{rel}}=0.2\%$		
2	*Speed and mileage meter for motor vehicles	mileage	Verification Regulation of Speed and mileage meter for non-disintegrated cars JJG 750	(1~999)km	$U_{\text{rel}}=1.0\%$		
		speed		(1~120) km/h	$U_{\text{rel}}=3.2\%$		
3	Test Equipment for Vehicle Speed Radar Measurement Meters	Frequency	Verification Regulation of Test Equipment for Vehicle Speed Radar Measurement Meters JJG 771	(1~40000) Hz	$U=0.56\text{Hz}$		
4	Verification Regulation of Standard Equipment for Taximeter	rotational speed	Verification Regulation of Standard Equipment for Taximeter JJG 738	(50~10000)r/min	$U_{\text{rel}}=0.12\%$		
		rotational number		(1~999999)r	$U_{\text{rel}}=0.03\%$		
		speed		40km/h	$U_{\text{rel}}=0.4\%$		
		noise		(1~500)mm	$U_{\text{rel}}=0.1\%$		
5	*Point-to-point Vehicle Speed Monitor Systems for Road Traffic	Mileage	Point-to-point Vehicle Speed Monitor Systems for Road Traffic JJF(Chuan) 109	(1~60000)m	$U_{\text{rel}}=0.8\%$		
		Time Interval		(1~99999) s	$U_{\text{rel}}=0.5\%$		
6	*Vehicles radar measuring speedometers	Simulation speed	Mobile Radar Vehicle Speed Measurement Device JJG 528, Fixed Radar Vehicle	(0~180) km/h	$U=0.7\text{km/h}$		
		Field speed		(20~100) km/h	$U=1.5\text{km/h}$		



No. CNAS L0893

第 38 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
			Speed Measurement Device JJG 527	($>100\sim180$) km/h	$U=2.3$ km/h		
		Microwave emission frequency		($0\sim40$)GHz	$U=2.0$ MHz		
7	Speed and mileage meter for motor vehicles	speed	Speed and mileage meter for motor vehicles JJG 559	($20\sim240$)km/h	$U=1.5$ km/h		
		mileage		($3\sim1000$)km	$U=0.01$ km		
8	Standard Equipment for Revolution Speed	revolution speed	Verification Regulation of Standard Equipment for Revolution Speed JJG 326	($30\sim40000$)r/min	$U_{rel}=0.01\%(k=3)$		
9	Tachometer	revolution speed	Verification Regulation of Tachometer JJG 105	($20\sim33000$)r/min	$U_{rel}=0.02\%$		
10	*Mixer	revolution speed	Calibration Specification for Mixer for cement paste JJF(Building materials) 104	($0\sim300$)r/min	$U=1.7$ r/min		
		Time		($0\sim6000$)s	$U=0.33$ s		
		Length		($0\sim300$)mm	$U=0.33$ mm		
11	*Weights	mass	Verification Regulation of Weights JJG 99	$E_1:1\text{mg}\sim500\text{mg}$	$U=0.0004\text{mg}\sim0.0006\text{mg}$		
				$E_1:1\text{g}\sim500\text{g}$	$U=0.0008\text{mg}\sim0.02\text{mg}$		
				$E_1:1\text{kg}\sim20\text{kg}$	$U=0.04\text{mg}\sim1.1\text{mg}$		
				$E_2:1\text{mg}\sim500\text{mg}$	$U=0.0006\text{mg}\sim0.0008\text{mg}$		
				$E_2:1\text{g}\sim500\text{g}$	$U=0.0012\text{mg}\sim0.024\text{mg}$		
				$E_2:1\text{kg}\sim20\text{kg}$	$U=0.04\text{mg}\sim1.5\text{mg}$		
				$F_1:1\text{mg}\sim500\text{mg}$	$U=0.001\text{mg}\sim0.0012\text{mg}$		
				$F_1:1\text{g}\sim500\text{g}$	$U=0.0016\text{mg}\sim0.026\text{mg}$		



No. CNAS L0893

第 39 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				F ₁ : 1kg~20kg	U=0.04mg~1.7mg		
				F ₂ : 1mg~500mg	U=0.001mg~0.0012mg		
				F ₂ : 1g~500g	U=0.0016mg~0.026mg		
				F ₂ : 1kg~500kg	U=0.04mg~1.2g		
				M ₁ : 1mg~500mg	U=0.001mg~0.0012mg		
				M ₁ : 1g~500g	U=0.0084mg~0.44mg		
				M ₁ : 1kg~20kg	U=0.44mg~1.9mg		
				M ₂ : 100m~500mg	U=0.05mg~0.06mg		
				M ₂ : 1g~500g	U=0.08mg~1.0mg		
				M ₂ : 1kg~20kg	U=2.6mg~50mg		
				M ₃ : 1g~500g	U=0.46mg~2.6mg		
				M ₃ : 1kg~20kg	U=5.0mg~0.1g		
12	*Mechanical Balance	Quality	Verification Regulation of Mechanical Balance JJG 98	Max:(0~20)g d=0.01mg	U=0.08mg		
				Max:(0~200)g d=0.1mg	U=0.3mg		
				Max:(0~20)kg d>0.1mg	U=0.72mg~42mg		
13	*Electronic Balance	Quality	Verification Regulation of Electronic Balance JJG 1036	Max:(0~500) g e=0.001mg~0.1mg,(0~ 5×10 ⁴)e	U=4.4 μg~0.15mg		
				Max:(0~500) g e=0.001mg~0.1mg,(5× 10 ⁴ ~2×10 ⁵)e	U=11 μg~0.18mg		



No. CNAS L0893

第 40 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				Max:(0~500) g $e=0.001\text{mg}\sim 0.1\text{mg}, (>2 \times 10^5)e$	$U=7.4 \mu\text{g}\sim 0.19\text{mg}$		
				Max:500g~5kg $e=1\text{mg}\sim 10\text{mg}, (0\sim 5 \times 10^4)e$	$U=0.44\text{mg}\sim 0.62\text{mg}$		
				Max:500g~5kg $e=1\text{mg}\sim 10\text{mg} (5 \times 10^4 \sim 2 \times 10^5)e$	$U=0.57\text{mg}\sim 16\text{mg}$		
				Max:500g~5 kg $e=1\text{mg}\sim 10\text{mg} (>2 \times 10^5)e$	$U=1.5\text{mg}\sim 18\text{mg}$		
				Max:(5~20) kg $e=20\text{mg}\sim 50\text{mg} (0\sim 5 \times 10^4)e$	$U=1.4\text{mg}\sim 2.0\text{mg}$		
				Max:(5~20) kg $e=20\text{mg}\sim 50\text{mg} (5 \times 10^4 \sim 2 \times 10^5)e$	$U=2.0\text{mg}\sim 2.4\text{mg}$		
				Max:(5~20) kg $e=20\text{mg}\sim 50\text{mg} (>2 \times 10^5)e$	$U=2.5\text{mg}\sim 32\text{mg}$		
				Max:(20~100) kg $(e=0.1\text{g}\sim 0.2\text{g}) (0\sim 5 \times 10^4)e$	$U=(0.0047\sim 0.046)\text{g}$		
				Max:(20~100) kg $(e=0.1\text{g}\sim 0.2\text{g}) (5 \times 10^4 \sim 2 \times 10^5)e$	$U=(0.0061\sim 0.18)\text{g}$		




No. CNAS L0893

第 41 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				Max:(20~100) kg ($e=0.1\text{g}\sim 0.2\text{g}$) ($>2\times 10^5$)e	$U=(0.015\sim 0.18)\text{g}$		
				Max:(100~3000) kg ($e=1\text{g}\sim 10\text{g}$) ($0\sim 5\times 10^4$)e	$U=(0.20\sim 1.8)\text{g}$		
				Max:(100~3000) kg ($e=1\text{g}\sim 10\text{g}$) ($5\times 10^4\sim 2\times 10^5$)e	$U=(0.21\sim 2.2)\text{g}$		
				Max:(100~3000) kg ($e=1\text{g}\sim 10\text{g}$) ($>2\times 10^5$)e	$U=(0.32\sim 2.4)\text{g}$		
14	*Calibration Specification for Mass Comparators	Quality	Calibration Specification for Mass Comparators JJF1326	Max:(0~3000) kg $d=0.1\mu\text{g}\sim 50\text{g}$	$U=0.002\text{mg}\sim 11\text{g}$		
15	*Relative Density Balance for Liquid	Quality	Verification Regulation of Relative Density Balance for Liquid JJG 171	0.001~2.0000	$U_{\text{rel}}=0.043\%$		
16	*Table Balances	Quality	Verification Regulation of Table Balances JJG 156	(0~20)kg	$U=(0.1\sim 10)\text{g}$		
17	*Non-self-indicating Weighing Instruments	Quality	Verification Regulation of Non-self-indicating Weighing Instruments JJG 14	Max:(0~100) kg ($e=5\text{g}\sim 50\text{g}$) (0~500)e	$U=0.69\text{g}\sim 7.0\text{g}$		
				Max:(0~100) kg ($e=5\text{g}\sim 50\text{g}$) (500~2000)e	$U=0.77\text{g}\sim 9.0\text{g}$		
				Max:(0~100) kg ($e=5\text{g}\sim 50\text{g}$) (2000~10000)e	$U=0.8\text{g}\sim 8.0\text{g}$		
				Max:(100~1000) kg ($e=50\text{g}\sim 500\text{g}$) (0~500)e	$U=7.0\text{g}\sim 70\text{g}$		



No. CNAS L0893

第 42 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				Max:(100~1000) kg ($e=50g\sim 500g$) (500~2000)e	$U=9.0g\sim 89g$		
				Max:(100~1000) kg ($e=50g\sim 500g$) (2000~10000)e	$U=9.0g\sim 92g$		
				Max:(1~30) t ($e=1kg\sim 10kg$) (0~500)e	$U=90g\sim 0.21kg$		
				Max:(1~30) t ($e=1kg\sim 10kg$) (500~2000)e	$U=92g\sim 0.34kg$		
				Max:(1~30) t ($e=1kg\sim 10kg$) (2000~10000)e	$U=95g\sim 0.41kg$		
18	*Digital Indicating Weighing Instruments	Quality	Verification Regulation of Digital Indicating Weighing Instruments JJG 539	Max:(0~30) kg $e=1g\sim 10g$ (0~500)e	$U=0.06g\sim 0.6g$		
				Max:(0~30) kg $e=1g\sim 10g$ (500~2000)e	$U=(0.06\sim 1.2)g$		
				Max:(0~30) kg $e=1g\sim 10g$ (2000~10000)e	$U=(0.15\sim 2.0)g$		
				Max:(30~1000) kg $e=20g\sim 500g$ (0~500)e	$U=1.4g\sim 31g$		
				Max:(30~1000) kg $e=20g\sim 500g$ (500~2000)e	$U=1.4g\sim 54g$		
				Max:(30~1000) kg $e=20g\sim 500g$ (2000~10000)e	$U=(1.8\sim 57)g$		
				Max:(1~30) t $e=1kg\sim 10kg$ (0~500)e	$U=(0.07\sim 1.2)kg$		



No. CNAS L0893

第 43 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
19	*Analogue Indicating Weighing Instruments	Quality	Verification Regulation of Analogue Indicating Weighing Instruments JJG 13	Max:(1~30)t e=1kg~10kg (500~2000)e	$U=(0.07\sim1.4)\text{kg}$		
				Max:(1~30) t e=1kg~10kg (2000~10000)e	$U=(0.10\sim2.0)\text{kg}$		
				Max:(30~150) t e=10kg~50kg (0~500)e	$U=(1.0\sim4.2)\text{kg}$		
				Max:(30~150) t e=10kg~50kg (500~2000)e	$U=(1.2\sim5.3)\text{kg}$		
				Max:(30~150) t e=10kg~50kg (2000~10000)e	$U=(2.1\sim6.4)\text{kg}$		
20	*Automatic Gravimetric Filling Instruments	Quality	Verification Regulation of Automatic Gravimetric Filling Instruments JJG 564	Max:(0~10) kg e=10g~50g (0~50)e	$U=(1.0\sim5.7)\text{g}$		
				Max:(0~10) kg e=10g~50g (50~200)e	$U=(1.3\sim6.3)\text{g}$		
				Max:(0~10) kg e=10g~50g (200~1000)e	$U=(1.4\sim6.4)\text{g}$		
				Max:(10~120) kg e=100g~500g (0~50)e	$U=(10\sim53)\text{g}$		
				Max:(10~120) kg e=100g~500g (50~200)e	$U=(14\sim61)\text{g}$		
				Max:(10~120) kg e=100g~500g (200~1000)e	$U=(16\sim65)\text{g}$		
20	*Automatic Gravimetric Filling Instruments	Quality	Verification Regulation of Automatic Gravimetric Filling Instruments JJG 564	10g~100t	$U_{\text{rel}}=0.09\%$		



No. CNAS L0893

第 44 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
21	*Continuous Totalizing Automatic Weighing Instruments (Belt Weigher)	Quality	Verification Regulation of Continuous Totalizing Automatic Weighing Instruments (Belt Weigher) JJG 195	(0.2~1000)t/h	$U_{rel}=0.09\%$		
22	*Discontinuous Totalizing Automatic Weighing Instruments	Quality	Verification Regulation of Discontinuous Totalizing Automatic Weighing Instruments (Totalizing Hopper Weighers) JJG 648	0.1t~1t	$U_{rel}=0.08\%$		
				>1t~10t	$U_{rel}=0.05\%$		
				>10t~200t	$U_{rel}=0.07\%$		
23	*Automatic Instruments for Weighing Road Vehicles in Motion	Quality	Verification Regulation of Automatic Instruments for Weighing Road Vehicles in Motion JJG 907	(1~150)t	$U_{rel}=0.36\%$		
24	Digital Weighing Indicators (Weighing Indicators)	Quality	Verification Regulation of Digital Weighing Indicators (Weighing Indicators) JJG 649	n:100~10000	$U=0.5d\sim 1d$		
25	*Taking Blood Electronic Scale	Quality	Verification Regulation of Taking Blood Electronic Scale JJG 815	Max:(0~600) g d=1g (0~50)d	$U=0.12g$		
				Max:(0~600) g d=1g (50~200)d	$U=0.14g$		
				Max:(0~600) g d=1g (>200)d	$U=0.26g$		
				Max:(0~1000) g d=2g (0~50)d	$U=0.24g$		
				Max:(0~1000) g d=2g (50~200)d	$U=0.26g$		



No. CNAS L0893

第 45 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				Max:(0~1000) g $d=2g$ (>200)d	$U=0.48g$		
				Dynamic:200g	$U_{rel}=0.1\%$		
				Dynamic:400g	$U_{rel}=0.1\%$		
26	*Steelyard Scales	Quality	Verification Regulation of Steelyard Scale JJG 17	Max:(0~500) g $e=(1\sim 2)g$	$U=(0.14\sim 0.28)g$		
27	Load Cell	mass	Verification Regulation of Load Cell JJG 669	(10g~1kg)	$U_{rel}=0.01\%$		
				(1kg~1t)	$U_{rel}=0.01\%$		
				(1t~100t)	$U_{rel}=0.03\%$		
				(100~200)t	$U_{rel}=0.05\%$		
28	Interface Tensiometers	Force value	C. S. for Interface Tensiometers JJF 1464	(0.1~100)N	$U_{rel}=0.2\%$		
		Length		(0.1~25)mm	$U=(1+L/100) \mu m, (L: mm)$		
29	Dynami e Force Sensors	Dynami e Force	V. R. of Dynamic Force Sensors JJG 632	200N~10kN	$U_{rel}=5.0\%$		
30	Working Force Measuring Machines for Special Purposes	Force value	C. S. for Working Force Measuring Machines for Special Purposes JJF 1134	(0.5~15)N	$U_{rel}=0.1\%$		
31	Standard Dynamometers	Force value	V. R. of Standard Dynamometers JJG 144	0.1N~10kN	$U_{rel}=0.01\%$		
				(10~3600)kN	$U_{rel}=0.03\%$		
32	Force Transducers	Force value	V. R. of Force Transducers JJG 391	0.1N~10kN	$U_{rel}=0.01\%$		
				(10~3600)kN	$U_{rel}=0.03\%$		
33	Working	Force value	V. R. of Working	0.1N~10kN	$U_{rel}=0.01\%$		



No. CNAS L0893

第 46 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	Dynamometers		Dynamometers JJG 455	(10~3600)kN	$U_{rel}=0.03\%$		
34	*Deadweight and Lever Force Standard Machines	Force value	V. R. of Force Standard Machines JJG 734	10N~1MN	$U_{rel}=0.003\%$		
35	*Hydraulic-amplification Force Standard Machines	Force value	V. R. of Hydraulic-amplification Force Standard Machines JJG 1117	10N~1MN (>1~3)MN	$U_{rel}=0.003\%$ $U_{rel}=0.01\%$		
36	*Build-up Force Standard Machines	Force value	V. R. of Build-up Force Standard Machines JJG 1116	10N~1MN (>1~3)MN (>3~3.6)MN (>3.6~10)MN (>10~30)MN	$U_{rel}=0.003\%$ $U_{rel}=0.01\%$ $U_{rel}=0.03\%$ $U_{rel}=0.1\%$ $U_{rel}=0.3\%$		
37	Hydraulic Jacks	Force value	V. R. of Hydraulic Jacks JJG 621	10N~3.6MN	$U_{rel}=0.5\%$		
38	Measuring Instrumentations for Strain Gauge Transducer	Voltage ratio	C. S. for Measuring Instrumentations for Strain Gauge Transducer JJF 1469	$\pm(0.1\sim100)\text{mV/V}$	$U_{rel}=0.0005\%$		
39	*Tension, Compression and Universal Testing Machines	Force value	V. R. of Tension, Compression and Universal Testing Machines	0.1N~300kN (>0.3~30)MN	$U_{rel}=0.2\%$ $U_{rel}=0.3\%$		
		Coaxiality	JJG 139, Metallic materials — Calibration and verification of static uniaxial testing machines — Part 1:	0.1%~40%	$U=2\%$		
		Displacement		(0.05~1)mm (1~500)mm	$U=1\mu\text{m}$ $U_{rel}=0.1\%$		



No. CNAS L0893

第 47 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Deformation	Tension/compression testing	(0.1~1)mm, (Extensometer)	U=1μm		

CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT
SCHEDULE OF ACCREDITATION CERTIFICATE



在线扫码获取验证

No. CNAS L0893

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
			machines — Calibration and verification of the force-measuring system ISO 7500-1, Metallic materials — Calibration and verification of static uniaxial testing machines -- Part 1: Tension/compression testing machines — Calibration and verification of the force-measuring system BS EN ISO 7500-1, Metallic materials -- Verification of static uniaxial testing machines -- Part 2: Tension creep testing machines -- Verification of the applied force ISO 7500-2, Metallic materials -- Verification of static uniaxial testing machines -- Part 2: Tension creep testing machines -- Verification of the applied force DIN EN ISO 7500-2, Standard Practices for Force Verification of Testing Machines ASTM E4-16, Standard Practice for Verification of Testing Frame and Specimen Alignment Under Tensile and Compressive Axial Force	(1~300)mm, (Extensometer)	$U_{rel}=0.1\%$		




No. CNAS L0893

Application ASTM E1012-14, Standard Practices for Verification of Displacement

The scope of the accreditation in Chinese remains the definitive version.

第 49 页 共 181 页



№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
40	*Flexure Testing Machines	Force value	V. R. of Flexure Testing Machines JJG 476	0.1N~300kN	$U_{rel}=0.2\%$		
				(>0.3~30)MN	$U_{rel}=0.3\%$		
41	*Electronic Universal Testing Machine	Force value	V. R. of Electronic Universal Testing Machine JJG 475, Metallic materials —	0.1N~300kN	$U_{rel}=0.2\%$		
			Calibration and verification of static uniaxial testing machines — Part 1:	(>0.3~30)MN	$U_{rel}=0.3\%$		
		Coaxality	Tension/compression testing machines — Calibration and verification of the force-measuring system ISO 7500-1, Metallic materials —	0.1%~40%	$U=2\%$		
		Displacement		(0.05~1)mm	$U=1\mu m$		
		Speed		(1~500)mm	$U_{rel}=0.1\%$		
		Deformation		(0.05~500)mm/min	$U_{rel}=0.2\%$		
				(0.1~1)mm, (Extensometer)	$U=1\mu m$		



No. CNAS L0893

第 50 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			Calibration and verification of static uniaxial testing machines — Part 1: Tension/compression testing machines — Calibration and verification of the force-measuring system BS EN ISO 7500-1, Metallic materials -- Verification of static uniaxial testing machines -- Part 2: Tension creep testing machines -- Verification of the applied force ISO 7500-2, Metallic materials -- Verification of static uniaxial testing machines -- Part 2: Tension creep testing machines -- Verification of the applied force DIN EN ISO 7500-2, Standard Practices for Force Verification of Testing Machines ASTM E4-16, Standard Practice for Verification of Testing Frame and Specimen Alignment Under Tensile and Compressive Axial Force Application ASTM E1012-14, Standard Practices for Verification of Displacement Measuring Systems and Devices Used in Material Testing Machines ASTM E2309/E2309M-16, Standard	(1~300)mm, (Extensometer)	$U_{rel}=0.1\%$		

No. CNAS L0893

第 51 页 共 181 页



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
42	*High-Temperature Creep and Stress-Rupture Testing Machines	Force value	V. R. of High-Temperature Creep and Stress-Rupture Testing Machines JJG	0.1N~300kN	$U_{rel}=0.2\%$		
				(>0.3~30)MN	$U_{rel}=0.3\%$		
		Coaxiality	276,Metallic materials — Calibration and verification of static uniaxial testing	0.1%~40%	$U=2\%$		
		Time	machines — Part 1:	(0~86400)s	$U=0.1s$		
		Deformatin	Tension/compression testing	(0.1~1)mm, (Extensometer)	$U=1\mu m$		



No. CNAS L0893

第 52 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		machines — Calibration and verification of the force-measuring system ISO 7500-1, Metallic materials — Calibration and verification of static uniaxial testing machines — Part 1: Tension/compression testing machines — Calibration and verification of the force-measuring system BS EN ISO 7500-1, Metallic materials -- Verification of static uniaxial testing machines -- Part 2: Tension creep testing machines -- Verification of the applied force ISO 7500-2, Metallic materials -- Verification of static uniaxial testing machines -- Part 2: Tension creep testing machines -- Verification of the applied force DIN EN ISO 7500-2, Standard Practices for Force Verification of Testing Machines ASTM E4-16, Standard Practices for Verification of Displacement Measuring Systems and Devices Used in Material Testing Machines ASTM		(1~300)mm, (Extensometer)	$U_{rel}=0.1\%$		

No. CNAS L0893

E2309/E2309M-16, Standard Practice for Verification of Testing Frame and Specimen

The scope of the accreditation in Chinese remains the definitive version.

第 53 页 共 181 页



在线扫码获取验证

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
43	*Electro-hydraulic Servo Universal Testing Machines	Force value	V. R. of Electro-hydraulic Servo Universal Testing Machines JJG 1063	10N~3MN	$U_{rel}=0.2\%$		
44	*Working Force Measuring Machines for Special Purposes	Force value	C. S. for Working Force Measuring Machines for Special Purposes JJF 1134	0.1N~300kN ($>0.3\sim30$)MN	$U_{rel}=0.2\%$ $U_{rel}=0.3\%$		
45	*Pendulum Impact Testing Machines	Length	V. R. of Pendulum Impact Testing Machines JJG 145	(0~800)mm, (Center distance)	$U=0.33$ mm		
		Energy		(0.1~300)J	$U_{rel}=0.33\%$		
46	Calibration Instrument for Torque Wrenches	Torque	V. R. of Calibration Instrument for Torque Wrenches JJG 797	(0.1~1)Nm ($>1\sim3000$)Nm	$U_{rel}=0.12\%$ $U_{rel}=0.1\%$		
47	Torque Wrenches	Torque	V. R. of Torque Wrenches JJG 707	(0.4~3000)Nm	$U_{rel}=0.3\%$		
48	*Electric and Pneumatic Torque Wrenches	Torque	C. S. for Electric and Pneumatic Torque Wrenches JJF 1610	(2~2500)Nm	$U_{rel}=0.3\%$		
49	*Torsion Testing Machines	Torque	V. R. of Torsion Testing Machines JJG 269	(2~10000)Nm	$U_{rel}=0.05\%$		
50	Static Torque Measuring Devices	Torque	V. R. of Static Torque Measuring Devices JJG 995	(0.1~50)Nm ($>50\sim5000$)Nm	$U_{rel}=0.12\%$ $U_{rel}=0.06\%$		
51	Tacho-Torque Measuring Device	Torque	V. R. of Tacho-Torque Measuring Device JJG 924	(0.1~50)Nm ($>50\sim5000$)Nm	$U_{rel}=0.12\%$ $U_{rel}=0.06\%$		
		Rotating velocity		(50~10000)r/min	$U_{rel}=0.12\%$		
52	*Centrifuge	Rotating velocity	Centrifuge NIMTT(CM) 106	(20~33000)r/min	$U_{rel}=0.02\%$		



No. CNAS L0893

第 54 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Temperature		(-20~40)°C	U=0.8°C		
53	Dynamic Elastic Modulus Measurement Instruments	Frequency	C. S. for Dynamic Elastic Modulus Measurement Instruments JJF 1373	(5~5000)Hz	U _{rel} =1.0%		
54	Equipment of Power Measuring	Rotating velocity	V. R. of Equipment of Power Measuring JJG 653	Non-contact: (50~10000)r/min	U _{rel} =0.12%		
		Torque		Non-contact: (1~2500)Nm	U _{rel} =0.2%		
		Driving force		Contact: (1~10 ⁴)N	U _{rel} =0.6%		
		Speed		Contact: (1~200)km/h	U _{rel} =0.2%		
55	Diamond Pressure Head	Angle	Calibration Specification for Diamond Pressure Head NIMTT (CM) 025	(90~140)°	U=2'		
		Length		((0.2~10)mm	U=1.2 μm		
56	*Metallic Rockwell Hardness Testing Machines	Hardness	Verification Regulation of Metallic Rockwell Hardness Testing Machines ((Scales A,B,C,D,E,F,G,H,K,N,T) JJG 112	(20~88)HRA	U=0.6HRA		
				(20~100)HRBW	U=0.6HRBW		
				(20~70)HRC	U=0.4HRC		
				(12~93)HRN/TW	U=(1.0~1.5)HRN/TW		
		Angle		90° ~140°	U=0.25°		
		Length		(0.2~10)mm	U=1.2um		
57	Metallic Rockwell Hardness Reference Blocks	Hardness	Verification Regulation of Metallic Rockwell Hardness Reference Blocks (Scales	(20~88)HRA	U=0.5HRA		
				(20~100)HRBW	U=0.5HRBW		



No. CNAS L0893

第 55 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	locks		A,B,C,D,E,F,G,H,K,N,T) JJG 113	(20~70)HRC	$U=0.5\text{HRC}$		
				(89~91)HR15N	$U=0.64\text{HR15N}$		
				(74~80)HR30N	$U=0.64\text{HR30N}$		
				(42~54)HR30N	$U=0.75\text{HR30N}$		
				(32~61)HR45N	$U=0.80\text{HR45N}$		
				(88~93)HR15TW	$U=0.76\text{HR15TW}$		
				(70~82)HR30TW	$U=0.76\text{HR30TW}$		
58	Brinell Hardness Reference Blocks	Hardness	Verification Regulation of Metallic Brinell Hardness Reference Blocks JJG 147	(8~650)HBW(S)	$U_{\text{rel}}=(0.8\sim1.0)\%$		
59	*Brinell Hardness Testers	Hardness	Verification Regulation of Metallic Brinell Hardness Testers JJG 150	(8~125) HB	$U_{\text{rel}}=1.3\%$		
				(125~225)HB	$U_{\text{rel}}=1.0\%$		
				(225~650)HB	$U_{\text{rel}}=0.9\%$		
60	Vickers Hardness Reference Blocks	Hardness	Verification Regulation of Vickers Hardness Reference Blocks JJG 148	(5~225)HV	$U_{\text{rel}}=2.1\%$		
				(225~400)HV	$U_{\text{rel}}=1.3\%$		
				(400~1000)HV	$U_{\text{rel}}=1.6\%$		
61	*Vickers Hardness Testers	Hardness	Verification Regulation of Metallic Vickers Hardness Testers JJG 151	(5~225)HV	$U_{\text{rel}}=2.7\%$		
				(225~400)HV	$U_{\text{rel}}=2.1\%$		
				(400~1000)HV	$U_{\text{rel}}=2.3\%$		
62	Epuotip Hardness	Hardness	Verification Regulation of Epuotip Hardness Testers JJG	(465~847)HLD	$U=5\text{HLD}$		



No. CNAS L0893

第 56 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	Reference Blocks		747	(465~847)HLD	$U=7\text{HLD}$		
				(448~645)HLG	$U=5\text{HLG}$		
				(448~645)HLG	$U=7\text{HLG}$		
63	ShoreHardness Reference Blocks	Hardness	Verification Regulation of Shore Hardness Reference Blocks JJG 347	(91~99)HSD	$U=1.1\text{HSD}$		
				(56~64)HSD	$U=0.9\text{HSD}$		
				(26~34)HSD	$U=0.8\text{HSD}$		
64	Shore Hardness testers	Hardness	Verification Regulation of Shore Hardness tester JJG 346	(26~99)HSD	$U=2.0\text{HSD}$		
65	Shore A Durometers	Force	Verification Regulation of Shore A Durometers JJG 304	(0~100)HA	$U=1.0\text{HA}$		
		Length		$(0.79 \pm 0.03)\text{mm}$	$U=0.1\text{mm}$		
		Angle		$0^\circ \sim 360^\circ$	$U=0.5^\circ$		
66	Measuring Instrument for Cement Bright Degumming Equipment	Frequency	Verification Regulation of Measuring Instrument for Cement Bright Degumming Equipment JJG 974	(20~100)Hz	$U_{\text{rel}}=1.0\%$		
		Acceleration		$(0.1 \sim 1000)\text{m/s}^2$	$U_{\text{rel}}=3.0\%$		
		Displacement		$(10 \sim 1000)\mu\text{m}$	$U_{\text{rel}}=3.0\%$		
67	Acceleration Sensor	Acceleration	Verification Regulation of Piezoelectrl Accelerometer JJG 233	$(1 \sim 300)\text{m/s}^2, 160\text{Hz}$	$U_{\text{rel}}=1.0\%$		
				$(1 \sim 300)\text{m/s}^2, (10 \sim 2000)\text{Hz}$	$U_{\text{rel}}=2.0\%$		
				shock: $(200 \sim 14000)\text{m/s}^2$	$U_{\text{rel}}=5.0\%$		
68	Measuring Instrument for Shock	Acceleration	Calibration Specification for Shock Measuring Instruments JJF 1943	$(200 \sim 14000)\text{m/s}^2$	$U_{\text{rel}}=5.0\%$		



No. CNAS L0893

第 57 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Time		(1~100) ms	$U_{rel}=3.0\%$		
69	*Calibration Set of Shock Acceleration by Impact Force	Acceleration	Verification Regulation of Calibration Set of Shock Acceleration by Impact Force JJG 791	(0.5~1000)m/s ²	$U_{rel}=3.0\%$		
70	Dynamic Balance Measuring Instruments	Acceleration	Calibration Specification for Dynamic Balance Measuring Instruments JJF 1570	(0.1~100)m/s ² , (1Hz~400Hz)	$U_{rel}=3.0\%$		
		Phase		0° ~360° , (1Hz~400Hz)	$U=3.0^\circ$		
		revolution speed		(30~30000)r/min	$U_{rel}=0.02\%$		
71	*Quantitative Filling Machine for Liquid State Material	Weight	Verification Regulation of Quantitative Filling Machine for Liquid Material JJG 687	Set the heavy type:(0~20)kg	$U=17g$		
72	Electromagnetic Velocity Transducer	Velocity	Verification Regulation of Electromagnetic Velocity Transducer JJG 134	(0.1~50)cm/s, (10~2000)Hz	$U_{rel}=3.0\%$		
73	*Verification regulation for cement Mortar mixer complying	revolution speed	Calibration Specification for Mixer for Mixing Mortars JJF (Building materials) 123	(0~300)r/min	$U=1.7r/min$		
		time		(0~6000)s	$U=0.33s$		
		Length		(0~300)mm	$U=0.33mm$		
74	High Strength Bolt Testers	torque	Calibration Specification for High Strength Bolt Testers JJF 1478	(50~3000) Nm	$U_{rel}=0.2\%$		
75	*Digital Electrodynamics Vibration Testing System	Frequency	Digital Electrodynamics Vibration Testing System JJG 948	(5~5000)Hz	$U_{rel}=1.0\%$		
		Acceleration		(0.1~1000)m/s ²	$U_{rel}=5.0\%$		



No. CNAS L0893

第 58 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Displacement		(10~1000) μm	$U_{\text{rel}}=5.0\%$		
76	*Cement mortar specimen jolting table	Time	C.S.for cement mortar specimen jolting table	(0~6000)s	$U=0.33\text{s}$		
		displacement	compiying with ISO 679 JJF(Building materials) 124	(0~300)mm	$U=0.33\text{mm}$		
77	Verification Regulation of Vibration meters	Frequency	Verification Regulation of Vibration meters JJG 676	(10~2000)Hz	$U_{\text{rel}}=0.2\%$		
		Acceleration		(0.1~300) m/s^2	$U_{\text{rel}}=2.0\%$		
		Velocity		(0.1~50)cm/s	$U_{\text{rel}}=2.0\%$		
		Displacement		(0.01~10)mm	$U_{\text{rel}}=2.0\%$		
78	*Verification Regulation of Mechanical Vibration Genetator for Testing	Frequency	Verification Regulation of Mechanical Vibration Genetator for Testing JJG 189	(5~5000)Hz	$U_{\text{rel}}=1.0\%$		
		Acceleration		(0.1~1000) m/s^2	$U_{\text{rel}}=5.0\%$		
		Displacement		(10~1000) μm	$U_{\text{rel}}=5.0\%$		
79	*Standard Vibrators	Frequency	Verification Regulation of Standard Vibrators JJG 298	(0.1~5000)Hz	$U_{\text{rel}}=1.0\%$		
		Acceleration		(0.1~1000) m/s^2	$U_{\text{rel}}=3.0\%$		
		Displacement		(10~1000) μm	$U_{\text{rel}}=3.0\%$		
80	*Hydraulic Vibration Testing System	Frequency	Hydraulic Vibration Testing System JJG 638	(0.1~5000)Hz	$U_{\text{rel}}=1.0\%$		
		Acceleration		(0.1~1000) m/s^2	$U_{\text{rel}}=5.0\%$		
		Displacement		(10~1000) μm	$U_{\text{rel}}=5.0\%$		



No. CNAS L0893

第 59 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
81	Vibration Displacement Transducer	Displacement	Vibration Displacement Transducer JJG 644	Dynamic: (0.01~10)mm, (10~2000)Hz	<i>U</i> _{rel} =3.0%		
				Static: (0.01~300)mm	<i>U</i> _{rel} =0.1%		
82	Charge Amplifiers	Charge	Charge Amplifiers JJG 338	(0.1~10 ⁴)pC	<i>U</i> _{rel} =0.05%		
83	Resistance Strain Gauge Indicators	Frequency	Resistance Strain Gauge Indicators JJG 623	10Hz~500kHz	<i>U</i> =0.5dB		
		Strain		(0.1~10 ⁵) μ ε	<i>U</i> _{rel} =0.1%		
84	Calibration Specification for IEPE Amplifiers	Voltage	Calibration Specification for IEPE Amplifiers JJF 1269	(0.01~10)V	<i>U</i> _{rel} =0.05%		
85	Dynamical Signal Analyzer	Frequency	Dynamical Signal Analyzer JJG 834	(1~200k)Hz	<i>U</i> _{rel} =1×10 ⁻⁴		
		Voltage		(0.02~10)V	<i>U</i> _{rel} =0.2%		
IV Acoustics measuring instrument							
1	Standard Sound Source	Sound Power Level	V. R. of Reference Sound Sources JJG 277	(36~148)dB, (100~500) Hz	<i>U</i> =1.0dB		
				(36~148)dB, (>500~2500) Hz	<i>U</i> =0.5dB		
				(36~148)dB, (>2.5~10) kHz	<i>U</i> =0.8dB		
2	Sound Calibrators	SPL	V.R. of Sound Calibrators JJG 176	(30~130)dB, (31.5Hz~16kHz)	<i>U</i> =0.50dB		
				94dB, 1kHz	<i>U</i> =0.12dB		
				114dB, 1kHz	<i>U</i> =0.15dB		
				124dB, 250Hz	<i>U</i> =0.10dB		



No. CNAS L0893

第 60 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		Frequency		20Hz～20kHz	<i>U</i> _{rel} =0.01%		
		Distortion		0.05%～100%	<i>U</i> =10%rdg		
3	Working Standard Microphoes	Frequency response of Sensitivitylevel	V.R. of Working Standard Microphones(Electrostatic Actuator Method) JJG 175,V.R. of Working Standard	(-60～20)dB, 20Hz～20kHz	<i>U</i> =0.5dB		
		Sensitivitylevel	Microphones(Coupler Comparison Method) JJG 1019,V.R. of Working Standard Microphones (Free-field Comparison Method) JJG 1172	(-60～20)dB, 20Hz～20kHz	<i>U</i> =0.2dB		
				-80dB～0dB	<i>U</i> =0.15dB		
		Free-field Sensitivity Level		(-80～0)dB, 20 Hz～20 kHz	<i>U</i> =0.4dB		
				(-80～0)dB, (20 ～40) kHz	<i>U</i> =2.0dB		
4	Octave andFraction OctaveBand Filter	Relative Attenuation	V.R. of Octave-Band and Fractional-Octave-Band Filters JJG 449	Stopband: (0～130)dB	<i>U</i> =0.08dB		
				Passband: (0～130)dB	<i>U</i> =0.05dB		
5	LaboratoryStandar dMicrophones	Free-field Sensitivity Level	V.R. of Laboratory Standard Microphones(Free-field Reciprocity Method) JJG 482,V.R. of Laboratory Standard Microphones(Coupler Reciprocity Method) JJG 790	-80dB～0dB, (1～20)kHz	<i>U</i> =0.16dB		
				-80dB～0dB, (>20～25)kHz	<i>U</i> =0.20dB		
		Pressure FieldSensitivi tyLevel		-80dB～0dB, (20～31.5)Hz	<i>U</i> =0.05dB		
				-80dB～0dB, (63～4000)Hz	<i>U</i> =0.04dB		
				-80dB～0dB, 8kHz	<i>U</i> =0.05dB		
				-80dB～0dB, 10kHz	<i>U</i> =0.06dB		
				-80dB～0dB, 25kHz	<i>U</i> =0.11dB		




No. CNAS L0893

第 61 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
6	Personal Sound Exposure Meters	Personal Sound Exposure Meters	V.R. of Personal Sound Exposure Meters JJG 980	(0.3~100)Pa ² h, (20~800)Hz	<i>U</i> _{rel} =11%		
				(0.3~100)Pa ² h, (1~4)kHz	<i>U</i> _{rel} =9%		
				(0.3~100)Pa ² h, (5~8)kHz)	<i>U</i> _{rel} =14%		
7	Measurement Instruments of Sound Intensity	Sound Intensity Level	V.R. of The Measurement Instruments of Sound Intensity JJG 992	30dB~130dB, 45Hz~7.1kHz	<i>U</i> =0.5dB		
8	*Anechoic Rooms and Hemi-anechoic Rooms	SPL	C. S. for Acoustic Performance of Anechoic Rooms and Hemi-anechoic Rooms JJF 1147	0dB~130dB, 20Hz~20kHz	<i>U</i> =0.8dB		
9	*Reverberation Rooms	Reverberation Time	C. S. for Acoustic Performance of Reverberation Rooms JJF 1143	0.05ms~30s	<i>U</i> _{rel} =8.7%		
10	Sound Level meters	SPL	Verification Regulation of Sound Level Meters JJG 188	(40~120)dB, (10~200)Hz	<i>U</i> =0.4dB		
				(40~120)dB, (250 Hz~1.25kHz)	<i>U</i> =0.3dB		
				(40~120)dB, (1.6~10)kHz	<i>U</i> =0.5dB		
				(40~120)dB, (12.5 ~20)kHz	<i>U</i> =0.9dB		
				(120~160)dB, 100Hz	<i>U</i> =0.5dB		
		Time Weighing		F:(1~50)dB/s, 4kHz	<i>U</i> =3.0dB/s		
				S:(1~50)dB/s, 4kHz	<i>U</i> =0.2dB/s		
		Burst Response		(-40~10)dB, (0.25~1000)ms	<i>U</i> =0.2dB		




在线扫码获取验证

No. CNAS L0893

第 62 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Level Linearity		(10~160)dB, (10Hz~20kHz)	$U=0.2\text{dB}$		
11	Psophometers	Impedance	Calibration Specification for Calibration Specification for Psophometers JJF 1167	600 Ω	$U=0.7\ \Omega$		
				10k Ω	$U=12\ \Omega$		
				10M Ω	$U=10\text{k}\ \Omega$		
		Level		(-90~40)dB	$U=0.05\text{dB}$		
12	Noise Level Statistical Analyzers	Sound Pressure Level	Verification Regulation of Noise Level Statistical Analyzers JJG 778	(40~140)dB, (10~200)Hz	$U=0.4\text{dB}$		
				(40~140)dB, (250Hz~1.25kHz)	$U=0.3\text{dB}$		
				(40~140)dB, (1.6~10)kHz	$U=0.5\text{dB}$		
				(40~140)dB, (12.5~20)kHz	$U=0.9\text{dB}$		
13	Audio-frequency Signal Generator	Amplitude Frequency Characteristics	Verification Regulation of Audio-frequency Signal Generator JJG 607	-80dB~20dB, (20Hz~20kHz)	$U=0.2\text{dB}$		
				-80dB~20dB, (>20~200)kHz	$U=0.3\text{dB}$		
14	Electro-acoustical Measurement Instruments for Telephone Set	Frequency Response	Verification Regulation of Electro-acoustical Measurement Instruments for Telephone Set JJG 869	(0~10)dB, (100Hz~8kHz)	$U=0.4\text{dB}$		
15	Measuring Amplifiers	Sensitivity	Calibration Specification for Measuring Amplifiers JJF 1157	-26.0dB, 1kHz	$U=0.05\text{dB}$		
		Frequency Weighting		-80dB~20dB, 10Hz~20kHz	$U=0.08\text{dB}$		



No. CNAS L0893

第 63 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Frequency Response		(-80~20)dB, 2Hz~200kHz	$U=0.08\text{dB}$		
16	Preamplifiers	Frequency Response	Calibration Specification for Microphone Preamplifiers JJF 1137	(-20~20)dB, 10Hz~50kHz	$U=0.09\text{dB}$		
		Transmission Loss		-20dB~20dB, 10Hz~50kHz	$U=0.09\text{dB}$		
17	Audio-frequency Power Amplifiers	Frequency Response	Calibration Specification for Audio-frequency Power Amplifiers JJF 1200	(-20~40)dB, 10Hz~20kHz	$U=0.05\text{dB}$		
		Distortion		0.05%~100%	$U=10\%\text{rdg}$		
18	Electrostatic Actuators	Response Level	Calibration Specification for Electrostatic Actuators JJF 1293	WS1:-60dB~20dB, DC~20kHz	$U=0.06\text{dB}$		
				WS2:-60dB~20dB, DC~20kHz	$U=0.10\text{dB}$		
19	Tone Burst Generators	Duration	Verification Regulation of Tone Burst Generators JJG 199	0.1 ms~5 s	$U_{\text{rel}}=0.5\%$		
		Repeat Cycle		0.1 ms~5 s	$U_{\text{rel}}=0.5\%$		
20	Sonic Belt Tension Meters	Frequency	Calibration Specification for Sonic Belt Tension Meters JJF 1216	(0~100)Hz	$U=0.3\text{Hz}$		
				(>100~1000)Hz	$U=3.0\text{Hz}$		
				(>1000~20000)Hz	$U_{\text{rel}}=0.3\%$		
21	Measurement Instruments of Otoacoustic Emissions	SPL	Calibration Specification for Measurement Instruments of Otoacoustic Emissions JJF 1289	(30~90)dB	$U=1.0\text{dB}$		
		Frequency		0.5kHz~8kHz	$U_{\text{rel}}=0.01\%$		
		Distortion		(0.05~100)%	$U=10\%$		
22	Impedance Audiometers	Hearing Level	Verification Regulation of Audiometric Equipment—Instruments for the Measurement of Aural	-20dB~100dB, 250Hz~8kHz	$U=0.9\text{dB}$		


No. CNAS L0893

第 64 页 共 181 页



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		SpoundPressureLevel	Acoustic Impedance/Admittance JJG 991	0dB~130dB, (226Hz、678Hz、1000Hz)	<i>U</i> =0.8dB		
23	*Audiometry Rooms	SPL	Calibration Specification for Acoustic Performance of Audiometry Rooms JJF 1191	0dB~60dB, 20Hz~20kHz	<i>U</i> =0.8dB		
24	Pure-tone Audiometer	Frequency	Verification Regulation of Audiological Equipment Pure-tone Audiometers JJG 388	(100~20000)Hz	<i>U</i> _{rel} =0.01%		
		Distortion		0.05%~100%	<i>U</i> =10%rdg		
		RETSPL		(-10~110)dB, (125~4000)Hz	<i>U</i> =1.0dB		
				(-10~110)dB, (6000~12500)Hz	<i>U</i> =1.5dB		
		RETFL		(-10~70)dB, (250~4000)Hz	<i>U</i> =1.5dB		
				(-10~70)dB, (6000~8000)Hz	<i>U</i> =2.0dB		
25	*Anechoic Watertank	Echoic interference value	Calibration Specification for Acoustic Characteristics of Anechoic Water Tank JJF 1146	(80~160)dB, (500Hz~200kHz)	<i>U</i> _{rel} =2.4%		
26	Standard hydrophone	Sound Pressure Sensitivity Level	Standard Hydrophones in the Frequency Range 0.5 MHz to 5 MHz (Two-transducer Reciprocity Method) JJG	(-220~-160)dB, 1Hz~1kHz	<i>U</i> =0.9dB		
				(-270~-180)dB, (>1~100)kHz	<i>U</i> =0.9dB		



No. CNAS L0893

第 65 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			1070, Standard Hydrophones in the Frequency Range 1kHz to 1 MHz JJG 1017, Standard Hydrophones in the Frequency Range 1 Hz to 2 kHz (Closed-chamber Comparison Method) JJG 340	(-270~-180)dB, > 100kHz~5MHz	$U=1.1\text{dB}$		
27	Equipment of Ultrasonic Inspection Detectors	Frequency	C.R for Equipment of Ultrasonic Inspection Detectors NIMTT(CM) 097	(0.5~15)MHz	$U_{\text{rel}}=0.07\%$		
		Attenuation		(0~81)dB	$U=0.06\text{dB}$		
		Voltage		(0.1~10)V	$U_{\text{rel}}=4.0\%$		
28	Acoustic Detector	Time of transmitted sound	Verification Regulation of Acoustic Detector JJG 990	$1\text{ }\mu\text{s} \sim 10\text{s}$	$U_{\text{rel}}=6.7 \times 10^{-4} \times (0.047/d^2 + 1.6^2)^{0.5} \times 100\%$		
29	Ultrasonic Standard Test Tube	Length	Ultrasonic Standard Test Tube NIMTT (CM) 041	(1~400)mm	$U=(1.5+L)\text{ }\mu\text{m}$	L is the length, m	
30	Air Ultrasound Measuring Instruments	Frequency	Calibration Specification for Air Ultrasound Measuring Instruments JJF 1504	100Hz~50kHz	$U=2.5\text{dB}$		
31	Acoustic Emission Instrumentation	Sound pressure level	Calibration Specification for Acoustic Emission Instrumentation JJF 1505	(0~60)dB	$U=0.4\text{dB}$		
		frequency		10Hz~100MHz	$U_{\text{rel}}=5\%$		
32	Environmental Noise Automatic Monitors	SPL	Verification Regulation of Environmental Noise Automatic Monitors JJG 1095	(0~120)dB, ((31.5~2k)Hz	$U=0.5\text{dB}$		
				(0~120)dB, (>2k~16k)Hz	$U=0.9\text{dB}$		
		Time Weighing		(60~120)dB, (10Hz~20kHz)	0.3dB/s(S); $U=1.5\text{dB/s(F)}$		



No. CNAS L0893

第 66 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Burst Response		(60~120)dB	$U=0.2\text{dB}$		
		Level Linearity		(10~160)dB	$U=0.2\text{dB}$		
33	Multi-Channels Sound Analyzers	SPL	Calibration Specification for Multi-Channels Sound Analyzers JJF 1288	(10~160)dB, (10Hz~20kHz)	$U=0.2\text{dB}$		
		Level Linearity		(10~150)dB	$U=0.2\text{dB}$		
				(0~50)dB	$U=0.2\text{dB}$		
		Time Weighing		(60~120)dB, (10Hz~20kHz)	$0.3\text{dB/s(S)}; U=1.5\text{dB/s(F)}$		
		amplitude frequency characteristics		(10~160)dB, (10Hz~20kHz)	$U=0.2\text{dB}$		
		Frequency		0%~100%	$U=0.05\%$		
		Amplitude		0%~100%	$U=0.3\%$		
		frequency spectrum features		(10~160)dB, (10Hz~20kHz)	$U=0.2\text{dB}$		
34	High Sound Pressure Measuring Microphones	Level Linearity	Calibration Specification of the Upper Limit of Dynamic Range of High Sound Pressure Measuring Microphones JJF 1738	(-2.0~2.0)dB	$U=0.2\text{dB}$		
		Total harmonic distortion		0%~20%	$U=0.8\%$		
		upper limit for dynamic range		0dB~171dB	$U=2\text{dB}$		
35	Elastic Wave-Based Measurement Apparatus	time indication	Calibration Specification for Elastic Wave-Based Measurement Apparatus NIMTT(CM) 031	0.1%~100%	$U_{\text{rel}}=0.3\%$		

No. CNAS L0893

第 67 页 共 181 页



The scope of the accreditation in Chinese remains the definitive version.



№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		voltage amplitude		0.1%~100%	$U_{rel}=2.0\%$		
36	Ultrasonic Flow Detectors by Time-of-Flight Diffraction	Receiver bandwidth	Calibration Specification for Ultrasonic Flow Detectors by Time-of-Flight Diffraction JJF 1447	100kHz~1MHz	$U_{rel}=2.5\%$		
37	Acoustic Emission Sensors	Sensitivity Level	Calibration Specification for Acoustic Emission Sensors(Comparative Method) JJF 1337	(40~120)dB, 100kHz~1MHz	$U=3.1\text{ dB}$		
38	Ultrasonic Flaw Detector for Rail	attenuator	Calibration Specification for Ultrasonic Flaw Detector for Rail JJG(铁道)130	40dB	$U=0.5\text{dB}$		
39	*Ultrasonic Flow Detectors	Attenuation	V.R. for Ultrasonic Flow Detectors JJG 746	(0~81)dB, (0.5~15)MHz	$U=0.2\text{ dB}$		
		Electrical noise		(0~100%)	$U=5.0\%$		
		horizontal linearity		(0~100)%, (0.5~15)MHz	$U=0.2\%$		
		Voltage		$(-4\times 10^8\sim 4\times 10^8)\mu\text{V}$, (0.5~15)MHz	$U=40\mu\text{V}$		
		Dynamic range		(0~81)dB, (0.5~15)MHz	$U=3\text{ dB}$		
		Linearity of Vertical Display		(0~100)%, (0.5~15)MHz	$U=0.7\%$		
40	*Ultrasonic Phased Array Flow Detectors	length	Calibration Specification for Ultrasonic Phased Array Flow Detectors JJF 1338	(0.1~55)mm	$U_{rel}=2.0\%$		
		angle		angle:(0.1~90)°	$U_{rel}=2.0\%$		
41	Transducers of Ultrasonic flow	frequency	Calibration Specification for Transducers of Ultrasonic	(0.5~15)MHz	$U_{rel}=2.3\%$		

No. CNAS L0893


第 68 页 共 181 页



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	detector	relative pulse echo	flaw detector JJF 1294	(-100~-0.1)dB	$U_{rel}=1.5\%$		
		Length		(0.1~50)mm	$U_{rel}=2.1\%$		
		angle		$35^{\circ} \sim 85^{\circ}$	$U_{rel}=3.6\%$		
42	Ultrasonic Power Meter for Mill watt Level	power	Verification Regulation of Ultrasonic Power Meter for Milliwatt Level JJG665	(1~5)mW	$U_{rel}=7\%$		
		power		(5~100)mW	$U_{rel}=5.8\%$		
		power		(100~150)mW	$U_{rel}=10\%$		
43	Ultrasound Phantoms	Sound velocity	Calibration Specification for Ultrasound Phantoms JJF 1556	(1000~3000)m/s	$U_{rel}=0.3\%$		
		Attenuation		(0.1~80)dB/cm	$U_{rel}=6\%$		
44	*Ultrasonic Source for Medical Ultrasonic Diagnostic Equipment	sound intensity	V.R. of Ultrasonic Source for Medical Ultrasonic Diagnostic Equipment JJG 639	(1~10)mW/cm ²	$U_{rel}=11\%$		
		definition		(0~190)mm	$U=0.1\text{mm}$		
V Electromagnetism measuring instrument							
1	Withstanding Voltage Tester	AC Voltage	Withstanding Voltage Tester	(0.1~100)kV	$U_{rel}=0.05\%$		
		DC Voltage	JJG795,C. S. of Withstanding Voltage Tester	(0.1~100)kV	$U_{rel}=0.1\%$		
		Breakdown current	(15kV~100kV)	0.1mA~2A	$U_{rel}=0.1\%$		
		Time	NIMTT(CM) 022	0.1s~999s	$U_{rel}=1.0\%$		
2	High Voltage Electrostatic Voltmeter	AC Voltage	High Voltage Electrostatic	(0.1~100)kV,50Hz	$U_{rel}=0.26\%$		
		DC Voltage	JJG494	(0.1~100)kV	$U_{rel}=0.16\%$		



№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
3	Digital high-Voltage meter	AC Voltage	Digital high-voltage meter verification regulation	(0.1~300)kV, 50Hz	$U_{rel}=0.05\%$		
		DC Voltage	JJG(川)101	(0.1~300)kV	$U_{rel}=0.1\%$		
4	High Voltage Divider at Power Frequency	Voltage	Verification Regulation of High Voltage Divider at Power Frequency JJG496	(0.1~300)kV, 50Hz	$U_{rel}=0.05\%$		
5	DC High Voltage Dividers	Voltage	DC High Voltage Dividers JJG1007	(0.1~300)kV	$U_{rel}=0.1\%$		
6	high-voltage prob	DC Voltage	C.S. of high-voltage prob NIMTT(CM) 020	(0.1~100)kV	$U_{rel}=0.1\%$		
		AC Voltage		(0.1~100)kV, 50Hz	$U_{rel}=0.05\%$		
		Spike Voltage		(1~15)kV	$U_{rel}=2.0\%$		
7	Industrial frequency, dc spark machine	voltage	verification procedure for test equipmet of rubber plastic wire and cable-part 10:spark tester JB/T 4278.10	(1~50)kV	$U_{rel}=0.5\%$		
8	Contactless electrostatic voltmeter	voltage	Contactless electrostatic voltmeter calibration specifications JJF 1517, Contactless electrostatic voltmeter calibration specifications GJB/J 5972	(0.1~20)kV	$U_{rel}=0.1\%$		
9	*Voltage Transformers of MeasUring Service	Difference value	V.R. of Voltage Transformers of MeasUring Service JJG 314	20%Un:(100/√3~500000/√3)/(100~100/√3)V	$U=2.4 \times 10^{-5}$		
				50%Un:(100/√3~500000/√3)/(100~100/√3)V	$U=1.8 \times 10^{-5}$		



No. CNAS L0893

第 70 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				80%~120%Un:(100/√3~500000/√3)/(100~100/√3)V	<i>U</i> =1.2×10 ⁻⁵		
		Phase difference		20%Un:(0~0.29)rad	<i>U</i> =2.4×10 ⁻⁵ rad		
		50%Un:(0~0.29)rad		<i>U</i> =1.8×10 ⁻⁵ rad			
		80%~120%Un:(0~0.29)rad		<i>U</i> =1.2×10 ⁻⁵ rad			
10	*Instrument Transformers in Power System	Difference value	V. R. of Instrument Transformers in Power System JJG1021	(80~120) %Un:(100/√3~500000/√3)/(100~100/√3)V	<i>U</i> =0.01%		
		Phase difference		(80~120) %Un:(0~999.9)′	<i>U</i> =0.8′		
		Difference value		1%In:(0.1~10000)/(1、5)A	<i>U</i> =0.02%		
				5%In:(0.1~10000)/(1、5)A	<i>U</i> =0.01%		
				20%~120%In:(0.1~10000)/(1、5)A	<i>U</i> =0.01%		
		Phase difference		1%In:(0~999.9)′	<i>U</i> =1.2′		
				5%In:(0~999.9)′	<i>U</i> =1.0′		
				20%~120%In:(0~999.9)′	<i>U</i> =0.8′		
11	Inductive Voltage Divider	Difference value	V.R of Inductive Voltage Divider JJG 244	(10~1000) / (10~200) V	<i>U</i> =1.2×10 ⁻⁵		
		Phase difference		(0~0.29)rad	<i>U</i> =1.2×10 ⁻⁵ rad		



No. CNAS L0893

第 71 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
12	Standard Cell	Electromotive force	Verification Regulation of Standard Cell JJG 153	1.018V(rating voltage)	$U_{\text{rel}}=2\times 10^{-7}$		
13	Standard Cell	Electromotive force	Verification Regulation of Working Sandard of DC Electromotive Force JJG 719	1.018V(rating voltage)	$U_{\text{rel}}=3\times 10^{-8}$		
14	DC Volt Ratio Box	Voltage Ratio	Verification Regulation of The DC Resistive Volt Ratio Box JJG531	10:1, 10V~1.0 kV	$U_{\text{rel}}=5\times 10^{-5}$		
				100:1, 10V~1.0 kV	$U_{\text{rel}}=2\times 10^{-5}$		
				200:1, 10V~1.0 kV	$U_{\text{rel}}=7\times 10^{-5}$		
				500:1, 10V~1.0 kV	$U_{\text{rel}}=7\times 10^{-5}$		
15	Residual Current Operated Protective Decice Operated Characteristic Tester	DC Current	Calibtation Specification for Residual Current Operated Protective Device Operated Chataceristic Tester JJF1283	(3~1500)mA	$U_{\text{rel}}=0.02\%$		
		AC Current		(3~1500)mA.(40~65)Hz	$U_{\text{rel}}=0.05\%$		
		Time		(0.01~999999)ms	$U=0.02\text{ms}\sim 3.9\text{ms}$		
16	Leakage current tester	voltage	V.R. of Leakage current tester JJG843	(1~1000)V	$U_{\text{rel}}=0.1\%$		
		current		0.01mA~10A	$U_{\text{rel}}=0.1\%$		
17	*Current Transformers of MeasUring Service	Difference value	V.R. of Current Transformers of MeasUring Service JJG 313	1%In:(0.1~10000)/(1、5)A	$U=5\times 10^{-6}$		
				5%In:(0.1~10000)/(1、5)A	$U=4\times 10^{-6}$		
				20%~120%In:(0.1~10000)/(1、5)A	$U=3\times 10^{-6}$		
		Phase difference		1%In:(0~0.29)rad	$U=5\times 10^{-6}\text{rad}$		
				5%In:(0~0.29)rad	$U=4\times 10^{-6}\text{rad}$		



No. CNAS L0893

第 72 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				20%~120%In:(0~0.29)rad	$U=3 \times 10^{-6}$ rad		
18	DC Electronic Loads	DC Voltage	Calibration Specification for DC Electronic Loads JJF1462	(1~240)V	$U_{rel}=0.01\%$		
		DC Current		(0.1~240)A	$U_{rel}=0.02\%$		
		DC Power		1W~12kW	$U_{rel}=0.02\%$		
19	lightning element tester	DC Voltage	Calibration method of lightning element tester NIMTT(CM) 084	10mV~2000V	$U_{rel}=0.2\%$		
		DC Current		1 μ A~10A	$U_{rel}=0.2\%$		
		Resistance		0.1 Ω ~100M Ω	$U_{rel}=0.2\%$		
20	Galvanometer	Current	Verification Regulation of DC Magnetoelectric Galvanometers JJG495	10^{-4} A~ 10^{-7} A	$U_{rel}=0.1\%$,		
				10^{-7} A~ 10^{-8} A	$U_{rel}= (0.1\% \sim 0.5\%)$		
				10^{-8} A~ 10^{-10} A	$U_{rel}= (0.5\% \sim 2\%)$		
				10^{-10} A~ 10^{-12} A	$U_{rel}= (2\% \sim 5\%)$		
21	DC Resistors	Resistances	Verification Regulation of DC Resistors JJG166	Working standard:($10^{-3} \sim 10^5$) Ω	$U_{rel}=(0.25 \sim 1) \times 10^{-6}$		
				Grade 1:($10^{-3} \sim 10^5$) Ω	$U_{rel}=(0.5 \sim 3) \times 10^{-6}$		
				Grade 2:($10^{-3} \sim 10^5$) Ω	$U_{rel}=3 \times 10^{-5}$		
22	DC Bridges for Measuring Temperature	Resistances	Verification Regulation of DC Bridges for Measuring Temperature JJG484	($10^{-2} \sim 10^{-1}$) Ω	$U_{rel}=1 \times 10^{-3}$		
				($10^{-1} \sim 1$) Ω	$U_{rel}=1 \times 10^{-4}$		
				($1 \sim 10^4$) Ω	$U_{rel}=1 \times 10^{-5}$		
23	DC Low Resistance Meters	Resistances	Verification Regulation of DC Low Resistance Meters	($10^{-6} \sim 10^{-4}$) Ω	$U_{rel}=1 \times 10^{-3}$		



在线扫码获取验证

No. CNAS L0893

第 73 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
			JJG837	$(10^{-4} \sim 10^{-1}) \Omega$	$U_{\text{rel}}=5 \times 10^{-4}$		
				$(10^{-1} \sim 10^5) \Omega$	$U_{\text{rel}}=5 \times 10^{-5}$		
24	Loop Resistance Tester and DC Resistance Meters	Resistances	Verification Regulation of Loop Resistance Tester and DC Resistance Meters JJG1052	$(10^{-6} \sim 10^{-4}) \Omega$	$U_{\text{rel}}=1 \times 10^{-3}$		
				$(10^{-4} \sim 10^{-1}) \Omega$	$U_{\text{rel}}=5 \times 10^{-4}$		
				$(10^{-1} \sim 10^5) \Omega$	$U_{\text{rel}}=5 \times 10^{-5}$		
25	DC Bridge	Resistances	Verification Regulation of DC Bridge JJG125	$(10^{-4} \sim 1) \Omega$	$U_{\text{rel}}=1 \times 10^{-3}$		
				$(1 \sim 10) \Omega$	$U_{\text{rel}}=1 \times 10^{-4}$		
				$(10 \sim 10^7) \Omega$	$U_{\text{rel}}=1 \times 10^{-5}$		
26	DC Comparison Bridge	Resistances	Verification Regulation of DC Comparison Bridge JJG546	$(10^{-2} \sim 1) \Omega$	$U_{\text{rel}}=1 \times 10^{-3}$		
				$(1 \sim 10^4) \Omega$	$U_{\text{rel}}=1 \times 10^{-5}$		
27	DC Shunts	Resistances	Verification Regulation of DC Shunts JJG1069	$(10^{-6} \sim 1) \Omega$, Rate Current: ($5 \sim 10000$) A	$U_{\text{rel}}=5 \times 10^{-5}$		
28	Earth Resistance Meter	Resistance	Verification Regulation of Earth Resistance Meters JJG366	$0.01 \Omega \sim 10 \Omega$	$U_{\text{rel}}= (0.5\% \sim 1\%)$		
				$10 \Omega \sim 2000 \Omega$	$U_{\text{rel}}= (0.5\% \sim 5\%)$		
29	Clamp Earth Resistance Meters	Resistance	Verification Regulation of Clamp Earth Resistance Meters JJG1054	$0.01 \Omega \sim 10 \Omega$	$U_{\text{rel}}= (0.5\% \sim 1\%)$		
				$10 \Omega \sim 2000 \Omega$	$U_{\text{rel}}= (0.5\% \sim 5\%)$		
30	surface resistance tester	Resistance	Calibration specification for surface resistance tester JJF1285	$10^2 \Omega \sim 10^6 \Omega$	$U_{\text{rel}}= (0.001\% \sim 0.002\%)$		
				$10^6 \Omega \sim 10^8 \Omega$	$U_{\text{rel}}= (0.002\% \sim 0.005\%)$		
				$10^8 \Omega \sim 10^{10} \Omega$	$U_{\text{rel}}= (0.005\% \sim 0.2\%)$		
				$10^{10} \Omega \sim 10^{13} \Omega$	$U_{\text{rel}}= (0.2\% \sim 5\%)$		

No. CNAS L0893

第 74 页 共 181 页



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				$10^{13} \Omega \sim 10^{15} \Omega$	$U_{\text{rel}} = (5\% \sim 10\%)$		
31	Industry Frequency Single-Phase Phase Meter	Mutually Phase(power Factor)	Verification Regulation of Industry Frequency Single-Phase Phase Meter JJG440	$0^\circ \sim 360^\circ$	$U=0.05^\circ$		
32	Low-frequency Phase Meters	Mutually Phase(power Factor)	Calibration Specification for Low-frequency Phase Meters JJF1756	$0^\circ \sim 360^\circ$, (45~65)Hz	$U=0.005^\circ$		
33	Instrument Transformer Test Set	Ratio variant	Verification Regulation of Instrument Transformer Test Set JJG 169	$1 \times 10^{-9} \sim 1.111 \times 10^{-2}$	$U_{\text{rel}}=0.4\% \sim 0.2\%$		
		Angular Difference		$1 \times 10^{-9} \sim 11.11 \times 10^{-2}$	$U_{\text{rel}}=0.4\% \sim 0.2\%$		
		Impedance		$(10^{-3} \sim 111.1) \Omega$	$U_{\text{rel}}=0.8\% \sim 0.4\%$		
		Admittance		$(10^{-8} \sim 0.1111) \text{S}$	$U_{\text{rel}}=0.4\%$		
34	Burden Box of Instrument Transformers	Impedance	Calibration Specification for Burden Box of Instrument Transformers JJF1264	$(10^{-3} \sim 111.1) \Omega$	$U_{\text{rel}}=0.8\% \sim 0.4\%$		
		Admittance		$(10^{-8} \sim 0.1111) \text{S}$	$U_{\text{rel}}=0.4\%$		
35	Volt-ampere characteristic and polarity Measuring Instrument	Current Transformation Ratio	Calibration Specification for Volt-ampere characteristic and polarity Measuring Instrument NIMTT(CM) 048	$0.1/5\text{A} \sim 10000/5\text{A}$	$U_{\text{rel}}=0.01\%$		
		Voltage Transformation Ratio		$(100/100\text{V} \sim 10000/100\text{V})$	$U_{\text{rel}}=0.01\%$		
				$10000/100\text{V} \sim 35/0.1\text{kV}$	$U_{\text{rel}}=0.01\% \sim 0.05\%$		
				$(0.1 \sim 50)\text{A}$	$U_{\text{rel}}=0.5\% \sim 0.2\%$		
				$(50 \sim 100)\text{A}$	$U_{\text{rel}}=0.2\% \sim 0.1\%$		
				$(100 \sim 1000)\text{A}$	$U_{\text{rel}}=0.1\%$		
		Current		$(1000 \sim 10000)\text{A}$	$U_{\text{rel}}=0.1\% \sim 0.5\%$		



No. CNAS L0893

第 75 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Voltage	ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	(0.1~50)V	$U_{rel}=0.5\%\sim 0.1\%$		
				50V~35kV	$U_{rel}=0.1\%$		
		resistant		(0.01~1000) Ω	$U_{rel}=0.1\%$		
		load		(0.01~80) kVA	$U_{rel}=0.1\%$		
36	AC and DC converter	Voltage	Calibration method for AC and DC converter NIMTT(CM) 088	5 V~600V	$U_{rel}=0.01\%$		
		Current		25mA ~10A	$U_{rel}=0.01\%$		
		Power		5V~600V/25mA~10A	$U_{rel}=0.01\%$		
37	Magnetic field coil	Coils constant	Calibration Specification for Magnetic Field Coil NIMTT(CM) 003	$(1 \times 10^{-4} \sim 2 \times 10^{-2})$ T/A	$U_{rel}=0.02\%$		
38	Power amplifier	Power Gain	calibration specification for RF and microwave power amplifier JJF1678	(0~80)dB, (9kHz~18GHz)	$U=0.49$ dB		
		1dB compression point Gain		(0~80)dB, (9kHz~18GHz)	$U=0.49$ dB		
		harmonic suppression		0~80)dBc,(9kHz~18GHz)	$U=0.49$ dB		
39	Magnetometer	Magnetic induction intensity	Verification Regulation of Tesla-Meter JJF 1832	(1~43)mT	$U_{rel}=0.1\%$		
				(43~2000) mT	$U_{rel}=0.02\%$		
40	Fluxmeter	Magnetic flux	Calibration Specification for Magnetic Flux Meters JJF 1905	(0.1~1)mWb	$U=0.002$ mWb		
				1mWb~10Wb	$U_{rel}=0.1\%$		



No. CNAS L0893

第 76 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
41	Alternating magnetic intensity meter	Magnetic induction intensity	Verification Regulation of Alternating Tesla-Meter for Weak Magnetic Field JJG 1049	(1～1000) μ T	<i>U</i> =(0.12～5) μ T		
42	Magnetometers Based Magnetic Force	Magnetic induction intensity	Calibration Specification for Magnetometers Based Magnetic Force JJF 1656	(0.05～10)mT	<i>U</i> _{rel} =1.3%		
43	DC Potentiometer	voltage	Verification Regulation of DC Potentiometer JJG 123	0.01 μ V～ 10mV	<i>U</i> =0.01 μ V		
				10mV～ 2.1111111V	<i>U</i> _{rel} =1×10 ⁻⁵		
44	DC Comparison Potentiometer	voltage	Verification Regulation of DC Comparison Potentiometer JJG 505	0.01 μ V～ 10mV	<i>U</i> =0.01 μ V		
				10mV～ 2.1111111V	<i>U</i> _{rel} =1×10 ⁻⁶		
45	DC Current Comparator Bridge	Resistances	Verification Regulation of DC Current Comparator Bridge JJG 506	(10 ⁻³ ～10 ⁵) Ω	<i>U</i> _{rel} =(0.5～1)×10 ⁻⁶		
		ratio		0.95～1.05	<i>U</i> _{rel} =5×10 ⁻⁸		
				0.01～0.95	<i>U</i> _{rel} =1×10 ⁻⁷		
				1.05～100	<i>U</i> _{rel} =1×10 ⁻⁷		
46	DC Comparator Bridge for measuring temperatures	Resistances	Calibration specification for DC Comparator Bridge for measuring temperatures JJF 1444	(10 ⁻³ ～10 ⁵) Ω	<i>U</i> _{rel} =(0.5～1)×10 ⁻⁶		
		ratio		0.95～1.05	<i>U</i> _{rel} =5×10 ⁻⁸		
				0.01～0.95	<i>U</i> _{rel} =1×10 ⁻⁷		
				1.05～100	<i>U</i> _{rel} =1×10 ⁻⁷		
47	Electromechanical watt hour meters	Alternating-current Electrical Energy	V.R. of Electromechanical Meters for Measuring Alternating-current Electrical Energy JJG 307	57.7V～380V,25mA～100A	<i>U</i> _{rel} =0.2%		



No. CNAS L0893

第 77 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
48	Electrical energy meters	Alternating-current Electrical Energy	V.R. of Electrical Meters for Measuring Alternating-current Electrical Energy JJG 596, V.R. of Multi-Rate Electricity Meters for Measuring Alternating-current Electrical Energy JJG 691, V.R. of Pre-payment Electrical Meters for Measuring Alternating-current Electrical Energy JJG 1099, V.R. of Electricity Meters with Maximum Demand Measurement Functions JJG 569	57.7V~380V, 1mA~100A	$U_{rel}=4 \times 10^{-4}$		
		time		(-99.99 ~ 99.99) s/d	$U=0.06\text{s/d}$		
49	Reference meters for electrical energy	Alternating-current Electrical Energy	V.R. of Reference Meters for Electrical Energy JJG 1085	3P4W active 0.05A~100A, 57.7V~380V	$U_{rel}=8 \times 10^{-5}$		
				3P3W active 0.05A~100A, 57.7V~380V	$U_{rel}=1 \times 10^{-4}$		
				3P4W reactive 0.05A~100A, 57.7V~380V	$U_{rel}=1.3 \times 10^{-4}$		
				3P4W active 0.01A~0.05A, 57.7V~380V	$U_{rel}=1.3 \times 10^{-4}$		
				3P3W active 0.01A~0.05A, 57.7V~380V	$U_{rel}=1.6 \times 10^{-4}$		
				3P4W active 0.001A~0.01A, 57.7V~380V	$U_{rel}=2.5 \times 10^{-4}$		
50	Electrical energy performance field	Alternating-current Electrical	C.S. for Electrical Energy Meter on-site calibrator NIMTT(CM) 074	57.7V~380V, port: 0.1~20A	$U_{rel}=1.2 \times 10^{-4}$		



No. CNAS L0893

第 78 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
	calibrator	Energy		57.7V~380V,clamp: 0.1~100A	$U_{\text{rel}}=2\times10^{-4}$		
		voltage		57.7V~380V	$U_{\text{rel}}=1.2\times10^{-4}$		
		current		port: 0.1~20A	$U_{\text{rel}}=1.2\times10^{-4}$		
				clamp: 0.1~100A	$U_{\text{rel}}=2\times10^{-4}$		
		power		57.7V~380V,port: 0.1~20A	$U_{\text{rel}}=1.2\times10^{-4}$		
				57.7V~380V,clamp: 0.1~100A	$U_{\text{rel}}=2\times10^{-4}$		
		phase		(0~360)°	$U=0.01^{\circ}$		
51	*verification equipment for AC electrical energy meters	Alternating- current	V.R. Verification Equipment for AC Electrical Energy Meters JJG 597	0.1A~100A,57.7V~ 380V	$U_{\text{rel}}=1.2\times10^{-4}$		
		Electrical Energy		0.001A~0.1A,57.7V~ 380V	$U_{\text{rel}}=3\times10^{-4}$		
		voltage		57.7V~380V	$U_{\text{rel}}=1.2\times10^{-4}$		
		current		0.1A~100A	$U_{\text{rel}}=1.2\times10^{-4}$		
				0.001A~0.1A	$U_{\text{rel}}=3\times10^{-4}$		
		power		0.1A~100A,57.7V~ 380V	$U_{\text{rel}}=1.2\times10^{-4}$		
				0.001A~0.1A,57.7V~ 380V	$U_{\text{rel}}=3\times10^{-4}$		
		phase		0° ~360°	$U=0.01^{\circ}$		
		frequency		45Hz~65Hz	$U=0.001\text{Hz}$		



No. CNAS L0893

第 79 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
52	*DC Digital Wattmeter	DC Power	Calibration method of DC digital power meter NIMTT(CM) 076	0.1V~600V,10 μ A~400A	<i>U</i> _{rel} =0.02%		
53	*Testers for Relaying Protection	DC Voltage	V.R. of Testers for Relaying Protection JJG 1112	10mV~1000V	<i>U</i> _{rel} =0.002%		
		DC Current		1mA~100A	<i>U</i> _{rel} =0.02%		
		AC Voltage		10mV~750V,10Hz~1kH	<i>U</i> _{rel} =0.02%		
		AC Current		1mA~100A,10Hz~1kHz	<i>U</i> _{rel} =0.02%		
		Mutually Phase		(0~360)° ,10Hz~1kHz	<i>U</i> =0.006°		
		Time		(1~999999)ms	<i>U</i> =0.02ms~3.9ms		
		Frequency		10Hz~1kHz	<i>U</i> _{rel} =0.001%		
54	*Harmonious and Flicker Analysis System (Power Quality Analyzer)	Voltage	Calibration Specification for Harmonious and Flicker Analysis System JJF 1205	(3~1000)V,45Hz~65Hz	<i>U</i> _{rel} =0.01%		
		Current		(0.05~100)A ,45Hz~65Hz	<i>U</i> _{rel} =0.01%		
				(100~3000)A,45Hz~65Hz	<i>U</i> _{rel} =0.2%		
		Power		(3~1000)V,(0.05~100)A ,45Hz~65Hz	<i>U</i> _{rel} =0.02%		
				(3~1000)V,(100~1000)A,45Hz~65Hz	<i>U</i> _{rel} =0.2%		
		Frequency		16Hz~850Hz	<i>U</i> _{rel} =0.001%		
		Phase		(0~360)° ,16Hz~180Hz	<i>U</i> =0.003°		



No. CNAS L0893

第 80 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Harmonic Voltage		(0.03~300)V ,(2~60)Times	$U_{rel}=0.01\% \sim 0.02\%$		
		Harmonic Current		(0.005~30)A,(2~60)Times	$U_{rel}=0.01\% \sim 0.02\%$		
				(30~100)A,(2~60)Times	$U_{rel}=0.2\%$		
		Flicker		(0.40~13.62)%	$U_{rel}=0.06\%$		
55	Transformer with load tap changer	Time	Calibration method of transformer with load tap changer NIMTT(CM) 078	(0.01~2000)ms	$U=0.02\text{ms}$		
		Resistance		(0.1~100) Ω	$U=0.10\text{ }\Omega$		
56	Electrical Quantities Transducer	AC Current	V.R. of Measuring transducers for converting a.c.electrical quantities into d.c.electrical quantities JJG 126	(0.1~100)A,40Hz~1kHz	$U_{rel}=0.02\%$		
		AC Voltage		10mV~1000V,40Hz~1kHz	$U_{rel}=0.01\%$		
		ACPower		$3\times(30\sim600)\text{V}$,(0.005~100)A,40Hz~65Hz	$U_{rel}=0.02\%$		
		Frequency		10Hz~100kHz	$U_{rel}=0.01\%$		
57	*DC Stabilized Power Supplies	DC Voltage	Calibration Specification for DC Stabilized Power Supplies JJF 1597	(0.01~1000)V	$U_{rel}=0.002\%$		
		DC Current		(0.01~10000)A	$U_{rel}=0.005\%$		
		AC Voltage		10mV~10V40Hz~100kHz	$U_{rel}=0.01\%$		
58	*Battery charging and discharging tester	DC Voltage	Calibration method for battery charging and discharging tester NIMTT(CM) 082	(0.01~1000)V	$U_{rel}=0.01\%$		
		DC Current		(0.01~1000)A	$U_{rel}=0.01\%$		
				(1000~10000)A	$U_{rel}=0.5\%$		
		Resistance		1m Ω ,10m Ω ,100m Ω ,1 Ω	$U_{rel}=0.02\%$		



No. CNAS L0893

第 81 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
59	*Digital AC Electrical Parameters Meter	AC Voltage	Calibration Specification for Digital AC Electrical Parameters Meter JJF 1491	0.01V~1000V,40Hz~1kHz	$U_{rel}=0.01\%$		
		AC Current		30 μ A~100A,(40~65)Hz	$U_{rel}=0.01\%$		
		AC Power		0.2A~10A,65Hz~1kHz	$U_{rel}=0.05\%$		
		Frequency		3V~600V,0.005A~100A,(40~65)Hz	$U_{rel}=0.01\%$		
		Phase		5V~600V,0.025A~10A,65Hz~1kHz	$U_{rel}=0.05\%$		
				40Hz~1kHz	$U_{rel}=1 \times 10^{-5}$		
				0° ~360° ,(40~69)Hz	$U=0.005^\circ$		
60	Process Calibrators	DC Voltage	Calibration Specification for Process Calibrators JJF 1472	0.01 V~1000V	$U_{rel}=0.001\%$		
		AC Voltage		0.01V~1000V,40Hz~10kHz	$U_{rel}=0.01\%$		
		DC Current		10 μ A ~50A	$U_{rel}=0.01\%$		
		AC Current		100 μ A~100A,(40~65)Hz	$U_{rel}=0.01\%$		
		Frequency		0.2A~10A,65Hz~5kHz	$U_{rel}=0.05\%$		
		Resistance		1Hz~100kHz	$U_{rel}=1 \times 10^{-5}$		
				10 Ω ~1M Ω	$U_{rel}=0.01\% \sim 0.002\%$		
	De high		C.S. for DC high voltagesource, dc highvoltage	(0.1~1)kV	$U_{rel}=0.1\%$		




No. CNAS L0893

第 82 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
	generator		generatorcalibration specifications NIMTT(CM) 054	(> 1~300)kV	<i>U</i> _{rel} =0.2%		
		current		10 μ A~1A	<i>U</i> _{rel} =0.1%		
62	*Lightning impulse voltage test system.	voltage	calibration specification of impulse voltage divider NIMTT(CM) 129	(0.1~500)kV	<i>U</i> _{rel} =1.0%		
		time		(1~3000) μ s	<i>U</i> _{rel} =3.0%		
63	*Lightning impulse current test system.	voltage	Lightning impulse current test system.Calibration specification NIMTT(CM) 011	(1~100) kA	<i>U</i> _{rel} =1.5%		
		current		(1~500) μ s	<i>U</i> _{rel} =3.0%		
64	Multifunction standard Source	DCV	Calibration Specification for Multifunction Standard Sources JJF 1638	10mV~100mV	<i>U</i> _{rel} =5.5×10 ⁻⁶ ~1.0×10 ⁻⁵		
				100mV~1V	<i>U</i> _{rel} =4.5×10 ⁻⁶ ~5.0×10 ⁻⁶		
				1V~10V	<i>U</i> _{rel} =1.5×10 ⁻⁶ ~5.0×10 ⁻⁶		
				10V~100V	<i>U</i> _{rel} =2.7×10 ⁻⁶ ~3.6×10 ⁻⁶		
				100V~1000V	<i>U</i> _{rel} =3.7×10 ⁻⁶ ~4.5×10 ⁻⁶		
		ACV		10mV~220mV,10Hz~ 40Hz	<i>U</i> _{rel} =2.5×10 ⁻⁴		
				10mV~220mV,40Hz~ 20kHz	<i>U</i> _{rel} =1.0×10 ⁻⁴		
				10mV~220mV,20kHz~ 100kHz	<i>U</i> _{rel} =3.4×10 ⁻⁴		
				10mV~ 220mV,100kHz~1MHz	<i>U</i> _{rel} =1.8×10 ⁻³		
				220mV~2.2V,10Hz~ 40Hz	<i>U</i> _{rel} =2.5×10 ⁻⁴		
				220mV~2.2V,40Hz~ 20kHz	<i>U</i> _{rel} =0.5×10 ⁻⁴		



No. CNAS L0893

第 83 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				220mV~2.2V,20kHz~100kHz	$U_{rel}=2.0 \times 10^{-4}$		
				220mV~2.2V,100kHz~1MHz	$U_{rel}=1.5 \times 10^{-3}$		
				2.2V~22V,10Hz~40Hz	$U_{rel}=2.5 \times 10^{-4}$		
				2.2V~22V,40Hz~20kHz	$U_{rel}=0.5 \times 10^{-4}$		
				2.2V~22V,20kHz~100kHz	$U_{rel}=2.0 \times 10^{-4}$		
				2.2V~22V,100kHz~1MHz	$U_{rel}=5.0 \times 10^{-3}$		
				22V~220V,10Hz~40Hz	$U_{rel}=2.5 \times 10^{-4}$		
				22V~220V,40Hz~20kHz	$U_{rel}=1.0 \times 10^{-4}$		
				22V~220V,20kHz~100kHz	$U_{rel}=2.0 \times 10^{-4}$		
				22V~220V,100kHz~500kHz	$U_{rel}=5.0 \times 10^{-3}$		
				220V~1000V,50Hz~1kHz	$U_{rel}=1.0 \times 10^{-4}$		
		ACI		10 μ A~220mA,10Hz~40Hz	$U_{rel}=2.0 \times 10^{-4}$		
				10 μ A~220mA,40Hz~1kHz	$U_{rel}=1.5 \times 10^{-4}$		
				10 μ A~220mA,1kHz~5kHz	$U_{rel}=2.0 \times 10^{-4}$		
				220mA~2.2A,20Hz~1kHz	$U_{rel}=1 \times 10^{-4}$		



No. CNAS L0893

第 84 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date		
		DCI		220mA~2.2A,1kHz~5kHz	$U_{rel}=4.5\times 10^{-4}$				
				2.2A~50A,20Hz~1kHz	$U_{rel}=1\times 10^{-4}$				
				2.2A~20A,1kHz~5kHz	$U_{rel}=1.0\times 10^{-3}$				
				10 μ A~22mA	$U_{rel}=3.0\times 10^{-5}$				
				22mA~220mA	$U_{rel}=5.0\times 10^{-5}$				
				220mA~2.2A	$U_{rel}=5.8\times 10^{-5}$				
				2.2A~20A	$U_{rel}=6.5\times 10^{-5}$				
				1 Ω ~100 Ω	$U_{rel}=6.5\times 10^{-5}$				
		Resistive		100 Ω ~100k Ω	$U_{rel}=1.0\times 10^{-5}$				
				100k Ω ~10M Ω	$U_{rel}=5.0\times 10^{-5}$				
				10M Ω ~100M Ω	$U_{rel}=5.0\times 10^{-4}$				
65	Insulation resistance meterMegohm-meter	Resistance	Verification regulation of megohmmeter JJG 622	100 Ω ~100M Ω	$U_{rel}=1\%\sim 5\%$				
				100M Ω ~1T Ω	$U_{rel}=5\%\sim 10\%$				
		Voltage		10V~5000V	$U_{rel}=0.5\%\sim 5\%$				
66	Megohm-meter(High Resistance meter)	Resistance	Verification Regulation of High Insulation Resistance Meters JJG 690	100 Ω ~100M Ω	$U_{rel}=0.2\%\sim 2\%$				
				100M Ω ~10G Ω	$U_{rel}=2\%\sim 5\%$				
				10G Ω ~1T Ω	$U_{rel}=5\%\sim 10\%$				
				10V~1000V	$U_{rel}=0.5\%$				
67	High Voltage	Resistance	Verification Regulation of	10 ² Ω ~10 ⁶ Ω	$U_{rel}=0.001\%\sim 0.002\%$				



No. CNAS L0893

第 85 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	and Value D.C.Resistors		High Voltage and Value D.C.Resistors JJG 1072	10 ⁶ Ω ~ 10 ⁸ Ω	U _{rel} =0.002%~0.005%		
				10 ⁸ Ω ~ 10 ¹⁰ Ω	U _{rel} =0.005%~0.2%		
				10 ¹⁰ Ω ~ 10 ¹² Ω	U _{rel} =0.2%~5%		
68	PicoampereSource	Current	Calibration Specification for Micro Current Source(Micro Current Tester) NIMTT(CM) 019	10 ⁻⁴ A~10 ⁻⁷ A	U _{rel} =0.1%		
				10 ⁻⁷ A~10 ⁻⁸ A	U _{rel} =0.1%~0.5%		
				10 ⁻⁸ A~10 ⁻¹⁰ A	U _{rel} =0.15%~0.5%		
				10 ⁻¹⁰ A~10 ⁻¹¹ A	U _{rel} =0.5%~2%		
				10 ⁻¹¹ A~10 ⁻¹² A	U _{rel} =2%~5%		
69	Electronic Insulation Resistance Meters	Resistive	Verification Regulation of Electronic Insulating Resistance Meters JJG 1005	100 Ω ~ 100M Ω	U _{rel} =1%~5%		
		Voltage		100M Ω ~ 1T Ω	U _{rel} =5%~10%		
				10V~5000V	U _{rel} =0.5%~5%		
70	Calibration device of leakage current switch tester	DC Current	Calibration Specification for Calibration device of leakage current switch tester NIMTT(CM) 102	(5~3000)mA	U _{rel} =0.08%		
		AC Current		(5~3000)mA,50Hz	U _{rel} =0.10%		
		Time		(20~5000)ms	U=0.08ms		
71	Calibration device of Transformer with load tap changer	Resistance	Calibration Specification for Calibration device of Transformer with load tap changer NIMTT(CM) 094	(0.1~10) Ω	U _{rel} =0.06%		
		Time		(10~40) Ω	U _{rel} =0.10%		
				(1~10)ms	U _{rel} =2.5×10 ⁻⁵		
				(10~100)ms	U _{rel} =2.5×10 ⁻⁶		
				(100~10000)ms	U _{rel} =3×10 ⁻⁷		



No. CNAS L0893

第 86 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
72	Calibration device of Fluxmeter (Volt-Second generator)	DC voltage	Calibration Specification for Calibration device of Fluxmeter NIMTT(CM) 100	(10~100)mV	$U_{rel}=1 \times 10^{-4}$		
				(100~10000)mV	$U_{rel}=2 \times 10^{-5}$		
		Time		(1~10)ms	$U_{rel}=3 \times 10^{-5}$		
				(10~10000)ms	$U_{rel}=3 \times 10^{-6}$		
73	Multi Function Clamp Meter Calibrator	DC Voltage	Calibration Specification for Multi Function Clamp Meter Calibrator NIMTT(CM) 114	0.01 V~1000V	$U_{rel}=0.002\%$		
		AC Voltage		0.01V~1000V,(40~1000)Hz	$U_{rel}=0.01\%$		
		DC Current		0.1A ~1000A	$U_{rel}=0.02\%$		
		AC Current		0.1A ~2000A,(45~65)Hz	$U_{rel}=0.02\%$		
		Resistance		0.1 Ω ~100M Ω	$U_{rel}=0.01\% \sim 0.05\%$		
74	Insulating Oil electrical Strength Tester	Voltage	Calibration Specification for Dielectric strength detector of insulating oils NIMTT(CM) 126	5kV~100kV	$U_{rel}=0.6\%$		
		time		0.1s~600s	$U_{rel}=1.0\%$		
75	Withstanding Voltage calibration device	Voltage	Calibration specification for safety gauge NIMTT(CM) 131	(1~500)V,(50Hz)	$U_{rel}=0.05\%$		
				>0.5kV~100kV,(50Hz)	$U_{rel}=0.1\%$		
		Voltage		(1~500)V	$U_{rel}=0.05\%$		
				>0.5kV~100kV	$U_{rel}=0.1\%$		
		current		0.1mA~100A,(DC,50Hz)	$U_{rel}=0.1\%$		
		resistance		10m Ω ~500m Ω	$U_{rel}=0.2\%$		
				1M Ω ~1000M Ω	$U_{rel}=0.2\%$		



No. CNAS L0893

第 87 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Time		0.1s~999s	U _{rel} =0.1%		
76	Testing transformer	Voltage	Calibration specification for high voltage tester devices NIMTT(CM)133	1kV~100kV,(DC,50Hz)	U _{rel} =0.2%		
				>100kV~600kV,(DC,50Hz)	U _{rel} =1%		
		current		0.1mA~1A,(DC,50Hz)	U _{rel} =0.2%		
				>1A~1000A,(DC,50Hz)	U _{rel} =0.3%		
		Time		0.1s~600s	U _{rel} =1.0%		
77	Zinc oxide lightning arrester test instrument	Voltage	Calibration specification for Resistive current testers for zinc oxide surge arrester NIMTT(CM) 128	1V~200V,(50Hz)	U _{rel} =0.2%		
		current		0.01mA~20mA,(50Hz)	U _{rel} =0.2%		
78	DC Spark Detector(Needle-hole Detector)	Voltage	Calibration Specification for Electric Spark Leak Detectors NIMTT(CM) 123	0.01kV~50kV	U _{rel} =1.2%		
79	Impulse Voltage Testers for Winding Interturn Insulation	Voltage	Calibration Specification for Impulse Voltage Testers for Winding Interturn Insulation JJF 1691,(15~50)kV Calibration Specification for Impulse Voltage Testers for Winding Interturn Insulation NIMTT(CM) 155	0.1kV~50kV	U _{rel} =1.2%		
		time		0.1 μ s~100 μ s	U _{rel} =3%		
80		Voltage	calibration specification for Electrostatic analyzer NIMTT(CM) 115	1V~1kV	U _{rel} =0.1%		
				>1kV~20kV	U _{rel} =0.2%		
		time		0.1s~20s	U _{rel} =5%		



No. CNAS L0893

第 88 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
81	Concrete Resistivity Tester	resistivity	Calibration specification for Concrete Resistivity Tester NIMTT(CM) 148	(1~2000)kΩ • cm	<i>U</i> _{rel} =0.05%		
82	Digital Multimeter	DC Voltage	Calibration Specification for Multimeters JJF 1587	10mV~100mV	<i>U</i> _{rel} =5.5×10 ⁻⁶ ~1.0×10 ⁻⁵		
				100mV~1V	<i>U</i> _{rel} =4.5×10 ⁻⁶ ~5.0×10 ⁻⁶		
				1V~10V	<i>U</i> _{rel} =1.5×10 ⁻⁶ ~5.0×10 ⁻⁶		
				10V~100V	<i>U</i> _{rel} =2.7×10 ⁻⁶ ~3.6×10 ⁻⁶		
				100V~1000V	<i>U</i> _{rel} =3.7×10 ⁻⁶ ~4.5×10 ⁻⁶		
		AC Voltage		10mV~220mV,10Hz~40Hz	<i>U</i> _{rel} =2.5×10 ⁻⁴		
				10mV~220mV,40Hz~20kHz	<i>U</i> _{rel} =1.0×10 ⁻⁴		
				10mV~220mV,20kHz~100kHz	<i>U</i> _{rel} =3.4×10 ⁻⁴		
				10mV~220mV,100kHz~1MHz	<i>U</i> _{rel} =1.8×10 ⁻³		
				220mV~2.2V,10Hz~40Hz	<i>U</i> _{rel} =2.5×10 ⁻⁴		
				220mV~2.2V,40Hz~20kHz	<i>U</i> _{rel} =0.5×10 ⁻⁴		
				220mV~2.2V,20kHz~100kHz	<i>U</i> _{rel} =2.0×10 ⁻⁴		
				220mV~2.2V,100kHz~1MHz	<i>U</i> _{rel} =1.5×10 ⁻³		
				2.2V~22V,10Hz~40Hz	<i>U</i> _{rel} =2.5×10 ⁻⁴		



No. CNAS L0893

第 89 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

在线扫码获取验证


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		AC Current		2.2V~22V,40Hz~20kHz	$U_{rel}=0.5 \times 10^{-4}$		
				2.2V~22V,20kHz~100kHz	$U_{rel}=2.0 \times 10^{-4}$		
				2.2V~22V,100kHz~1MHz	$U_{rel}=5.0 \times 10^{-3}$		
				22V~220V,10Hz~40Hz	$U_{rel}=2.5 \times 10^{-4}$		
				22V~220V,40Hz~20kHz	$U_{rel}=1.0 \times 10^{-4}$		
				22V~220V,20kHz~100kHz	$U_{rel}=2.0 \times 10^{-4}$		
				22V~220V,100kHz~500kHz	$U_{rel}=5.0 \times 10^{-3}$		
				220V~1000V,40Hz~10kHz	$U_{rel}=1.0 \times 10^{-4}$		
				10 μ A~220mA,10Hz~40Hz	$U_{rel}=2.0 \times 10^{-4}$		
				10 μ A~220mA,40Hz~1kHz	$U_{rel}=1.5 \times 10^{-4}$		
				10 μ A~220mA,1kHz~5kHz	$U_{rel}=2.0 \times 10^{-4}$		
				220mA~2.2A,20Hz~1kHz	$U_{rel}=1 \times 10^{-4}$		
				220mA~2.2A,1kHz~5kHz	$U_{rel}=4.5 \times 10^{-4}$		
				2.2A~50A,20Hz~1kHz	$U_{rel}=1 \times 10^{-4}$		
				2.2A~20A,1kHz~5kHz	$U_{rel}=1.0 \times 10^{-3}$		



No. CNAS L0893

第 90 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		DC Current	ilac-MAS CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	50A~100A,40Hz~65Hz	$U_{rel}=1\times 10^{-4}$		
				10 μ A~22mA	$U_{rel}=3.0\times 10^{-5}$		
				22mA~220mA	$U_{rel}=5.0\times 10^{-5}$		
				220mA~2.2A	$U_{rel}=5.8\times 10^{-5}$		
				2.2A~20A	$U_{rel}=6.5\times 10^{-5}$		
				20A~100A	$U_{rel}=1\times 10^{-4}$		
		Resistances		(10 ⁻⁴ ~1) Ω	$U_{rel}=5\times 10^{-4}$		
				1 Ω ~100 Ω	$U_{rel}=5\times 10^{-5}$		
				100 Ω ~100k Ω	$U_{rel}=1.0\times 10^{-5}$		
				100k Ω ~10M Ω	$U_{rel}=5.0\times 10^{-5}$		
				10M Ω ~10G Ω	$U_{rel}=5.0\times 10^{-4}$		
83	*Amperemeters, Voltsmeters, Wattmeters	DC Current	V.R. of Amperemeters, Voltsmeters, Wattmeters and Ohmmeters JJG 124	10 μ A ~50A	$U_{rel}=0.05\%$		
		AC Current		100 μ A ~100A,40Hz~65Hz	$U_{rel}=0.05\%$		
				0.2A~10A,65Hz~5kHz	$U_{rel}=0.06\%$		
		DC Voltage		10mV~1000V	$U_{rel}=0.05\%$		
		AC Voltage		0.1V~1000V,40Hz~10kHz	$U_{rel}=0.05\%$		
		DC Power		0.1V~600V,10 μ A~100A	$U_{rel}=0.02\%$		
		AC Power		10V~600V,5mA~100A,(40~65)Hz	$U_{rel}=0.05\%$		



No. CNAS L0893

第 91 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				5V~600V,0.025A~10A, 65Hz~1kHz	<i>U</i> _{rel} =0.05%		
84	*DigitalAC./DC. Voltmeter	AC Voltage	V.R. of AC Digital VoltMeter JJG(JG) 72	0.01V~1000V,40Hz~10kHz	<i>U</i> _{rel} =0.01%		
85	*DigitalAC./DC. Amperemeter	AC Current	V.R. of AC Digital Current Meter JJG(JG) 68	100 μ A~100A,40Hz~65Hz	<i>U</i> _{rel} =0.01%		
				0.2A~10A,65Hz~5kHz	<i>U</i> _{rel} =0.05%		
86	Calibrators for Electrical Meters	DC Voltage	Calibration Specification of Calibrators for Electrical Meters JJF 1284	0.01 V~1000V	<i>U</i> _{rel} =0.001%		
		AC Voltage		0.01V~1000V,40Hz~10kHz	<i>U</i> _{rel} =0.01%		
		DC Current		10 μ A ~100A	<i>U</i> _{rel} =0.01%		
		AC Current		100 μ A ~100A,(40~65)Hz	<i>U</i> _{rel} =0.01%		
				0.2A~10A,65Hz~5kHz	<i>U</i> _{rel} =0.05%		
		Frequency		1Hz~100kHz	<i>U</i> _{rel} =1×10 ⁻⁵		
		Resistance		0.1 Ω ~10 Ω	<i>U</i> _{rel} =1%~0.05%		
				10 Ω ~1M Ω	<i>U</i> _{rel} =0.01%~0.002%		
				1M Ω ~100M Ω	<i>U</i> _{rel} =0.01%		
87	Electrical (Safety) Analyzer	AC Voltage	Calibration Specification for Electrical (Safety) Analyzer NIMTT(CM) 113	10V~300V,40Hz~65Hz	<i>U</i> _{rel} =0.02%		
		AC Current		0.1A ~20A,40Hz~65Hz	<i>U</i> _{rel} =0.02%		
		DC Leakage Current		10 μ A ~1A	<i>U</i> _{rel} =0.3%~0.05%		
		AC Leakage Current		30 μ A ~1A,40Hz~5kHz	<i>U</i> _{rel} =0.5% ~0.1%		



在线扫码获取验证

No. CNAS L0893

第 92 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		Resistance		0.01 Ω ~ 100 Ω	<i>U</i> _{rel} =5%~0.02%		
				100 Ω ~ 100M Ω	<i>U</i> _{rel} =0.3%~1.5%		
88	AC Digital Powermeter	AC Power	V.R. of AC Digital Powermeter JJG 780	3×3V~600V,0.005A~ 100A,(40~65)Hz	<i>U</i> _{rel} =0.01%		
				5V~600V,0.025A~ 10A,65Hz~1kHz	<i>U</i> _{rel} =0.05%		
89	*Clamp Meter	AC Current	Calibration Specification of Clamp-on Multifunctional Electricity Meter NIMTT(CM) 085	(0.001~1000)A45Hz~ 65Hz	<i>U</i> _{rel} =0.02%		
				(0.001~1000)A65Hz~ 1kHz	<i>U</i> _{rel} =0.3%		
		DC Current		(0.001~1000)A	<i>U</i> _{rel} =0.02%		
		AC Power		(0.1~1000)V,(1~ 1000)A(45Hz~65Hz)	<i>U</i> _{rel} =0.2%		
		DC Voltage		0.01V~1000V	<i>U</i> _{rel} =0.002%		
		AC Voltage		0.01V~1000V40Hz~ 10kHz	<i>U</i> _{rel} =0.01%		
		Resistance		10 Ω ~ 1M Ω	<i>U</i> _{rel} =0.01%~0.002%		
				1M Ω ~ 100M Ω	<i>U</i> _{rel} =0.01%		
		Phase		0° ~360° (16~69)Hz	<i>U</i> =0.005°		
				0° ~360° (69~450)Hz	<i>U</i> =0.03°		
		Frequency		10Hz~100kHz	<i>U</i> _{rel} =0.001%		
90	*Clamp Ammeters	AC Current	Calibration Specification for Clamp Ammeters JJF1075	(0.1~1000)A,(45~ 65)Hz	<i>U</i> _{rel} =0.02%		



No. CNAS L0893

第 93 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
91	Large current tester		Calibration Specification of Large Current Tester NIMTT(CM) 087	(0.1~1000)A,(65~400)Hz	$U_{rel}=0.3\%$		
		DC Current		(0.1~2000)A	$U_{rel}=0.02\%$		
		DC Current		(0.001~10000)A	$U_{rel}=0.005\%$		
		AC Current		(0.001~0.1)A,(45~65)Hz	$U_{rel}=0.012\%$		
92	Single phase and three phase multifunction calibration source	AC Current	Calibration method for single phase and three phase multifunction calibration source (calibration method for multi-functional instrument calibration source) NIMTT(CM) 079, Calibration Specification for Verification Equipment of Electrical Measuring Devices JJF1923	(0.1~10000)A,(45~65)Hz	$U_{rel}=0.005\%$		
		Time		0.5s~10s	0.01s		
		DC Voltage		0.01V~1000V	$U_{rel}=0.001\%$		
		AC Voltage		0.01V~1000V40Hz~10kHz	$U_{rel}=0.01\%$		
		DC Current		10 μ A~100A	$U_{rel}=0.01\%$		
		AC Current		100 μ A~100A(40~65)Hz	$U_{rel}=0.01\%$		
		AC Power		0.2A~10A65Hz~10kHz	$U_{rel}=0.05\%$		
		DC Power		3×3V~600V/0.005A~100A(40~65)Hz	$U_{rel}=0.01\%$		
		Frequency		5V~600V/0.025A~10A65Hz~1kHz	$U_{rel}=0.05\%$		
		Resistance		(3~1000)V/(0.001~500)A	$U_{rel}=0.02\%$		
				1Hz~100kHz	$U_{rel}=1\times 10^{-5}$		
				0.1 Ω ~10 Ω	$U_{rel}=1\%\sim 0.05\%$		



No. CNAS L0893

第 94 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				10 Ω ~ 1M Ω	<i>U</i> _{rel} =0.01%~0.002%		
				1M Ω ~ 100M Ω	<i>U</i> _{rel} =0.01%		
		Phase		0° ~ 360° (16~69)Hz	<i>U</i> =0.005°		
				0° ~ 360° (69~450)Hz	<i>U</i> =0.03°		
		Harmonic Voltage		(0.03~300)V,(2~60)times	<i>U</i> _{rel} =0.012%~0.05%		
		Harmonic Current		(0.005~30)A,(2~60)times	<i>U</i> _{rel} =0.012%~0.05%		
93	Grounding (conduction) resistance tester	resistance	Earth-continuity testers JJG984	0.1m Ω ~ 11 Ω	<i>U</i> _{rel} =0.05%~0.2%		
		current		(1~100)A	<i>U</i> _{rel} =0.05%		
		voltage		10mV~20V	<i>U</i> _{rel} =0.1%		
94	field meter	Low Frequency Electric-Field Strength	Calibration Specification for Electric Field meters JJF(J11)154,Calibration Specification for Electromagnetic Field Probes from 10kHz to 100MHz JJF 1884,Calibration Specification for Electric Field Probes JJF 1886	10H~400kHz1V/m~100V/m	<i>U</i> =0.46~0.83dB		
		high Frequency Electric-Field Strength		(0.1~10)kV/m(50Hz)	<i>U</i> =0.46~0.62dB		
				10kHz~100MHz0.2V/m~300V/m	<i>U</i> =0.6~0.8dB		
				100MHz~1GHz0.01A/m~1A/m	<i>U</i> =1.04~1.33dB		
				10kHz~100MHz0.01A/m~1A/m	<i>U</i> =0.6~0.8dB		
				1GHz~18GHz0.1V/m~300V/m	<i>U</i> =0.76~1.25dB		
				100MHz~1GHz1V/m~300V/m	<i>U</i> =1.04~1.33dB		



No. CNAS L0893

第 95 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
95	DC total electric field strength measuring instrument	Electric-Field Strength	Calibration Specification for DC total electric field strength measuring instrument NIMTT(CM) 009	$(-40 \sim 40) \text{ kV/m}$	$U_{\text{rel}}=4.2\%$		
96	AC charging sport tester	Alternating-current Electrical Energy	C.S. for AC charging sport tester NIMTT(CM) 092	$57.7\text{V} \sim 380\text{V}, 0.01\text{A} \sim 100\text{A}$	$U_{\text{rel}}=1.2 \times 10^{-4}$		
		AC voltage		$57.7\text{V} \sim 380\text{V}, 50\text{Hz}$	$U_{\text{rel}}=1.2 \times 10^{-4}$		
		AC current		$0.01\text{A} \sim 100\text{A}, 50\text{Hz}$	$U_{\text{rel}}=1.2 \times 10^{-4}$		
		AC power		$57.7\text{V} \sim 380\text{V}, 0.01\text{A} \sim 100\text{A}, 50\text{Hz}$	$U_{\text{rel}}=1.2 \times 10^{-4}$		
97	Off-board charger tester	Electrical Energy	C.S. for Off-board charger tester NIMTT(CM) 086	$100\text{mV} \sim 1000\text{V}, 100 \mu\text{A} \sim 400\text{A}$	$U_{\text{rel}}=1.2 \times 10^{-4}$		
		DC voltage		$100\text{mV} \sim 1000\text{V}$	$U_{\text{rel}}=1.2 \times 10^{-4}$		
		DC current		$100 \mu\text{A} \sim 400\text{A}$	$U_{\text{rel}}=1.2 \times 10^{-4}$		
		DC power		$100\text{mV} \sim 1000\text{V}, 100 \mu\text{A} \sim 400\text{A}$	$U_{\text{rel}}=1.2 \times 10^{-4}$		
98	*DC Standard Resister	Resistances	Verification Regulation of DC Standard Resister JJG982, Calibration specification of Calibrators for DC Low-resistance Meters JJF(J1)173	$(10^{-6} \sim 10^{-5}) \Omega$	$U_{\text{rel}}=3 \times 10^{-3}$		
		Resistances		$(10^{-5} \sim 10^{-4}) \Omega$	$U_{\text{rel}}=1 \times 10^{-3}$		
		Resistances		$(10^{-4} \sim 10^{-2}) \Omega$	$U_{\text{rel}}=1 \times 10^{-4}$		
		Resistances		$(10^{-2} \sim 1) \Omega$	$U_{\text{rel}}=3 \times 10^{-5}$		
		Resistances		$(1 \sim 10^7) \Omega$	$U_{\text{rel}}=1 \times 10^{-5}$		
		Resistances		$(10^7 \sim 10^8) \Omega$	$U_{\text{rel}}=5 \times 10^{-4}$		



No. CNAS L0893

第 96 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
99	Hall Current (Voltage) Transduce	DC Current	Hall Current (Voltage) Transduce JJG(Ⅲ)136	(0.001～10000)A	<i>U</i> _{rel} =0.005%		
		AC Current		(0.001～0.1)A,(45～65)Hz	<i>U</i> _{rel} =0.012%		
				(0.001～1000)A,65Hz～1kHz	<i>U</i> _{rel} =0.3%		
				(0.1～10000)A,45Hz～65Hz	<i>U</i> _{rel} =0.005%		
				DC Voltage	10mV～1000V		<i>U</i> _{rel} =0.002%
		AC Voltage		10mV～1000V,45Hz～1kHz	<i>U</i> _{rel} =0.005%		
100	*Electric recorders	DC Voltage	Calibration Specification of Electric recorders NIMTT(CM) 150	0.01V～1000V	<i>U</i> _{rel} =0.001%		
		AC Voltage		0.01V～1000V,45Hz～10kHz	<i>U</i> _{rel} =0.01%		
		DC Current		10 μ A～20A	<i>U</i> _{rel} =0.01%		
		AC Current		10mA～20A,(45～65)Hz	<i>U</i> _{rel} =0.01%		
		AC Current		0.2A～10A,65Hz～1kHz	<i>U</i> _{rel} =0.05%		
		Frequency		10Hz～100kHz	<i>U</i> _{rel} =0.001%		
101	*Current Clamp、 Current Coil	DC Current	Calibration Specification for Current Clamp、Current Coil NIMTT(CM) 151	(0.001～10000)A	<i>U</i> _{rel} =0.1%		
		AC Current		(0.001～10000)A45Hz～65Hz	<i>U</i> _{rel} =0.1%		
		AC Current		(0.001～1000)A65Hz～1kHz	<i>U</i> _{rel} =0.3%		
102	Electronic Meters for Measuring Direct- current Electrical	Electrical Energy	Electronic Meters for Measuring Direct-current Electrical Energy JJG 842	100mV～1000V10mA～500A	<i>U</i> _{rel} =3×10 ⁻⁴		

No. CNAS L0893

第 97 页 共 181 页



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	Energy	time		$(-99.99 \sim 99.99) \text{ s/d}$	$U=0.06\text{s/d}$		
103	*Off-board Charger for Electric Vehicles	Electrical Energy	Off-board Charger for Electric Vehicles JJG1149	$100\text{V} \sim 1000\text{V}$ $5\text{A} \sim 250\text{A}$	$U_{\text{rel}}=1 \times 10^{-3}$		
104	*AC Charging Spot for Electric Vehicles	Electrical Energy	AC Charging Spot for Electric Vehicles JJG1148	220V $1\text{A} \sim 70\text{A}$	$U_{\text{rel}}=1 \times 10^{-3}$		
105	Calibration Device of Power Capacitance Inductance Tester	Capacitance	Calibration Specification for Calibration Device of Power Capacitance Inductance Tester NIMTT(CM) 152	$1\mu\text{F} \sim 25000\mu\text{F}$	$U_{\text{rel}}=0.02\%$		
		Inductance		$1\text{mH} \sim 10\text{H}$	$U_{\text{rel}}=0.02\%$		
106	Calibration Device of Zinc Oxide Arrester Tester	AC Voltage	Calibration Specification for Calibration Device of Zinc Oxide Arrester Tester NIMTT(CM) 153	$1\text{V} \sim 250\text{V}$ 50Hz	$U_{\text{rel}}=0.02\%$		
		AC Current		$0.1\text{mA} \sim 20\text{mA}$ 50Hz	$U_{\text{rel}}=0.02\%$		
		Phase		$0^\circ \sim 360^\circ$, 50Hz	$U=0.01^\circ$		
107	Textile Frictional Static Charges Tester	charge quantity	Calibration Specification for Textile Frictional Static Charges Tester JJF(纺织)071	$(0.01 \sim 2) \mu\text{C}$	$U_{\text{rel}}=0.1\%$		
108	Eddy Current Conductivity Meters	conductivity	Calibration Specification for Eddy Current Conductivity Meters JJF 1692	$(1.05 \sim 95)\% \text{ IACS}$	$U=(0.05 \sim 0.5)\% \text{ IACS}$		
109	power transformers winding deformation tester	frequency	Calibration specification of power transformers winding deformation tester DL/T 1952	$1\text{kHz} \sim 1\text{MHz}$	$U_{\text{rel}}=0.01\%$		
		amplitude radio		$(0, -10, -20, -40, -60, -80)\text{dB}$	$U=(0.01 \sim 0.3)\text{dB}$		
110	*DC high- current source	Current	DC high-current source NIMTT(CM) 154	$(1000 \sim 10000)\text{A}$	$U_{\text{rel}}=1 \times 10^{-5}$		
111	*Buchholz relay & overspeed shut-off	velocity of flow	Buchholz relay & overspeed shut-off valve detection	$(1 \sim 100)\text{m}^3/\text{h}$ $((0.566 \sim 5.662)\text{m/s})$	$U_{\text{rel}}=1\%$		



No. CNAS L0893

第 98 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	valve detection device	volume	device NIMTT(CM) 072	(10~1000)mL	U _{rel} =1.2%		
		pressure		(10~250)kPa	U _{rel} =1%		
112	Medical Diagnostic X-ray Non-invasive Current Meters	Current	Calibration Specification for Medical Diagnostic X-ray Non-invasive Current Meters JJF 1473	0.1mA~1A	U _{rel} =0.4%		
113	*Magnetic particle Flaw Detectors	Current	Calibration Specification for Magnetic Particle Flaw Detectors JJF 1273	(0.1~10000) A _s (DC,50Hz)	U _{rel} =1.5%		
		Magnetic induction intensity		(0~1)mT	U=0.04 mT		
		Illuminance		(50~3000)lx	U _{rel} =1.6%~3.2%		
		Ultraviolet Irradiance (UVA)		(1~10000) μ W/cm2	U _{rel} =19%		
114	Magnetic Yoke Detectors	Irradiance	Calibration Specification for Magnetic Yoke Detectors JJF 1458	(0.1~500) N	U _{rel} =4.0%		
		Current		(0.1~40)A _s (DC,50Hz)	U _{rel} =2.2%~5.1%		
115	Eddy Current Flaw Detector	Frequency	V.R. of Eddy Current Flaw Detector JJG(MH) 0061	1Hz~10MHz	U _{rel} =0.02%		
		Voltage		(0.1~10) V _s (1kHz~1MHz)	U _{rel} =1.2%		
VI Radio measuring instrument							
1	Electronic Voltmeter	Frequency response	Verification Regulation of Electronic Voltmeter JJG250, Calibration Specification for Low-frequency Volt-meters JJF1925	10Hz~500MHz	U _{rel} =3%		
		Voltage		1mV~1V	U _{rel} =0.5%		
				1V~300V	U _{rel} =1%		



No. CNAS L0893

第 99 页 共 181 页

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
2	UHF Electronic Mill voltmeter	Voltage	Verification Regulation of RF Voltmeters JJG308	1 mV~1 V (DC~1000MHz)	$U_{rel}=1\%$		
				1 V~10 V (DC~1000MHz)	$U_{rel}=2\%$		
3	Frequency Response Test Set	Voltage	Verification Regulation of 300 MHz Frequency Response Test Set JJG359	50mV~300mV (1MHz~300MHz)	$U_{rel}=5\%$		
4	Network analyzer	Reflection	Verification Regulation for automatic network analyzer GJB/J3608, Calibration Specification for Vector network analyzers JJF1495	1.00~2.00(50MHz~ 50GHz)	$U=0.02\sim0.06$		
		Transmission		(0~50)dB(50MHz~ 50GHz)	$U=0.20\text{dB}\sim0.50\text{dB}$		
		Source output frequency		10Hz~40GHz	$U_{rel}=1\times10^{-9}$		
		Source output power		(-80~+20)dBm	$U=0.15\text{dB}\sim0.20\text{dB}$		
		Background noise		(-160~-50) dB	$U=2.3\text{dB}\sim3.0\text{dB}$		
		Crosstalk		(-160~-50) dB	$U=1.1\text{dB}\sim3.0\text{dB}$		
		Modulus value Trace noise		(0~1)dB	$U=0.0013\text{dB}$		
		Phase Trace noise		(0~1) °	$U=0.012^\circ$		
5	Underground pipeline detector	frequency	calibration specification for Underground pipeline detector NIMTT(CM) 010	20Hz~1MHz	$U_{rel}=4.3\times10^{-6}$		



No. CNAS L0893

第 100 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
6	Digital Signal Generator	Level	Calibration Specification for Digital Signal Generator JJF1174	20dBm~-120dBm (100kHz~6GHz)	U=0.2dB~0.5dB		
		Frequency		100kHz~6GHz	U _{rel} =1×10 ⁻⁹		
		Harmonics		0dBc~100dBc (100kHz~6GHz)	U=0.5dB~2dB		
		Vector Magnitude		0.5%~18%(FSK、MSK、16QAM)	U=0.5%		
		RMS Phase		0.5°~15° (FSK、MSK、16QAM)	U=0.5°		
		RMS frequency		1kHz~250kHz(FSK、MSK、16QAM)	U=2Hz		
7	Data Acquisition System	Voltage	Calibration Specification for Data Acquisition System JJF1048	1mV~1V	U _{rel} =0.3%		
				1V~100V	U _{rel} =0.2%		
		Frequency		(0.1~250)MHz	U _{rel} =4×10 ⁻⁸		
				250MHz~2.7GHz	U _{rel} =6×10 ⁻⁸		
8	Standard Capacitors	Capacitance	Verification Regulation of Standard Capacitors JJG183	1pF~1 μ F (1 kHz)	U _{rel} =2×10 ⁻⁵		
				(1~99999) μ F (100 Hz)	U _{rel} =5×10 ⁻⁴		
9	Standard Inductors	Inductance	Verification Regulation of Standard Inductors JJG726	1 μ H~10 μ H (1kHz)	U _{rel} =2%		
				10 μ H~100 μ H (1kHz)	U _{rel} =0.1%		
				100 μ H~1H (1kHz)	U _{rel} =0.01%		
				1H~99999H (100Hz)	U _{rel} =0.05%		
	Alternating	Resistances	Verification Regulation of	1 Ω ~1M Ω (1kHz)	U _{rel} =0.01%		

No. CNAS L0893

第 101 页 共 181



The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
	Current Bridge (Capacitance Bridge)	Inductances	Alternating Current Bridge JJG441	10 μ H~1H (1kHz)	<i>U</i> _{rel} =0.01%		
		Capacitances		1pF~1 μ F (1kHz)	<i>U</i> _{rel} =0.01%		
11	High Voltage Capacitance Bridge	ratio of capacitances	Verification Regulation of High Voltage Capacitance Bridges JJG563	X:(0.1~1)	<i>U</i> _{rel} =0.002%		
		dissipation factor		X: (1~1000)	<i>U</i> _{rel} =0.2%		
					1×10 ⁻⁴ ~0.1		<i>U</i> _{rel} =0.2%
12	Dielectric Dissipation Factor Tester	capacitances	High Voltage Standard Capacitors JJG1075	10pF~100nF	<i>U</i> _{rel} =0.05%		
		dissipation factor		1×10 ⁻⁵ ~1×10 ⁻³	<i>U</i> =6×10 ⁻⁵		
13	High frequency Q meter	Q	Calibration Specification of HF Q Meter JJF1073	10~500	<i>U</i> _{rel} =5%~10%		
		Frequency		50kHz~500kHz	<i>U</i> _{rel} =2%		
					500kHz~50MHz		<i>U</i> _{rel} =2.4%
14	Transistor Character Scope	Voltage	Calibration Specification for Semiconductor Device Curve Tracers JJF1236	(0.1~1)V	<i>U</i> _{rel} =1%		
				(1~200)V	<i>U</i> _{rel} =0.5%		
		Current		200 μ A~1mA	<i>U</i> _{rel} =0.8%		
				1mA~1A	<i>U</i> _{rel} =0.6%		
				1A~10A	<i>U</i> _{rel} =0.5%		
15	Radio Communication Te ster	RF frequency	Ca1ibration Specification for RF Communication Test Set JJF1065	20kHz~6GHz	<i>U</i> _{rel} =2×10 ⁻⁹		
		RF power		20dBm~- 120dBm(20kHz~6GHz)	<i>U</i> =0.2dB		



No. CNAS L0893

第 102 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Frequency modulation (FM)	ilac-MRA CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT SCHEDULE OF ACCREDITATION CERTIFICATE	20Hz~200kHz	$U_{rel}=3\%$		
		amplitude modulation (AM)		1%~99%	$U_{rel}=3\%$		
		AF frequency		10Hz~20kHz	$U_{rel}=2 \times 10^{-9}$		
		AF level		0.1V~750V(10Hz~20kHz)	$U_{rel}=0.1\%$		
		DC level		0.1V~1000V	$U_{rel}=0.01\%$		
16	Bluetooth tester	output frequency	Calibration Specification for Bluetooth Test Set JJF1278	0.1Hz~3GHz	$U_{rel}=5 \times 10^{-8}$		
		output level		(-127~30) dBm(0.1Hz~3GHz)	$U=0.5\text{dB}$		
		Output signal frequency response		(-60~-20)dBm(0.1Hz~3GHz)	$U=0.5\text{dB}$		
		Output signal harmonic		(0~-120)dBc(0.1Hz~3GHz)	$U=1\text{dB}$		
		Modulation frequency deviation		10Hz~350kHz(GFSK、8DPSK)	$U_{rel}=2\%$		
17	Wireless LAN tester	output frequency	Calibration Specification for WLAN Test Set JJF1277	0.01Hz~13GHz	$U_{rel}=5 \times 10^{-8}$		
		output level		(-127~30)dBm(0.01Hz~13GHz)	$U=0.5\text{dB}$		
		Output signal harmonic		(0~-120)dBc	$U=1\text{dB}$		



No. CNAS L0893

第 103 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
18	CDMA digital mobile communication comprehensive test instrument	Single side band phase noise of output signal	Calibration Specification for CDMA Digital Radio Communication Testers JJF1177	(0~-120)dBc/Hz	$U=1\text{dB}$		
		modulation		5%~20%	$U_{\text{rel}}=2\%$		
		Level		(-127~30)dBm(0.01Hz~13GHz)	$U=0.3\text{dB}$		
		Vector Magnitude		0.5%~18%	$U_{\text{rel}}=1\%$		
		output frequency		0.1GHz~3GHz	$U_{\text{rel}}=2 \times 10^{-8}$		
		output level		(-120~-5)dBm(0.1GHz~3GHz)	$U=0.5\text{dB}$		
		Output signal harmonic		(0~-120)dBc	$U=0.8\text{dB}$		
		Single side band phase noise of output signal		(0~-120)dBc/Hz	$U=1.0\text{dB}$		
		CDMA signal generator Rho		0.9~1.0(0.1GHz~3GHz)	$U=0.0005$		
		CDMA generator EVM		1.5%~50%(0.1GHz~3GHz)	$U=2.0\%$		
		CDMA power level		(-75~-5)dBm	$U=0.5\text{dB}$		
		Audio generator frequency		100Hz~50kHz	$U=0.0008\text{Hz}$		



在线扫码获取验证

No. CNAS L0893

第 104 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Audio generator level		1V~5V(100Hz~5kHz)	$U_{rel}=0.04\%$		
19	TD-SCDMA digital mobile communication comprehensive test instrument	output frequency	Calibration Specification for TD-SCDMA Digital Radio Communication Testers JJF1204	30MHz~2.7GHz	$U_{rel}=2 \times 10^{-8} < / - 8 >$		
		output level		(-120~-10)dBm(30MHz~2.7GHz)	$U=0.5\text{dB}$		
		Output signal harmonic		(0~-120)dBc	$U=0.8\text{dB}$		
		Single side band phase noise of output signal		(0~-120)dBc/Hz	$U=1.0\text{dB}$		
		CDMA signal generator Rho		0.9~1.0(30MHz~2.7GHz)	$U=0.0005$		
		CDMA generator EVM		1.5%~50%(30MHz~2.7GHz)	$U=2.0\%$		
		Audio generator frequency		100Hz~5kHz	$U=0.0008\text{Hz}$		
		Audio generator level		1V~5V(100Hz~5kHz)	$U_{rel}=0.06\%$		
20	*artificial mains network	Voltage Division Factor	Calibration Specification of Artificial Mains Networks JJF 1705 2018	0dB~30dB (9kHz~108MHz)	$U=1.6\text{dB}$		
		Impedance		0 Ω ~300 Ω (9kHz~108MHz)	$U_{rel}=6\%$		
		phase		0° ~360°	$U=2.8^\circ$		



在线扫码获取验证

No. CNAS L0893

第 105 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
21	*coupling/Decoupling network	Insert Loss	calibration specification of Coupling/decoupling network JJF(苏)213-2018	0dB~30dB (150kHz~200MHz)	$U=0.8\text{dB}$		
		common-mode impedance		10 Ω ~300 Ω (150kHz~200MHz)	$U_{\text{rel}}=7.3\%$		
22	Standard Dissipation Factor	capacitance	High Voltage Standard Capacitors JJG 1075, Calibration Regulation of Standard Dissipation Factor NIMTT(CM) 050	10pF~500nF	$U_{\text{rel}}=0.05\%$		
		dissipation factor		$1 \times 10^{-5} \sim 1 \times 10^{-1}$	$U=6 \times 10^{-5} \sim 6 \times 10^{-4}$		
23	*Dielectric Dissipation Factor Tester	capacitances	High Voltage Dielectric Loss Tester JJG 1126	100pF~500nF	$U_{\text{rel}}=0.1\%$		
		dissipation factor		0.000%~10%	$U=0.5\% \times \tan \delta + 5 \times 10^{-5}$		
24	Cable and Antenna Analyzers	Level	Calibration Specification for SWR Tester JJF(川) 137, Calibration Specification for Cable and Antenna Analyzers JJF 1740	(-70~20)dBm(2MHz~18GHz)	$U=0.07\text{dB}$		
		Frequency		2MHz~18GHz	$U_{\text{rel}}=2 \times 10^{-7}$		
		Attenuation		(-30~0)dB(2MHz~18GHz)	$U=0.05\text{dB}$		
		SWR		1.00~2.00(2MHz~18GHz)	$U=0.02 \sim 0.06$		
25	Spectrum Analyzer	Frequency	Calibration Specification for Spectrum JJF 1396	20Hz~50GHz	$U_{\text{rel}}=3 \times 10^{-10}$		
		Amplitude		(-30~20)dBm(20Hz~50GHz)	$U=0.12\text{dB} \sim 0.5\text{dB}$		
		Amplitude		(0~80)dB(20Hz~50GHz)	$U=0.08\text{dB} \sim 0.5\text{dB}$		
		Sweep Width		1kHz~50GHz	$U_{\text{rel}}=2 \times 10^{-4}$		



No. CNAS L0893

第 106 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
26	Attenuator	Attenuation	Verification Regulation of Coaxial Attenuator JJG 387, Verification Regulation of Variable Attenuator JJG 737	(0~120)dB (250kHz~26.5GHz)	$U=0.09\text{dB} \sim 0.14\text{dB}$		
		Attenuation		(0~50)dB (26.5GHz~40GHz)	$U=0.14\text{dB} \sim 0.30\text{dB}$		
		Attenuation		(0~50)dB (40GHz~50GHz)	$U=0.30\text{dB} \sim 0.40\text{dB}$		
		SWR		1.00~2.00	$U=0.02 \sim 0.06$		
27	Power divider/Coupler	Insertion loss	Calibration Specification for Microwave Components	0~80dB(250kHz~40GHz)	$U=0.2\text{dB}$		
		Standing wave ratio	NIMTT(CM) 101, Calibration Specification for Directional Coupler and SWR Bridges JJF 1680	1.00~2.00(250kHz~40GHz)	$U=0.02 \sim 0.06$		
28	Low Frequency Signal Generator	Voltage	Verification Regulation of Low Frequency Signal Generator JJG 602	1mV~1V	$U_{\text{rel}}=3\%$		
		Voltage		1V~300V	$U_{\text{rel}}=1.5\%$		
		Frequency		10mHz~1MHz	$U_{\text{rel}}=1 \times 10^{-7}$		
		Distortion		0.01%~20%	$U_{\text{rel}}=3\%$		
29	Signal Generator	Level	Calibration Specification for Signal Generators JJF 1931	20dBm~-120dBm (20Hz~40GHz)	$U=0.16\text{dB} \sim 0.50\text{dB}$		
		Frequency		20Hz~40GHz	$U_{\text{rel}}=5 \times 10^{-10}$		
		Amplitude modulation		6%~99%(fm:50Hz~400kHz)	$U_{\text{rel}}=0.7\% \sim 1.8\%$		
		Frequency modulation		4kHz~400kHz(fm:10Hz~200kHz)	$U_{\text{rel}}=0.6\% \sim 3.5\%$		
		Phase modulation		(4~400)rad (50Hz~100kHz)	$U_{\text{rel}}=0.6\% \sim 1.2\%$		



No. CNAS L0893

第 107 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
30	Distortion Meter	Distortion	Calibration Specification for Distortion Meters JJF 1852	0.3%~100%(10Hz~10kHz)	$U_{rel}=1\%$		
		Distortion		0.3%~100%(10kHz~100kHz)	$U_{rel}=3\%$		
		Distortion		0.3%~100%(100kHz~200kHz)	$U_{rel}=3.5\%$		
		Distortion		0.003%~0.3%(10Hz~10kHz)	$U_{rel}=6\%$		
		Distortion		0.003%~0.3%(10kHz~100kHz)	$U_{rel}=6.5\%$		
		Distortion		0.003%~0.3%(100kHz~200kHz)	$U_{rel}=9\%$		
		Voltage		(0.001~300)V	$U_{rel}=1\%$		
31	Digital Oscilloscope	Vertical deflection coefficient	Calibration Specification for Digital Storage Oscilloscope JJF 1057, Verification Regulation of Digital Oscilloscope GJB 7691	(1mV~20V)/div	$U_{rel}=0.6\%$		
		Frequency bandwidth		50kHz~50GHz	$U_{rel}=2.5\%$		
		Rise time		17ps~500ns	$U_{rel}=2\%$		
		Sweep time		200ps~5s	$U_{rel}=0.6\%$		
32	Analog Oscilloscope	Vertical deflection coefficient	Verification Regulation of Analogue Oscilloscope JJG 262	(1mV~20V)/div	$U_{rel}=0.6\%$		
		Frequency bandwidth		50kHz~500MHz	$U_{rel}=2.5\%$		
		Rise time		350ps~500ns	$U_{rel}=20\%$		
		Sweep time		200ps~5s	$U_{rel}=0.6\%$		



No. CNAS L0893

第 108 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
33	Sampling Oscilloscope	Vertical deflection coefficient	Verification Regulation of 1 GHz Sampling Oscilloscope JJG 491	10mV~200V	$U_{rel}=0.6\%$		
		Frequency bandwidth		50kHz~50GHz	$U_{rel}=3\%$		
		Time base		200ps~5s	$U_{rel}=0.6\%$		
		Rise time		17ps~500ns	$U_{rel}=2\%$		
34	Oscilloscope Calibrator	Voltage	Verification Regulation of Oscilloscope Calibrator JJG 278	1mV~200V	$U_{rel}=0.05\%$		
		Time base		25ps~5s	$U_{rel}=4 \times 10^{-8}$		
		Rise Time		25ps~500ns	$U_{rel}=3\%$		
35	Pulse Generator	Frequency	Verification Regulation of Pulse Generator JJG 490	0.1Hz~3GHz	$U_{rel}=3.4 \times 10^{-8}$		
		Pulse wide		10000s~1ns	$U_{rel}=0.35\%$		
		Amplitude		1mV~100V	$U_{rel}=1.0\%$		
		Rise Time		17ps~500ns	$U_{rel}=3\%$		
36	Audio Analyzer	Voltage	Calibration Specification for Audio Analyzer JJF 1395	(1~220)mV(10Hz~1MHz)	$U_{rel}=0.07\% \sim 0.35\%$		
				(0.22~22)V(10Hz~1MHz)	$U_{rel}=0.06\% \sim 0.28\%$		
				(22~220)V(10Hz~1MHz)	$U_{rel}=0.06\% \sim 1.16\%$		
				(220~300)V(15Hz~1kHz)	$U_{rel}=0.04\% \sim 0.01\%$		
		Distortion		0.003%~0.05%(10Hz~10kHz)	$U_{rel}=10\% \sim 2\%$		



No. CNAS L0893

第 109 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date		
		Source Frequency		0.05%~0.3%(10Hz~200kHz)	<i>U</i> _{rel} =1%~10%				
				0.3%~100%(10Hz~200kHz)	<i>U</i> _{rel} =1%~4%				
				10Hz~1MHz	<i>U</i> _{rel} =1×10 ⁻⁶				
		Source Voltage		(1~220)mV(10Hz~1MHz)	<i>U</i> _{rel} =0.08%~0.40%				
				(0.22~22)V(10Hz~1MHz)	<i>U</i> _{rel} =0.10%~0.48%				
				(22~220)V(10Hz~1MHz)	<i>U</i> _{rel} =0.08%~1.26%				
				(220~300)V(15Hz~1kHz)	<i>U</i> _{rel} =0.04%~0.01%				
37	Measuring Receiver	Frequency	Calibration Specification for Measuring Receivers JJF 1173	250kHz~67GHz	<i>U</i> _{rel} =1.3×10 ⁻⁹				
		Level		(0~-120)dB (250kHz~50GHz)	<i>U</i> =0.006dB~0.15dB				
		Amplitude modulation		5%~99%	<i>U</i> _{rel} =0.2%~0.5%				
		Frequency modulation		1Hz~400kHz	<i>U</i> _{rel} =0.2%~0.5%				
		Phase modulation		(1~400)rad	<i>U</i> _{rel} =0.5%~2%				
38	EMI Receiver	Frequency	Calibration Specification for EMI Testing Receivers JJF1144	9kHz~40GHz	<i>U</i> _{rel} =1×10 ⁻⁸				
		Level		(0~125)dB μ V	<i>U</i> =0.3dB				
		Attenuation		(0~100)dB	<i>U</i> =0.2dB				
		Standing wave ratio		1.00~2.00	<i>U</i> =0.02				



No. CNAS L0893

第 110 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Impulse frequency response		(0~70)dB	$U=1.2\text{dB}$		
39	LCR Digital Bridge(LCR Meter)	Resistances	Verification regulation for wide range digital RLC meter GJB/J5412	$1\ \Omega \sim 1\text{M}\ \Omega$	$U_{\text{rel}}=0.01\%$		
		Inductances		$10\ \mu\text{H} \sim 100\ \mu\text{H}$	$U_{\text{rel}}=0.05\%$		
		Inductances		$100\ \mu\text{H} \sim 1\text{H}$	$U_{\text{rel}}=0.01\%$		
		Inductances		$1\text{H} \sim 9999\text{H}$	$U_{\text{rel}}=0.05\%$		
		Capacitances		$1\text{pF} \sim 1\ \mu\text{F}$	$U_{\text{rel}}=0.01\%$		
		Capacitances		$1\ \mu\text{F} \sim 9999\ \mu\text{F}$	$U_{\text{rel}}=0.05\%$		
		Dielectric Dissipation Factor		$1 \times 10^{-5} \sim 1$	$U=5 \times 10^{-5}$		
40	Power Meter	Standing wave ratio	Verification Regulation of Lower Power Mount	$1.00 \sim 2.00(9\text{kHz} \sim 40\text{GHz})$	$U=0.02 \sim 0.07$		
		Calibration factor	GJB/J3598, Calibration Specification for RF &	$30\% \sim 200\%(250\text{kHz} \sim 40\text{GHz})$	$U_{\text{rel}}=2.0\% \sim 3.0\%$		
		Calibration factor	Microwave Power Sensors JJF1887	$30\% \sim 200\%(40\text{GHz} \sim 50\text{GHz})$	$U_{\text{rel}}=3.0\% \sim 5.0\%$		
41	Function Signal Generator	Frequency	Verification Regulation of Function Generators	$0.1\text{Hz} \sim 2.7\text{GHz}$	$U_{\text{rel}}=4 \times 10^{-8}$		
		Voltage	JJG840, Calibration Specification for Arbitrary	$1\text{mV} \sim 55\text{V}$	$U_{\text{rel}}=1\%$		
		Distortion	Waveform Ge-nerator JJF1152	$0.003\% \sim 30\%$	$U_{\text{rel}}=1\% \sim 6\%$		
42	*Insulating Oil Dielectric	Capacitors	Calibration Specification for Insulating Oil Dielectric Dissipation Factor and	100pF	$U_{\text{rel}}=0.1\%$		



No. CNAS L0893

第 111 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	and Volume Resistivity Testers	Dielectric Dissipation Factor	Volume Resistivity Testers JJF1618	0.001%~10%	$U_{rel}=0.5\%$		
		Volume Resistivity		1M Ω ~1T Ω	$U_{rel}=0.5\%\sim 3\%$		
43	*A.C.Resistance Boxes	Resistance	Calibration Specification for A.C.Resistance Boxes JJF1636	1 Ω ~100 Ω	$U_{rel}=0.03\%$		
		Resistance		100 Ω ~100k Ω	$U_{rel}=0.02\%$		
		Resistance		100k Ω ~10M Ω	$U_{rel}=0.1\%$		
44	Oscilloscope voltage probes	Attenuation ratio	Calibration Specification for Oscilloscope Voltage Probes JJF1437	1~1000	$U_{rel}=0.6\%$		
		Bandwidth		250kHz~2GHz	$U_{rel}=5\%$		
		Rise time		58ps~50ns	$U_{rel}=5\%$		
VII Time and frequency measuring instrument							
1	Crystal Oscillators Inside the Electrical Measuring Instrument	Frequency	Verification Regulation of Crystal Oscillator inside the Electrical Measurement Instrument JJG180	1 MHz, 5MHz, 10MHz	$U_{rel}=2\times 10^{-9}$		
2	Quartz crystal frequency standard	Frequency	Verification Regulation of Quartz Crystal Frequency Standards JJG181	1 MHz, 5MHz, 10MHz	$U_{rel}=1\times 10^{-10}$		
3	Frequency Meters	Frequency	Verification Regulation of Frequency Meters JJG603	10Hz~20kHz	$U_{rel}=5\times 10^{-5}$		
4	Frequency Comparator	Comparator uncertainty	Verification Regulation of Frequency Comparator JJG545	0.1s~10s	$U_{rel}=1\times 10^{-11}/\tau$		



No. CNAS L0893

第 112 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
5	Vibrating Wire Frequency Readouts	Frequency	Calibration Specification for Vibrating Wire Frequency Readouts JJF1401	(300~6000)Hz	$U=(0.03\sim0.6)\text{Hz}$		
6	High Voltage Capacitance Bridges	Time	High Voltage Switch Operation Characteristic Testers JJG1120	(0.01~9999.99)ms	$U=0.02\text{ms}$	变更	
		Length		(0~300)mm	$U=0.03\text{mm}$		
7	Mechanical stopwatch	Time	Verification Regulation of Stopwatches JJG237	分度盘 T 240s/900s/1800s/3600s	$U=2\times10^{-7}\times T+3\text{ms}$		
				秒盘 T 60s/30s/6s	$U=2\times10^{-7}\times T+3\text{ms}$		
8	Electronic stopwatch	Time	Verification Regulation of Stopwatches JJG237	(0~24)h、10s、10min、1h、1d	$U=2\times10^{-7}\times T+3\text{ms}$		
9	Timer and Charge System of SPC Exchange	Time Length	Verification Regulation of the Timer and Charge System of SPC Exchange JJG(川)87	(0.1~864000)s	$U_{\text{rel}}=0.1\%$		
10	Time Interval Generator	Frequency	Calibration Specification for Time Interval Generators JJF 1902	5MHz, 10MHz	$U_{\text{rel}}=2\times10^{-9}$		
		Time		10ns~10000s	$U=2\times10^{-7}\times T+3\mu\text{s}$		
11	Time Interval Meter	Crystal oscillator Frequency accuracy	Verification Regulation of Time Interval Meters JJG238	5MHz 10MHz	$U_{\text{rel}}=5\times10^{-9}$		
		Time Interval		1 μs ~10000s	$U=2\times10^{-7}\times T+3\mu\text{s}$		
12	Coast-down Time Testers	Velocity	Calibration Specification for Coast-down Time Testers JJF1360	0.1km/h~20km/h	$U=0.01\text{km/h}$		
				20km/h~130km/h	$U_{\text{rel}}=4.6\times10^{-4}$		
		Time		0.1s~150s	$U=1.1\text{ms}$		
		Diameter		0mm~300mm	$U=0.03\text{mm}$		



No. CNAS L0893

第 113 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
13	GNSS Vehicle Comprehensive Performance Tester	Response time of localization	Calibration Specification for Vehicle Comprehensive Performance Tester JJF(Mechanics)1017	(0~1000)s	$U=0.02s$		
		Time		(0~1000)s	$U=0.02s$		
		Velocity		(5~300)km/h	$U=0.08km/h$		
		Distance		(0.1~10000) m	$U_{rel}=0.02\%$		
		Positioning accuracy		经度(0~180)°, 纬度(0~90)°	$U=0.0002'$		
14	GNSS Signal Simulators	Power range	Calibration Specification for GNSS Signal Simulators JJF1471	(-50~-90) dB	$U=0.2dB$		
		Power resolution		0.1dB~2dB	$U=0.06dB$		
		Harmonic		(-100~0)dBc	$U=2.0dB$		
		Frequency		1GHz~2.7GHz	$U_{rel}=3.6 \times 10^{-10}$		
		1s Frequency stability		$5 \times 10^{-11}/s$	$U_{rel}=1.6 \times 10^{-12}$		
		Speed range		(0~2.0×10 ⁵)m/s	$U=0.04m/s$		
		Acceleration range		(0~2×10 ⁴)m/s ²	$U=0.02m/s^2$		
		Acceleration range		(0~2×10 ³)m/s ³	$U=0.01m/s^3$		
		Pseudorange resolution		(0.01~0.1)m	$U=0.06m$		
		Pseudorange ratio resolution		(0.01~0.1)m/s	$U=0.006m/s$		



No. CNAS L0893

第 114 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
15	Calibration device of high voltage switch characteristic tester	Time	Calibration device of high voltage switch characteristic tester NIMTT(CM) 070	(1~20000)ms	$U=2 \times 10^{-4}\text{ms}$		
16	Time calibrator	Crystal oscillator Frequency accuracy	Verification Regulation of Time Interval Generator JJG601	5MHz 10MHz	$U_{\text{rel}}=5 \times 10^{-9}$		
		Time Interval		1ms~1s	$U=2 \times 10^{-7} \times T+3 \mu \text{ s}$		
		Time Interval		1s~1d	$U=2 \times 10^{-7} \times T+3\text{ms}$		
17	Microwave Frequency Counter	Crystal oscillator Frequency accuracy	Verification Regulation of Microwave Frequency Counters JJG841	5MHz 10MHz	$U_{\text{rel}}=5 \times 10^{-10}$		
		Frequency		100kHz~67GHz	$U_{\text{rel}}=5 \times 10^{-10}$		
18	Universal Counters	Crystal oscillator Frequency accuracy	Verification Regulation of Universal Counters JJG349	5MHz 10MHz	$U_{\text{rel}}=5 \times 10^{-10}$		
		Frequency		0.1Hz~18GHz	$U_{\text{rel}}=5 \times 10^{-10}$		
VIII Optics measuring instrument							
1	Illuminance Meter	Illuminance	Illuminance Meter JJG245	(10 ⁻³ ~10 ⁻¹)lx	$U_{\text{rel}}=(2.5\% \sim 2.0\%)$		
				(10 ⁻¹ ~10)lx	$U_{\text{rel}}=(2.0\% \sim 1.0\%)$		
				(10~3000)lx	$U_{\text{rel}}=1.0\%$		
				(3000~10000)lx	$U_{\text{rel}}=(1.0\% \sim 2.5\%)$		



No. CNAS L0893

第 115 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
2	Standard Lamp of Luminous Intensity	Luminance	Standard Lamp of Luminous Intensity JJG 246	(1~10)cd	$U_{rel}=(1.2\%\sim 0.8\%)$		
				(10~1200)cd	$U_{rel}=0.8\%$		
				(1200~5000)cd	$U_{rel}=(0.8\%\sim 2.0\%)$		
3	Standard Incandescent Lamp	Total Luminous Flux	Standard Incandescent Lamp for Total Luminous Flux JJG 247	(50~2×10 ⁴)lm	$U_{rel}=1.0\%$		
4	Luminance Meter(Chromatic Luminance Meter)	Luminance	Luminance Meter JJG 211	(0.05~0.3) cd/m ² /	$U_{rel}=2.4\%$		
				(0.3~3000) cd/m ²	$U_{rel}=2.0\%$		
				(3000~50000) cd/m ²	$U_{rel}=3.0\%$		
				x,y:0.0~0.9	$U(x)=0.0004,U(y)=0.0004$		
5	Spectral Radiometers	Wavelength	Spectral Radiometers NIMTT(CM) 060,Colour Temperature Meters JJG 212,Calibration Specification for Spectralradiometers JJF1975	(250~2500)nm	$U=0.1\text{nm}$		
		Chromaticity Coordinate		x,y:0.0~0.9	$U(x)=0.0008,U(y)=0.0006$		
		Color Temperature		(2042~9500)K	(2042~3200)K: $U=(6\sim 15)\text{K},(3200\sim 9500)\text{K}:U=(15\sim 80)\text{K}$		
		spectral radioation		(250~2500)nm	(250~400)nm: $U_{rel}=(4.1\sim 2.8)\%$, (400~800)nm: $U_{rel}=2.8\%$, (800~2500)nm: $U_{rel}=(2.8\sim 5.3)\%$		
				(200-250) nm	$U_{rel}=12\%\sim 4.1\%$		
				(250~400) nm	$U_{rel}=(4.1\sim 2.8)\%$		



No. CNAS L0893

第 116 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(250~2500)nm	$U_{rel}=(250\sim 400)\text{nm}: U_{rel}=(4.1\sim 2.8)\%, (400\sim 800)\text{nm}: U_{rel}=2.8\%, (800\sim 2500)\text{nm}: U_{rel}=(2.8\sim 5.3)\%$		
		Stray Light		(0~5)%	$U=0.2\%$		
		Rendering Index		0~100	$U=0.3$		
6	Color Analyzers	Color Temperature	Calibration Specification of Cathode Ray Tubes(CRT) Analyzers JJF 1079	6500K~9300K	$U=1.0\times 10^2\text{K}$		
		Chromaticity Coordinate		x,y:0.0~0.9	$U=0.002$		
		Luminance		(50~500)cd/m ²	$U_{rel}=3.5\%$		
7	Lovibond Comparable Colormeter	Lovibond Chromaticness	Lovibond Comparable Colormeter JJG 758	R:(0.1~79.9)	$U=0.6$		
				Y:(0.1~79.9)	$U=0.6$		
				B:(0.1~49.9)	$U=0.6$		
				N:(0.1~3.9)	$U=0.6$		
8	Colormeters and Color Difference Meters	Stimulus value Y	Colormeters and Color Difference Meters JJG 595	Y:0~100	$U=0.9\sim 1.3$		
		Chromaticity coordinate		x,y:0.0~0.9	$U=0.0074\sim 0.0086$		
9	Color Standard Plates	Spectral reflectance	Color Standard Plates JJG 453, Whiteness Meter JJG 512, Filter-Type Smokemeters JJG 847, V. R. of	(0~100) %	$U=0.96\%$		
		Spectral reflection	Colorimeters and Color Difference Meters JJG 595	(0~100) %	$U=1.0\%\sim 1.3\%$		
		Stimulus value		X:0~100,Y:0~100,Z:0~150	$U=0.13\sim 1.3$		



No. CNAS L0893

第 117 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Stimulus value		X:0~100,Y:0~100,Z:0~150	$U=0.01\sim0.77$		
		Chromaticity coordinate		x,y:0.0~0.9	$U=0.0017\sim0.0046$		
		Chromaticity coordinate		x,y:0.0~0.9	$U=0.0013\sim0.0025$		
		Lightness index		0~100	$U=0.36\sim0.53$		
		Chroma index		a*: (-100~100) ,b*: (-100~100)	$U=0.53\sim2.0$		
		Chromatic aberration		0~100	$U=0.4$		
		Blue Whiteness		0~100	$U=1.0\sim1.2$		
		G-s Whiteness		0~100	$U=1.2\sim1.3$		
		Light tone index		(-3~3)	$U=0.5$		
		Hunter Whiteness		0~100	$U=1.1\sim1.4$		
		Smokemetric tablet		0.1~10.0	$U=0.3$		
10	Whiteness Meter and standard white plate	Whiteness	Whiteness Meter JJG 512	W:0~100	$U=1.0\sim1.3$	合格评定国家认可 CNAS 认可证书专用章	
11	Laser for Medicine	Laser Power	Verification Regulation of Laser for Medicine JJG 581	0.1mW~100mW	$U_{rel}=2.0\%$		
		Laser Power		0.1W~150W	$U_{rel}=5.0\%$		
12	Laser Energy Meter	Laser Energy	Verification Regulation of laser energy meter JJG312	0.4mJ~850mJ	$U_{rel}=4\%$		



No. CNAS L0893

第 118 页 共 181

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
13	laser beams Analysis	Laser beam width	Calibration Specification for laser beam Analysis NIMTT(CM) 118	0.2mm~5mm	$U_{rel}=5\%$		
14	Laser Power Meter	Laser Power	Verification Regulation of 0.1mW~200W laser power meter JJG249	0.1mW~100mW	$U_{rel}=2.0\%$		
		Laser Power		0.1W~150W	$U_{rel}=4.0\%$		
15	Irradiance Meter	Ultraviolet Irradiance (UVA)	Ultraviolet Irradiance Meters JJG 879, Calibration Specification for Wide-band Irradiance Meters JJF 1660	UVA:(1~30000) $\mu\text{W}/\text{cm}^2$	$U_{rel}=16\%$		
		Ultraviolet Irradiance(UVB&UVC)		UVB:(1~1000) $\mu\text{W}/\text{cm}^2$, UVC:(1~1000) $\mu\text{W}/\text{cm}^2$	$U_{rel}=13\%$		
		Radiation Exposure Amount		(100~5000) mJ/cm^2	$U_{rel}=16\%$		
		Wide-band Irradiance		250nm~2500nm	(250~400)nm: $U_{rel}=(6.1\sim 5.1)\%$, (400~800)nm: $U_{rel}=5.1\%$, (800~2500)nm: $U_{rel}=(5.1\sim 6.6)\%$		
16	Standard Light Source	Spectral Irradiance	Verification Regulation of Spectral Irradiance Standard Lamp JJG 384, Verification Regulation of Spectral Radiance Standard Lamp JJG 383, Calibration Specification of radiometers used in aging test JJF 1525, V. R. of Standard Lamps for Distribution (Colour) Temperature JJG 213,	(250~2500)nm	(250~400)nm: $U_{rel}=(4.1\sim 2.8)\%$, (400~800)nm: $U_{rel}=2.8\%$, (800~2500)nm: $U_{rel}=(2.8\sim 5.3)\%$		
		Color Rendering Index		(250~2500)nm	(250~400)nm: $U_{rel}=(4.6\sim 3.5)\%$, (400~800)nm: $U_{rel}=3.5\%$, (800~2500)nm: $U_{rel}=(3.5\sim 5.7)\%$		
		Color Temperature		(2042~2353)K	$U=(5.0\sim 6.5)\text{K}$		



No. CNAS L0893

第 119 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
			NIMTT(CM) 060	(2353~2856)K	$U=(6.5\sim9.0)K$		
				(2856~3200)K	$U=(9.0\sim14)K$		
				(3200~6500)K	$U=(14\sim30)K$		
				(6500~9000)K	$U=(30\sim50)K$		
		Chromaticity Coordinate		x,y:0.0~0.9	$U(x)=0.0008, U(y)=0.0006$		
17	Radiation Fluxmeters	Irradiance	Verification Regulation of Radiation Fluxmeters JJF1572	(0.1 ~2.0)kW/m ²	$U_{rel}=5\%$		
18	Stabilized Laser Sources	output power	Verification Regulation of Stabilized Laser Sources for Optical Transmit JJG 958	(-70~30)dBm	$U=0.09dB$		
		Central wavelength		(600~1700)nm	$U=0.05nm$		
		Spectral bandwidth		(600~1700)nm	$U=0.05nm$		
19	Optical Attenuator	Attenuation value	Calibration Specification of Optical Attenuator for Telecommunications JJF1199	(800~1700)nm: (0~60)dB	$U=0.08dB$		
		Insertion loss		(800~1700)nm: (0~60)dB	$U=0.08dB$		
20	Optical Return Loss Meters	Optical return loss	Calibration Specification of Optical Return Loss Meters for Telecommunications JJF1325	(800~1700) nm, (0~60) dB	$U=(0.4\sim0.6) dB$		
21	Filter	Spectral transmittance	Interference filter JJG 812, Light filter NIMTT(CM) 066	(0.01~100)%	$U_{rel}=0.3\%$		
		Absorbance		(0~3) A	$U=0.003$		
		Haze		0.1~30.0	$U=0.2$		



在线扫码获取验证

No. CNAS L0893

第 120 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		centerwavelength	ilac-M	(200~850) nm	$U=0.2$ nm		
				(850~2600) nm	$U=(0.2\sim0.5)$ nm		
		Full of halfmaximum		(200~850) nm	$U=0.2$ nm		
				(850~2600) nm	$U=(0.2\sim0.5)$ nm		
		Absorptance		(0~100) %	$U=0.3\%$		
		Effective Absorptance		(0~100) %	$U=0.3\%$		
22	Fiber Optical power meter	Optical power	Verification Regulation of Fiber Optical Power Meters JJG813	(-70~10)dBm	$U_{rel}=2.3\%$		
23	Passive Optical Network Power Meters	Burst optical power	Calibration Specification for Passive Optical Network Power Meters JJF 1755	(-50~10)dBm	$U=0.2$ dB		
		Continuous optical power		(-70~10)dBm	$U_{rel}=2.3\%$		
24	*Solar simulator	Spectral match	Calibration Specification of Solar Simulator JJF 1615	(300~1100)nm	$U_{rel}=9.0\%$		
		Non-uniformity of irradiance		(200~2000)W/m ²	$U_{rel}=2.6\%$		
		Temporal instability of irradiance		(200~2000)W/m ²	$U_{rel}=2.6\%$		
25	Verification device of Pupil Distance Meters	Length	Calibration specification for verification device of pupillary distance meter JJF (川) 177	(55~85)mm	$U=(3+L/100)$ μm (L:mm)		
26	Reflectometers	Reflectivity	Calibration Specification for Reflectometers JJF 1232	0~100	$U=0.9\sim1.3$		



No. CNAS L0893

第 121 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
27	Standard Lenses of Vertex Power	Vertex Power	Verification Regulation of Standard Lenses of Vertex Power JJG 866	$(-25.0 \sim +25.0) \text{m}^{-1}$	$U = (0.02 \sim 0.03) \text{m}^{-1}$		
28	Trial case lenses	Vertex Power	Verification Regulation of Trial case lenses JJG 579	$(-25.0 \sim +25.0) \text{m}^{-1}$	$U = (0.02 \sim 0.03) \text{m}^{-1}$		
29	Focimeters	Vertex Power	Verification Regulation of Focimeters JJG 580	$(-25.0 \sim +25.0) \text{m}^{-1}$	$U = (0.02 \sim 0.03) \text{m}^{-1}$		
30	Eye Refractometers	Vertex Power	Verification Regulation of Eye Refractometers JJG 892	Objective: $(-20 \sim +20) \text{m}^{-1}$	$U = (0.07 \sim 0.10) \text{m}^{-1}$		
		Vertex Power		Subjective: $(-15 \sim +15) \text{m}^{-1}$	$U = 0.04 \text{m}^{-1}$		
31	Black and White Step Tablet	Black and White Step Tablet	Black and White Step Tablet JJG 452	D: $(0.0 \sim 4.0)$	$U = 0.02$		
				D: $(4.0 \sim 5.10)$	$U = 0.03$		
32	Visual Densitometer	Density	Verification Regulation of Diffuse Transmission Visual Densitometer JJG 920	D: $(0.00 \sim 4.00)$	$U = 0.02$		
				D: $(4.00 \sim 5.00)$	$U = 0.03$		
33	Abbe Refractometer	Refractive index	Abbe Refractometer JJG 625	$n_D: 1.47001 \sim 1.67248$	$U = 1 \times 10^{-4}$		
		Dispersion		$n_F - n_C: 0.00708 \sim 0.02086$	$U = 7 \times 10^{-5}$		
34	Specular Gloss Meters and Gloss Plates	Gloss	V.R. of Specular Gloss Meters and Gloss Plates JJG 696	Specular Gloss Meters: $(0.0 \sim 120.0) \text{GU}$	1.0 GU		
				Gloss Plates $(0.0 \sim 120.0) \text{gloss unit}$	1.2 GU		
35	Transmittance Measuring Equipment For Ophthalmic Products	Spectral transmittance	Calibration Specification of Transmittance Measuring Equipment For Ophthalmic Products JJG 1106	$0.1\% \sim 100\%$	$U = 1.5\%$		
		Spectral	Calibration Specification for Relative Spectral	$(300 \sim 400) \text{nm}$	$U_{\text{rel}} = 3.0\%$		

No. CNAS L0893

第 122 页 共 181



The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			Responsivity for Photoelectric Detectors JJF 1150	(400~1100)nm	$U_{rel}=2.0\%$		
37	Retroreflection Meters	Coefficient of Retroreflection	Calibration Specification Of Retroreflection Coefficient Meters for Motor Vehicle's Reflecting Marking JJF 1747, C. S. for Retroreflectometer JJF 1809	(0.1~1999) $cd \cdot lx^{-1} \cdot m^{-2}$	$U_{rel}=7.5\%$		
		Coefficient of Retroreflected Luminance		(0.1~1999) $mcd \cdot lx^{-1} \cdot m^{-2}$	$U_{rel}=12\%$		
		Coefficient of luminous intensity		(1~1999) $mcd \cdot lx^{-1}$	$U_{rel}=12\%$		
38	Water Colorimeter	Water Chromaticness	Water Colorimeter JJF 1689	(0~100) PCU	0.8 PCU		
39	Reflection densitometer	Optical Density	Reflection densitometer JJF 1492	0.07~1.00	$U=0.03$		
				1.00~2.04	$U=0.04$		
40	Transmittance meter	Transmittance	Transmittance meter NIMTT(CM) 146	(0~100) %	$U=0.8\%$		
		Stimulus value Y		Y:0~100	$U=1.0$		
		Chromaticity coordinate		x,y: 0.0~0.9	$U=0.0064$		
41	*UV pretreatment test chamber of photovoltaic module	Distribution of UV spectrum radiation	Calibration Specification for UV Preconditioning Test of Photovoltaic Module NIMTT(CM) 143	(250~400) nm	$U_{rel}=12\%$		
		Nonuniformity of UV irradiance		(0.1~40) mW/cm^2	$U_{rel}=8\%$		



No. CNAS L0893

第 123 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		UV irradiance		(0.1~40)mW/cm ²	<i>U</i> _{rel} =15%		
42	UV Analyzers	Peak Wavelength	Calibration Specification for UV Analyzers JJF 1936	(250-400) nm	<i>U</i> =0.5 nm		
		UV Irradiance		(1~1000) μ W/cm2	<i>U</i> _{rel} =15%		
43	Standard Lamps of Distribution Temperature	Distribution Temperature	Standard lamps for Distribution(Colour) Temperature JJG 213	(2042~2353)K	<i>U</i> =(5.0~6.5)K		
				(2353~2856)K	<i>U</i> =(6.5~9.0)K		
				(2856~3200)K	<i>U</i> =(9.0~14)K		
44	*I-V Curve measuring instrument	Open circuit voltage	Calibration Specification of I-V Curve Measuring Instrument NIMTT(CM) 120	10mV~80V	<i>U</i> _{rel} =1.0%		
		Short-circuit current		(0.1~20) A	<i>U</i> _{rel} =1.4%		
45	*Hazemeter	Haze	Calibration Specification for Hazemeter JJF1303	1~30	<i>U</i> =0.31		
		Transmission ratio		0.7~0.9	<i>U</i> =0.008		
46	*Transmittance Reference Filter	Wavelength	Verification Regulation of Transmittance Reference Filter JJG1034	(200~2600)nm	<i>U</i> =(0.10~0.35)nm		
		Transmittance		0.02~1.00	<i>U</i> _{rel} =(0.26~0.40)%		
IX Chemistry measuring instrument							
1	*Wavelength Dispersive X-Ray Fluorescence Spectrometers	Count rate	Verification Regulation for Wavelength Dispersive X-Ray Fluorescence Spectrometers JJG810	(10~2000)kcps	<i>U</i> _{rel} =0.3%		
2	*Hand Saccharimeter (Content-meter) and Hand Refractometer	Sugar content	Verification Regulation of Hand Saccharimeter (Content-meter) and Hand Refractometer JJG820	(0~60) %	<i>U</i> =0.06%		
		Refractive index		<i>n</i> _D : (1.3~1.5)	<i>U</i> _{rel} =1%		

No. CNAS L0893

第 124 页 共 181



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
3	*Colorimeter	content	Calibration Specification for Colorimeter NIMTT(CM) 073	(0.2~10.0) mg/L	<i>U</i> _{rel} =1%		
4	*Polarimeter and Polarimetric Saccharimeters	specific rotation	Verification Regulation of Polarimeter and Polarimetric Saccharimeters JJG536	(-45~+45)°	<i>U</i> =0.002°		
5	*Mercury Analyzers	Detection limit	Verification Regulation of Mercury Analyzers JJG548	absorption: ≤1.0 ng	<i>U</i> _{rel} =48%		
				Fluorescence: ≤0.1 ng	<i>U</i> _{rel} =48%		
6	*Flame Photometer	Detection limit	Verification Regulation of Flame Photometer JJG630	K: ≤0.004 mmol/L	<i>U</i> _{rel} =48%		
				Na: ≤0.008 mmol/L	<i>U</i> _{rel} =48%		
7	*UV-VIS-NIR Spectrophotometers	Wavelength	Verification Regulation of UV-VIS-NIR Spectrophotometers JJG178	(200~700)nm	<i>U</i> =0.01 nm		
				(700~900)nm	<i>U</i> =0.2 nm		
				(900~2600)nm	<i>U</i> =0.1 nm		
		Transmittance		(7~55)%	<i>U</i> _{rel} =0.37%		
8	*Fourier Transform Infrared Spectrometer	wave number	Calibration Specification for Fourier Transform Infrared Spectrometer JJF1319	(900~3300)cm ⁻¹	<i>U</i> =(0.03~0.10) cm ⁻¹		
9	*Dispersive Infrared Spectrophotometers	wave number	Verification Regulation of Dispersive Infrared Spectrophotometers JJG681	(900~3300)cm ⁻¹	<i>U</i> =(0.03~0.10) cm ⁻¹		
10	*Atomic Absorption Spectrophotometers	Wavelength	Verification Regulation of Atomic Absorption Spectrophotometers JJG694	(190~1000)nm	<i>U</i> =0.1 nm		
		detection limit		Cu:≤0.02 μg/mL	<i>U</i> _{rel} =48%		
				Cd:≤4 pg	<i>U</i> _{rel} =48%		



No. CNAS L0893

第 125 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
11	*Fluorescence Spectrophotometer	Wavelength	Verification Regulation of Fluorescence Spectrophotometer JJG537	(240~550)nm	<i>U</i> =0.1 nm		
		Detection limit		type A: ≤5×10 ⁻¹⁰ g/mL	<i>U</i> _{rel} =48%		
				type B: ≤1×10 ⁻⁸ g/mL	<i>U</i> _{rel} =48%		
12	*Atomic Fluorescence Spectrophotometers	Detection limit	Verification Regulation of Atomic Fluorescence Spectrophotometers JJG939	As、Sb: ≤0.4 ng	<i>U</i> _{rel} =48%		
13	*Emission Spectrometer	Wavelength	Emission Spectrometer JJG768	(190~1000)nm	<i>U</i> =0.01 nm		
		Detection limit		Zn: ≤0.01 mg/L(ICP)	<i>U</i> _{rel} =48%		
				Ni: ≤0.03 mg/L(ICP)	<i>U</i> _{rel} =48%		
				Mn: ≤0.005 mg/L(ICP)	<i>U</i> _{rel} =48%		
				Cr: ≤0.02 mg/L(ICP)	<i>U</i> _{rel} =48%		
				Cu: ≤0.02 mg/L(ICP)	<i>U</i> _{rel} =48%		
				Ba: ≤0.005 mg/L(ICP)	<i>U</i> _{rel} =48%		
				C: ≤0.02%(Direct reading spectrometer)	<i>U</i> _{rel} =48%		
				Si: ≤0.02%(Direct reading spectrometer)	<i>U</i> _{rel} =48%		
				Mn: ≤0.02%(Direct reading spectrometer)	<i>U</i> _{rel} =48%		
				Cr: ≤0.01%(Direct reading spectrometer)	<i>U</i> _{rel} =48%		
				Ni: ≤0.02%(Direct reading spectrometer)	<i>U</i> _{rel} =48%		




在线扫码获取验证

No. CNAS L0893

第 126 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				V: ≤0.01%(Direct reading spectrometer)	<i>U</i> _{rel} =48%		
				Spectrograph: ≤0.003%	<i>U</i> _{rel} =48%		
14	*Flow Analyzers with Spectrophotography	Detection limit	Calibration Specification for Flow Analyzers with Spectrophotography JJF 1568	cyanide: ≤0.002 mg/L	<i>U</i> _{rel} =48%		
				Volatile phenol: ≤0.002 mg/L	<i>U</i> _{rel} =48%		
				Cr ⁶⁺ : ≤0.004 mg/L	<i>U</i> _{rel} =48%		
				sulfide:≤0.005 mg/L	<i>U</i> _{rel} =48%		
				TP: ≤0.01 mg/L	<i>U</i> _{rel} =48%		
				TN: ≤0.04 mg/L	<i>U</i> _{rel} =48%		
				Amino Nitrogen: ≤0.04 mg/L	<i>U</i> _{rel} =48%		
				Anion Active Detergent: ≤0.05 mg/L	<i>U</i> _{rel} =48%		
		wavelength		(210~1100) nm	<i>U</i> =0.2 nm		
15	*Capillary Electrophoresis Instruments	Detection limit	Verification Regulation of Capillary Electrophoresis Instruments JJG964	≤1×10 ⁻⁶ g/mL (VB ₆)	<i>U</i> _{rel} =10%		
16	*Gas Chromatographs	Temperature	Verification Regulation of Gas Chromatographs JJG700	(20~300)℃	<i>U</i> =0.06℃		
		Sensitivity		TCD: ≥800 mV·mL / mg	<i>U</i> _{rel} =3.5%		
		Detection limit		FID:≤5 ng/s	<i>U</i> _{rel} =14%		
				FPD:≤0.5 ng/s (S)	<i>U</i> _{rel} =14%		
				FPD:≤0.1 ng/s (P)	<i>U</i> _{rel} =14%		

No. CNAS L0893

第 127 页 共 181



在线扫码获取验证

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
				NPD:≤5 pg/s （N）	<i>U</i> _{rel} =14%		
				NPD:≤10 pg/s （P）	<i>U</i> _{rel} =14%		
				ECD:≤5 pg/mL	<i>U</i> _{rel} =14%		
17	*on-line Gas Chromatograph	Temperature	Verification Regulation of On-line Gas Chromatograph JJG1055	(20～300)℃	<i>U</i> =0.06℃		
		Sensitivity		TCD: ≥1000 mV·mL / mg	<i>U</i> _{rel} =3.5%		
		Detection limit		PID: ≤5×10 ⁻¹² g/mL	<i>U</i> _{rel} =14%		
18	*Liquid Chromatographs	Flow	Verification Regulation of Liquid Chromatographs JJG705	(0.5～10) mL/min	<i>U</i> _{rel} =0.3%		
		Temperature		(10～90)℃	<i>U</i> =0.06℃		
		Wavelength		(200～400)nm	<i>U</i> =0.2nm		
		Detection limit		UV-VIS/DAD: ≤5×10 ⁻⁸ g/mL	<i>U</i> _{rel} =10%		
				FLD: ≤5×10 ⁻⁹ g/mL	<i>U</i> _{rel} =10%		
				RID: ≤5×10 ⁻⁶ g/mL	<i>U</i> _{rel} =10%		
				ELSD: ≤5×10 ⁻⁶ g/mL	<i>U</i> _{rel} =10%		
19	*Gel Chromatograph	Flow	Verification Regulation of Gel Chromatograph JJG342	(0.5～10) mL/min	<i>U</i> _{rel} =0.3%		
		Temperature		(10～90)℃	<i>U</i> =0.06℃		
		molecular weight		Organic phase: （1×10 ³ ～5×10 ⁵ ） g/mol	<i>U</i> _{rel} =8%		
				Water phase: （1×10 ³ ～5×10 ⁵ ） g/mol	<i>U</i> _{rel} =8%		



No. CNAS L0893

第 128 页 共 181

The scope of the accreditation in Chinese remains the definitive version.


No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
20	*Ion Chromatograph	Flow	Verification Regulation of Ion Chromatograph JJG823	(0.5~10) mL/min	$U_{rel}=0.3\%$		
		Temperature		(10~90)℃	$U=0.06^{\circ}\text{C}$		
		Detection limit		Conductivity detector (Cl ⁻ 、Li ⁺) : ≤0.02 μg/mL	$U_{rel}=10\%$		
				UV-VIS detector (NO ₂) : ≤0.02 μg/mL	$U_{rel}=10\%$		
				Electrochemical detector (I ⁻) : ≤0.02 μg/mL	$U_{rel}=10\%$		
21	*Portable Gas Chromatography-Mass Spectrometers	SNR	Calibration Specification for Portable Gas Chromatography-Mass Spectrometers NIMTT(CM) 015	≥10:1	$U_{rel}=14\%$		
22	*Quadrupole Inductively Coupled Plasma Mass Spectrometers	detection limit	Calibration Specification for Quadrupole Inductively Coupled Plasma Mass Spectrometers JJF1159	Be: ≤30 ng/L	$U_{rel}=48\%$		
				In: ≤10 ng/L	$U_{rel}=48\%$		
				Bi: ≤10 ng/L	$U_{rel}=48\%$		
23	*Bench Top Gas Chromatography-Mass Spectrometers	SNR	Calibration Specification for Gas Chromatography-Mass Spectrometries JJF1164	≥10:1	$U_{rel}=14\%$		
24	*Liquid Chromatography-Mass Spectrometers	SNR	Calibration Specification for Liquid Chromatography-Mass Spectrometers JJF1317	≥10:1	$U_{rel}=14\%$		
25	*Time of Flight Mass Spectrometers	M/R	Calibration Specification for Time of Flight Mass Spectrometers JJF1528	(100~5000)	$U_{rel}=1\times10^{-5}$		



No. CNAS L0893

第 129 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
26	*Nitrate-Nitrogen Automatic Analyzers	concentration	Verification Regulation of Nitrate-Nitrogen Automatic Analyzers JJG656	(0~500) μ g/mL	$U=(0.30\sim6.0)\text{mg/L}$		
27	*Analyzers for Oil Content in Water	concentration	Verification Regulation of Analyzers for Oil Content in Water JJG950	(0~1000) mg/L	$U=(0.2\sim20)\text{mg/L}$		
28	*Silicate Analyzers	Concentration	Calibration Specification for Silicate Analyzers JJF1539	(0.02~100) μ g/mL	$U_{\text{rel}}=(4\sim2)\%$		
29	*Ammonia-Nitrogen Automatic Analyzers	Concentration	Verification Regulation of Ammonia-Nitrogen Automatic Analyzers JJG631	(0.01~500) μ g/mL	$U_{\text{rel}}=1.5\%$		
30	*Total Organic Carbon Analyzer	Concentration	Verification Regulation of Total Organic Carbon Analyzer JJG821	TOC: (0.01~1000) μ g/mL	$U_{\text{rel}}=2.1\%$		
		Concentration		TIC: (0.01~1000) μ g/mL	$U_{\text{rel}}=2.1\%$		
31	*Chemical Oxygen Demand(COD) Meters	Concentration	Verification Regulation of Chemical Oxygen Demand(COD) Meters JJG975	A type instrument: (0.01~1500) (mg/L)	$U_{\text{rel}}=1\%$		
		Concentration		B type instrument: (0.01~1500) (mg/L)	$U=0.52\text{mg/L}$		
		Temperature		(100~200) $^{\circ}\text{C}$	$U=0.06^{\circ}\text{C}$		
32	*On-line Automatic Determination of Chemical Oxygen Demand (COD)	Concentration	Verification Regulation of On-line Automatic Determination of Chemical Oxygen Demand (COD) JJG1012	(16~1000) mg/L	$U_{\text{rel}}=1\%$		
33	*Water Quality On-line Analyzers of Total Phosphorus and Total Nitrogen	Concentration	Verification Regulation of Water Quality On-line Analyzers of Total Phosphorus and Total Nitrogen JJG1094	TP: (0.01~500) mg/L	$U_{\text{rel}}=2.3\%$		
		Concentration		TN: (0.01~100) mg/L	$U_{\text{rel}}=2.2\%$		

No. CNAS L0893

第 130 页 共 181



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
34	*Turbidimeter	Turbidity	Verification Regulation of Turbidimeter JJG880	(0.1~400)NTU	$U_{rel}=3.1\%$		
35	*Dissolved Oxygen Meters	Dissolved Oxygen	Verification Regulation of Dissolved Oxygen Meters JJG291	(0~20)mg/L	$U=0.02$ mg/L		
		Temperature		(0~50) °C	$U=0.18$ °C		
36	*Water Quality On-line Analyzers of Heavy Metals	Concentration	Calibration Specification for Water Quality On-line Analyzers of Heavy Metals JJF1565	Pb、Cd、Hg、As、Cr ⁶⁺ 、Cr、Cu、Zn、Ni、Fe、Mn: (0.001~100) mg/L	$U_{rel}=(1\sim3)\%$		
37	*Instrument for KF Coulometry Titration	weight	Verification Regulation of Instrument for KF Coulometry Titration JJG1044	10 μg	$U=1.4$ μg		
				100 μg	$U=14$ μg		
				1000 μg	$U=71$ μg		
				5000 μg	$U=71$ μg		
38	*Thermogravimetric Moisture Meters	weight	Verification Regulation of Thermogravimetric Moisture Meters JJG658	(0~210)g	$U=0.5$ mg		
		Concentration		(94.98~95.02)%	$U=0.20\%$		
39	*Instruments for measuring the moisture content of grain with capacitance and resistance method	Moisture	Verification Regulation of Instruments for Measuring the Moisture Content of Grain with Capacitance and Resistance Method JJG891	(8~33)%	$U=0.1\%$	合格评定国家认可 CNAS 认可证书专用章	
40	*Wood moisture content measuring meters	Concentration	Verification Regulation of Wood Moisture Content Measuring Meters JJG986	(0~50)%	$U=0.04\%$		
41	*Moisture	Concentration	Calibration Specification for Moisture meter NIMTT(CM) 110	(0~100)%	$U=0.8\%$		



No. CNAS L0893

第 131 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
42	Routon Capillary Viscometer	Kinematic viscosity	Verification Regulation of Routon Capillary Viscometer JJG155	$(1 \sim 10^5) \text{mm}^2/\text{s}$	$U_{\text{rel}} = (0.16\% \sim 0.61\%)$		
43	Viscosiuefer for Roll-Down Ball Type	Kinematic viscosity	Verification Regulation of Viscosiuefer for Roll-Down Ball Type JJG214	$(1 \sim 10^5) \text{mm}^2/\text{s}$	$U_{\text{rel}} = (0.16\% \sim 0.61\%)$		
44	*Rotational Viscometers	Viscosity	Verification Regulation of Rotational Viscometers JJG1002	$(1 \sim 10^5) \text{mPa.s}$	$U_{\text{rel}} = (0.16\% \sim 0.61\%)$		
45	Engler Viscosimeter	time	Verification Regulation of Engler Viscosimeter JJG742	$(50 \sim 52) \text{s}$	$U_{\text{rel}} = 0.2\%$		
46	Flow Cup Viscometers	Kinematic viscosity	Verification Regulation of Flow Cup Viscometers JJG743	$(1 \sim 10^5) \text{mm}^2/\text{s}$	$U_{\text{rel}} = (0.16\% \sim 0.61\%)$		
47	*Kinematic Viscosity Tester	Kinematic viscosity	Calibration Specification for Kinematic Viscosity Tester JJF1274	$(0.3 \sim 30000) \text{mm}^2/\text{s}$	$U_{\text{rel}} = (0.24 \sim 0.61)\%$		
		Temperature		$(20 \sim 100) ^\circ\text{C}$	$U = 0.003 ^\circ\text{C}$		
48	*Oscilloscopic Polarograph	Concentration	Verification Regulation of Oscilloscopic Polarograph JJG748	$\text{Cd}: (0.01 \sim 100) \text{mg/L}$	$U_{\text{rel}} = (1.2 \sim 2)\%$		
49	Verifying Meter for pH Meters	potential	Verification Regulation of Verifying Meter for pH Meters JJG919	$(-2000 \sim +2000) \text{mV}$	$U = (0.00058 \sim 0.048) \text{mV}$		
		pH		pH: $(0 \sim 14)$	pH: $U = (0.00006 \sim 0.00018)$		
50	*pH Meters	pH	Verification Regulation of Laboratory pH Meters JJG119	electric unit: $(0 \sim 14)$	$U = 0.001$		
				instrument: $(1 \sim 14)$	$U = 0.006$		
		Voltage		$(-2000 \sim 2000) \text{mV}$	$U = 0.1 \text{mV}$		
51	*Ion Meters	pX	Verification Regulation of Laboratory Ion Meters JJG757	$(0 \sim 14)$	$U = 0.001$		



No. CNAS L0893

第 132 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Potential		(-2000~2000) mV	$U=0.1$ mV		
52	*Automatic Potentiometric Titrators	Potential	Verification Regulation of Automatic Potentiometric Titrators JJG 814	(-2000~2000) mV	$U=0.1$ mV		
		Capacity		(0.1~50) mL	$U=0.001$ mL		
		Concentration		(0.09~0.11) mol/L	$U_{\text{rel}}=0.3\%$		
53	*Electrolytic Conductivity Meters	Conductance	Verification Regulation of Electrolytic Conductivity Meters JJG376	(100 pS · m ⁻¹ ~ 0.25 μS · cm ⁻¹ (electronic unit)	$U_{\text{rel}}=0.7\%$		
				(>0.25~2.5) μS · cm ⁻¹ (electronic unit)	$U_{\text{rel}}=0.2\%$		
				(>2.5 μS · cm ⁻¹ ~ 10 S · cm ⁻¹ (electronic unit)	$U_{\text{rel}}=0.05\%$		
				(0.000118~0.13110)S · cm ⁻¹ (instrument)	$U_{\text{rel}}=0.3\%$		
54	*On-line pH Meters	pH	Calibration Specification for on-line pH Meters JJF 1547	pH:(0~14)	$U=0.001$ (electric unit)		
		Potential		pH:(1~14)	$U=0.01$ (instrument)		
		Temperature		(-2000~2000) mV	$U=0.1$ mV		
				(0~60)°C	$U=0.18$ °C		
55	*alarmer detectors of combustible gas	gas concentration	Verification regulation of alarmer detectors of combustible gas JJG693	Methane, isobutane,propane, hydrogen: (1~100)%LEL	$U_{\text{rel}}=0.7\%$		
				Methane, hydrogen: (3~98)×10 ⁻²	$U_{\text{rel}}=0.7\%$		
56	*carbon monoxide detector	gas concentration	Verification Regulation of carbon monoxide detector	(10.0~200.0)×10 ⁻⁶	$U_{\text{rel}}=2.0\%~0.8\%$		



No. CNAS L0893

第 133 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
			JJG915	$(>200 \sim 1000) \times 10^{-6}$	$U_{\text{rel}}=0.8\%$		
				$(>1000 \sim 3000) \times 10^{-6}$	$U_{\text{rel}}=0.7\%$		
57	*carbon monoxide and carbon dioxide infrared gas analyzer	gas concentration	Verification Regulation of carbon monoxide and carbon dioxide infrared gas analyzer JJG635	CO:(10.0~200.0)×10 ⁻⁶	$U_{\text{rel}}=1.7\% \sim 1.1\%$		
				CO:(>200~1000)×10 ⁻⁶	$U_{\text{rel}}=1.4\% \sim 1.2\%$		
				CO:(>1000~10000)×10 ⁻⁶	$U_{\text{rel}}=1.1\%$		
				CO:(>1.00~5.00)×10 ⁻²	$U_{\text{rel}}=2.1\% \sim 1.4\%$		
				CO:(>5.00~20.00)×10 ⁻²	$U_{\text{rel}}=1.2\%$		
				CO ₂ :(0.050~1.000)×10 ⁻²	$U_{\text{rel}}=1.7\% \sim 1.1\%$		
				CO ₂ :(>1.00~5.00)×10 ⁻²	$U_{\text{rel}}=1.4\% \sim 1.2\%$		
				CO ₂ :(>5.00~20.00)×10 ⁻²	$U_{\text{rel}}=1.1\%$		
58	*Electrochemical Oxygen Meter	gas concentration	Verification Regulation of Electrochemical Oxygen Meter JJG 365	$0.1 \times 10^{-2} \sim 10 \times 10^{-2}$	$U_{\text{rel}}=2.8\% \sim 0.9\%$		
				$>10 \times 10^{-2} \sim 50 \times 10^{-2}$	$U_{\text{rel}}=0.8\%$		
				$>50 \times 10^{-2} \sim 100 \times 10^{-2}$	$U_{\text{rel}}=0.7\%$		
59	*Zirconia Oxygen Analyzers	gas concentration	Verification Regulation of Zirconia Oxygen Analyzers JJG 535	$0.1 \times 10^{-2} \sim 10 \times 10^{-2}$	$U_{\text{rel}}=2.8\% \sim 0.9\%$		
				$>10 \times 10^{-2} \sim 50 \times 10^{-2}$	$U_{\text{rel}}=0.8\%$		
				$>50 \times 10^{-2} \sim 100 \times 10^{-2}$	$U_{\text{rel}}=0.7\%$		
60	*Paramagnetic Oxygen Analyzer	gas concentration	Verification Regulation of Paramagnetic Oxygen Analyzer JJG 662	$0.1 \times 10^{-2} \sim 10 \times 10^{-2}$	$U_{\text{rel}}=2.7\% \sim 0.9\%$		



No. CNAS L0893

第 134 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				$> 10 \times 10^{-2} \sim 50 \times 10^{-2}$	$U_{\text{rel}}=0.8\%$		
				$> 50 \times 10^{-2} \sim 100 \times 10^{-2}$	$U_{\text{rel}}=0.7\%$		
61	*thermal conductivity hydrogen analyzer	gas content	verification regulation of thermal conductivity hydrogen analyzer JJG663	$(0.50 \sim 9.99) \times 10^{-2}$	$U_{\text{rel}}=1.4\% \sim 0.8\%$		
				$(10.0 \sim 100.0) \times 10^{-2}$	$U_{\text{rel}}=1.1\% \sim 0.8\%$		
62	*Sulfur Hydrogen Gas Detectors	gas concentration	Verification Regulation of Sulfur Hydrogen Gas Detectors JJG 695	$5 \times 10^{-6} \sim 100 \times 10^{-6}$	$U_{\text{rel}}=2.5\% \sim 1.5\%$		
				$> 100 \times 10^{-6} \sim 500 \times 10^{-6}$	$U_{\text{rel}}=1.7\% \sim 1.5\%$		
63	*Micro Oxygen Analyzers	gas concentration	Verification Regulation of Micro Oxygen Analyzers JJG 945	1×10^{-6}	$U_{\text{rel}}=3.5\%$		
				$2 \times 10^{-6} \sim 10 \times 10^{-6}$	$U_{\text{rel}}=2.4\% \sim 1.6\%$		
				$> 10 \times 10^{-6} \sim 100 \times 10^{-6}$	$U_{\text{rel}}=1.4\% \sim 0.8\%$		
				$> 100 \times 10^{-6} \sim 1000 \times 10^{-6}$	$U_{\text{rel}}=0.7\%$		
64	*Volatile Organic Compounds Photo Ionization Detectors	gas concentration	Calibration Specification for Volatile Organic Compounds Photo Ionization Detectors JJF 1172	$1 \times 10^{-6} \sim 2000 \times 10^{-6}$	$U_{\text{rel}}=3.0\% \sim 1.4\%$		
65	*Ammonia Gas Detectors	Gas concentration	Verification Regulation of Ammonia Gas Detectors JJG1105	$(20 \sim 300) \times 10^{-6}$	$U_{\text{rel}}=3.4\% \sim 2.1\%$		
66	*Chlorine Alarm Detectors	Gas concentration	Calibration Specification for Chlorine Alarm Detectors JJF1433	$(10 \sim 100) \times 10^{-6}$	$U_{\text{rel}}=4.8\% \sim 2.2\%$		
67	*Alarmer detector of Sulfur Hexafluoride	Gas concentration	Calibration Specification for the Alarmer Detector of Sulfur Hexafluoride JJF1263	$(10 \sim 100) \times 10^{-6}$	$U_{\text{rel}}=6.2\% \sim 1.6\%$		
				$(> 100 \sim 1000) \times 10^{-6}$	$U_{\text{rel}}=1.5\% \sim 1.2\%$		
68	*Chemiluminescence NO/NO _x	Gas concentration	Verification Regulation of Chemiluminescent NO/NO _x	NO: $(10 \sim 1000) \times 10^{-6}$	$U_{\text{rel}}=3.8\% \sim 0.8\%$		

No. CNAS L0893

第 135 页 共 181



The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	Analyzers		Analyzers JJG801	NO: ($>1000 \sim 5000$) $\times 10^{-6}$	$U_{rel}=0.7\%$		
69	*Air Samplers	Flow	Verification Regulation of Air Samplers JJG956	(0.1~6)L/min	$U_{rel}=1.4\%$		
		Time		(0~3600) s	$U=0.1s$		
70	*Dust Sampler	Flow	Verification Regulation of Dust Sampler JJG520	(0.1~60)L/min	$U_{rel}=1.2\%$		
		Time		(0~3600) s	$U=0.1s$		
71	Flue Gas Samplers	Flow	Flue Gas Samplers JJG 1169	(0.1~2)L/min	$U_{rel}=1.4\%$		
		Temperature		(0~300)°C	$U=0.11^{\circ}C$		
		pressure		(-20~0) kPa	$U=0.15kPa$		
		atmospheric pressure		(80~106) kPa	$U=2.9 kPa$		
		time		(10~3600) s	$U=0.1s$		
72	*Flue Gas Analyzers	gas concentration	Verification Regulation of Flue Gas Analyzers JJG 968	CO: $10 \times 10^{-6} \sim 1000 \times 10^{-6}$	$U_{rel}=2.0\% \sim 1.2\%$		
				CO: $>1000 \times 10^{-6} \sim 10000 \times 10^{-6}$	$U_{rel}=1.1\%$		
				O ₂ : $1 \times 10^{-2} \sim 3 \times 10^{-2}$	$U_{rel}=2.0\% \sim 1.1\%$		
				O ₂ : $>3 \times 10^{-2} \sim 30 \times 10^{-2}$	$U_{rel}=0.8\%$		
				SO ₂ : $2 \times 10^{-6} \sim 500 \times 10^{-6}$	$U_{rel}=2.7\% \sim 1.4\%$		
				SO ₂ : $>500 \times 10^{-6} \sim 10000 \times 10^{-6}$	$U_{rel}=1.4\%$		
				NO: $10 \times 10^{-6} \sim 500 \times 10^{-6}$	$U_{rel}=2.0\% \sim 0.7\%$		



No. CNAS L0893

第 136 页 共 181

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				NO: $> 500 \times 10^{-6} \sim 10000 \times 10^{-6}$	$U_{rel}=0.7\%$		
73	Samplers for Stack Dust	Flow	Samplers for Stack Dust JJG 680	(0.1~100)L/min	$U_{rel}=1.4\%$		
		Temperature		(80~300)°C	$U=0.9^{\circ}\text{C}$		
74	*Total Suspended Particulates Sampler	Flow	Verification Regulation of Total Suspended Particulates Sampler JJG943	(80~150)L/min	$U_{rel}=1.4\%$		
		Time		(0.8~1.2)m ³ /min	$U_{rel}=1.4\%$		
75	*Airborne Particle Counter	Particles concentration	Calibration Specification for Airborne Particle Counter JJF1190	(35~350 × 10 ⁶)particale/m ³ <sup>su	$U_{rel}=14\%$		
		Time		(0~3600) s	$U=0.1\text{s}$		
76	Ozone Gas Analyzers	Gas concentration	Verification Regulation of Ozone Gas Analyzers JJG1077	(0.1~400) × 10 ⁻⁶	$U_{rel}=(2.3 \sim 1.7)\%$		
77	*Static Light Scattering Particle Size Analyzers	Length	Calibration Specification for Static Light Scattering Particle size Analyzers JJF1211	Median particle size: (1~150) μm	$U_{rel}=(4.6 \sim 1.6)\%$		
78	Calibration device for formaldehyde gas analyzers	Gas concentration	Calibration Specification of Calibration Device for Formaldehyde Gas Analyzers NIMTT(CM) 091	(0.1~10)mg/m ³	$U_{rel}=2.9\%$		
79	formaldehyde gas analyzer	Gas concentration	Verification Regulation of Formaldehyde Gas Analyzers JJG 1022	(0.08~2.0) × 10 ⁻⁶	$U_{rel}=2.6\%$		
80	*Verifying Meter for Breath Alcohol Analyzers	Gas concentration	Calibration Specification of Verification Device for for Breath Alcohol Analyzers	(0.05~2) mg/L	$U_{rel}=(1.4 \sim 1.1)\%$		
		temperature		(0~40) °C	$U=0.2^{\circ}\text{C}$		



No. CNAS L0893

第 137 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		flow rate	NIMTT(CM) 059	(50~50000) mL/min	$U_{rel}=0.8\%$		
		pressure		(0~2500) Pa	$U=10$ Pa		
81	Breath Alcohol Analyzers	concentration	Verification Regulation of Breath Alcohol Analyzers JJG 657	(0.05~2)mg/L	$U_{rel}=1.1\%$		
82	*Melting-point Measurement Instruments	Temperature	Verification Regulation of Melting-point Measurement Instruments JJG701	Melting-point: (50~300)°C	$U=0.13^{\circ}\text{C}$		
83	*Bomb Calorimeter	Calorific Value	Verification Regulation of The bomb calorimeter JJG672	(26000~27000) J/g	$U=27\text{J/g}$		
84	*Differential Scanning Calorimeters	Temperature	Verification Regulation of the Differential Scanning Calorimeters JJG936	Fusion Temperature: (50~600)°C	$U=0.06^{\circ}\text{C}$		
		Heat		Fusion heat: (20~110)J/g	$U=0.18\text{J/g}$		
85	*Open/Closed Cup Flash Point Testers	Temperature	Calibration Specification for Open/Closed Cup Flash Point Testers JJF1384	Open cup flash point: (80~260)°C	$U=(6.2\sim 8.8)^{\circ}\text{C}$		
				Closed cup flash point: (60~200)°C	$U=(3.2\sim 4.4)^{\circ}\text{C}$		
86	*Osmometer	Osmometer	Verification Regulation of Osmometers JJG1089	(99~101)mOsmol/kg	$U=1.5\text{mOsmol/kg}$		
				(199~201)mOsmol/kg	$U=1.5\text{mOsmol/kg}$		
				(299~301)mOsmol/kg	$U=1.5\text{mOsmol/kg}$		
				(399~401)mOsmol/kg	$U=1.9\text{mOsmol/kg}$		
				(499~501)mOsmol/kg	$U=2.3\text{mOsmol/kg}$		
				(599~601)mOsmol/kg	$U=2.8\text{mOsmol/kg}$		



No. CNAS L0893

第 138 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				(699~701)mOsmol/kg	$U=3.2\text{mOsmol/kg}$		
87	*Automatic Amino Acid Analyzer	Detection limit	Verification Regulation of Automatic Amino Acid Analyzer JJG1064	histidine: $\leq 1\text{mmol/L}$	$U_{\text{rel}}=20\%$		
88	*Elemental Analyzers	Content	Calibration Specification for Elemental Analyzers JJF 1321	Carbon, hydrogen and nitrogen analyzer、Hydrogen detector C: 50%~80%	$U_{\text{rel}}=0.9\%\sim 0.3\%$		
				Carbon, hydrogen and nitrogen analyzer、Hydrogen detector H: 2%~5%	$U_{\text{rel}}=6.5\%\sim 2\%$		
				Carbon, hydrogen and nitrogen analyzer、Hydrogen detector N: 0.5%~2%	$U_{\text{rel}}=7.5\%\sim 2\%$		
				Full (semi) automatic nitrogen analyzer N: (0.5~20) mg/mL	$U_{\text{rel}}=0.4\%$		
				Oxygen, nitrogen and hydrogen detector O: 0.00058%~0.0120%	$U_{\text{rel}}=19\%\sim 5\%$		
				Oxygen, nitrogen and hydrogen detector N: 0.00075%~0.142%	$U_{\text{rel}}=15\%\sim 3\%$		
				Oxygen, nitrogen and hydrogen detector H: $(0.9\sim 6.6) \times 10^{-6}$	$U_{\text{rel}}=23\%\sim 3.5\%$		



No. CNAS L0893

第 139 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
89	*Determinators for Total Sulfur in Coal	Content	Verification Regulation of Determinators for Total Sulfur in Coal JJG 1006	0.1%~<1.00%	$U=0.02\%$		
				1.00%~4.00%	$U=0.05\%$		
				>4.00%~6.00%	$U=0.06\%$		
90	*Carbon-sulfur Analyzers	Content	Verification Regulation for carbon-sulfur Analyzers JJG 395	Automatic high-speed carbon-sulfur analyzer(C): 0.030%~0.100%	$U=0.001\%$		
				Automatic high-speed carbon-sulfur analyzer(C): >0.100%~0.500%	$U=0.003\%$		
				Automatic high-speed carbon-sulfur analyzer(C): >0.500%~1.000%	$U=0.007\%$		
				Automatic high-speed carbon-sulfur analyzer(C): >1.00%~4.00%	$U=0.01\%$		
				Infrared carbon-sulfur analyzer(C): 0.005%~0.010%	$U=0.0002\%$		
				Infrared carbon-sulfur analyzer(C): >0.010%~0.100%	$U=0.001\%$		
				Infrared carbon-sulfur analyzer(C): >0.100%~1.00%	$U=0.003\%$		



No. CNAS L0893

第 140 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				Infrared carbon-sulfur analyzer(C): $> 1.00\% \sim 4.00\%$	$U=0.01\%$		
				Automatic high-speed carbon-sulfur analyzer(S): $0.003\% \sim 0.010\%$	$U=0.0002\%$		
				Automatic high-speed carbon-sulfur analyzer(S): $> 0.010\% \sim 0.050\%$	$U=0.001\%$		
				Automatic high-speed carbon-sulfur analyzer(S): $> 0.050\% \sim 0.100\%$	$U=0.001\%$		
				Automatic high-speed carbon-sulfur analyzer(S): $> 0.100\% \sim 0.200\%$	$U=0.003\%$		
				Infrared carbon-sulfur analyzer(S): $0.003\% \sim 0.010\%$	$U=0.0002\%$		
				Infrared carbon-sulfur analyzer(S): $> 0.010\% \sim 0.100\%$	$U=0.001\%$		
				Infrared carbon-sulfur analyzer(S): $> 0.100\% \sim 0.200\%$	$U=0.003\%$		
91	*Industrial Analyzers	Content	Verification Regulation of Industrial Analyzers JJG	ash content: $1\% \sim 15\%$	$U=0.10\% \sim 0.11\%$		



No. CNAS L0893

第 141 页 共 181

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
			1140	ash content: $>15\% \sim 30\%$	$U=0.13\% \sim 0.19\%$		
				ash content: $>30\% \sim 40\%$	$U=0.12\%$		
				volatile matter: $1\% \sim 20\%$	$U=0.19\% \sim 0.33\%$		
				volatile matter: $>20\% \sim 40\%$	$U=0.31\% \sim 0.34\%$		
92	*Hand-held X-ray Fluorescence Spectrometer	Detection limit	Calibration Specification for Hand-held X-ray Fluorescence Spectrometer JJF(chuan) 165	Cr: $\leq 0.05\%$	$U_{rel}=48\%$		
				Ni: $\leq 0.05\%$	$U_{rel}=48\%$		
93	*Low-level Total Organic Carbon Analyzer	Concentration	Calibration Specification for Low-level Total Organic Carbon Analyzer JJF(chuan) 141	$(200 \sim 1000) \mu\text{g/L}$	$U_{rel}=5\%$		
94	*Oxidation-reduction Potential Meters	Potential	Calibration Specification for Oxidation-reduction Potential Meters NIMTT(CM) 052	electronic unit: $-2000 \text{ mV} \sim +2000 \text{ mV}$	$U=0.1 \text{ mV}$		
		temperature		solution: $(30 \sim 500) \text{ mV} (25^\circ\text{C})$	$U=0.4 \text{ mV}$		
				$(0 \sim 50)^\circ\text{C}$	$U=0.1^\circ\text{C}$		
95	*Liquid chromatograph-Atomic Fluorescence Spectrometers	Minimum detection quantity	Verification Regulation of Liquid chromatograph-Atomic Fluorescence Spectrometers JJG 1151	As(V) $\leq 1.0 \text{ ng}$; DMA、MMA $\leq 0.7 \text{ ng}$	$U_{rel}=58\%$		
96	*Gas Chromatographs with Helium Ionization Detector	LOD	C. S. for Gas Chromatographs with Discharge Helium Ionization Detector JJF(Chuan) 167	$\leq 10 \text{ pg/s}$	$U_{rel}=14\%$		



No. CNAS L0893

第 142 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
97	*Gas Chromatographs with Discharge Helium Ionization Detector	LOD	Calibration Specification for Gas Chromatographs with Sulfur Chemiluminescence Detector JJF 1953	≤ 10 pg/s	$U_{\text{rel}}=25\%$		
98	*Residual Chlorine Meter	Concentration	C. S. for Residual Chlorine Meter JJF 1609	$(0.1 \sim 10)$ mg/L	$U_{\text{rel}}=1.6\%$		
99	*Alarmer Detectors of Benzene	gas concentration	Calibration Specification for Alarmer Detectors of Benzene JJF 1674	$1 \times 10^{-6} \sim 10 \times 10^{-6}$	$U_{\text{rel}}=5.1\% \sim 2.3\%$		
				$> 10 \times 10^{-6} \sim 100 \times 10^{-6}$	$U_{\text{rel}}=2.1\%$		
100	*Sulfur Hexafluoride Decomposition Products Detectors	gas concentration	Calibration Specification for Sulfur Hexafluoride Decomposition Products Detectors JJF 1711	SO ₂ : $1 \times 10^{-6} \sim 100 \times 10^{-6}$	$U_{\text{rel}}=3.3\% \sim 2.4\%$		
				H ₂ S: $1 \times 10^{-6} \sim 100 \times 10^{-6}$	$U_{\text{rel}}=3.3\% \sim 2.4\%$		
				CO: $10 \times 10^{-6} \sim 500 \times 10^{-6}$	$U_{\text{rel}}=2.5\% \sim 2.3\%$		
101	*Alarms and Detectors of Chloroethylene Gas	gas concentration	Verification Regulation of Alarms and Detectors of Chloroethylene Gas JJG 1125	$10 \times 10^{-6} \sim 100 \times 10^{-6}$	$U_{\text{rel}}=2.2\%$		
102	Optical Cavity Ring-Down Moisture Meter	Dew temperature	Calibration Specification of Optical Cavity Ring-Down Moisture Meter JJF(chuan) 166	$(-90 \sim +20)$ °C	$U=(0.4 \sim 0.2)$ °C		
		moisture content		$(0.096 \sim 23625) \times 10^{-6}$ mol/mol	$U_{\text{rel}}=0.3\% \sim 0.8\%$		
103	*Ammonia-Nitrogen Analyzers	Concentration	Calibration Specification for Off-line Ammonia-Nitrogen Analyzers JJF(chuan)181—2021	$(0.01 \sim 1000)$ mg/L	$U_{\text{rel}}=1.1\%$		
104	*Total Phosphorus Meters	Concentration	Calibration Specification for Total Phosphorus Meters NIMTT(CM) 162	$(0.01 \sim 100)$ mg/L	$U_{\text{rel}}=2.3\%$		



No. CNAS L0893

第 143 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
105	*Total Nitrogen Meters	Concentration	Calibration Specification for Total Nitrogen Meters NIMTT(CM) 161	(0.01~100)mg/L	$U_{rel}=2.3\%$		
106	*Nitrite Meters	Concentration	Calibration Specification for Nitrite Meters NIMTT(CM) 159	Nitrite: (0.001~500)mg/L Nitrite Nitrogen: (0.001~100)mg/L	$U_{rel}=1.4\%$ $U_{rel}=2.5\%$		
107	*Gas-phase Molecular Absorption Spectrometers	detection limit	Calibration Specification for Gas-phase Molecular Absorption Spectrometers NIMTT(CM) 158	Nitrite Nitrogen: ≤ 0.01 mg/L Nitrate Nitrogen: ≤ 0.01 mg/L Sulfide: ≤ 0.01 mg/L Ammonia Nitrogen: ≤ 0.02 mg/L Total Nitrogen: ≤ 0.05 mg/L	$U_{rel}=45\%$ $U_{rel}=45\%$ $U_{rel}=45\%$ $U_{rel}=45\%$ $U_{rel}=45\%$		
108	*Water Quality Meters of Heavy Metals	Concentration	Calibration Specification for Water Quality Meters of Heavy Metals NIMTT(CM) 160	Cr(VI), Cr, Fe, Cu, Ni, Zn, Pb, Cd, Hg, As, Mn: (0.001~100)mg/L	$U_{rel}=0.5\%\sim 1.6\%$		
109	*Ultraviolet Fluorescence Sulfur Analyzers	Concentration	Calibration Specification for Ultraviolet Fluorescence Sulfur Analyzers JJF 1685	(1~10)mg/L (>10~200)mg/L	$U=0.11$ mg/L $U_{rel}=2.2\%$		
110	*Karl Fischer Volumetric Titrators for Water Content	quality	Verification Regulation of Karl Fischer Volumetric Titrators for Water Content JJG 1154	(1~20)mg	$U_{rel}=(1.6\sim 1.3)\%$		



No. CNAS L0893

第 144 页 共 181

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
111	*Energy Dispersive X-Ray Fluorescence Spectrometers	detection limit	Calibration Specification for Energy Dispersive X-Ray Fluorescence Spectrometers JJF (min) 1047	≤ 30 mg/kg	$U_{rel}=45\%$		
112	*Urea Analyzers of Water Sample Detection	Urea	Calibration Specification for Urea Analyzers of Water Sample Detection JJF1822	(0.1~5)mg/L	$U_{rel}=0.6\%$		
113	*on-line Automatic Monitors of Permanganate Index	Concentration	Calibration Specification for on-line Automatic Monitors of Permanganate Index JJF1875	(0.01~20)mg/L	$U_{rel}=3.2\%$		
114	*Sulfur Dioxide Gas Detectors	gas concentration	Verification Regulation of Sulfur Dioxide Gas Detectors JJG 551	$2 \times 10^{-6} \sim 5000 \times 10^{-6}$	$U_{rel}=1.8\% \sim 1.4\%$		
115	Dust Concentration Measuring Instruments	dust concentration	Dust Concentration Measuring Instruments JJG 846	(0.01~50)mg/m ³	$U_{rel}=(8.5 \sim 7.4)\%$		
116	*Dilution device	flow	Calibration Specification for Dilution device NIMTT(CM)016	(5~50000)mL/min	$U_{rel}=0.33\%$		
		dilution ratio		0.1%~100%	$U_{rel}=0.96\%$		
117	*Pour Point and Cloud Point Testers of Petroleum Product	Temperature	Calibration Specification for Pour Point and Cloud Point Testers of Petroleum Products JJF 1869	Pour Point: (-30.0~0)°C	$U=2.8^{\circ}\text{C}$		
				Cloud Point: (-17.0~2.0)°C	$U=1.8^{\circ}\text{C}$		
118	McFarland Bacterial Turbidity Analyzer	Turbidity	Calibration Specification for McFarland Bacterial Turbidity Analyzers JJF 1825	(0.1~6)MCF	$U=(0.04 \sim 0.22)\text{MCF}$		
119	*Nucleic acid extraction	Temperature	(Automatic) nucleic acid extractor calibration	(30~90)°C	$U=0.5^{\circ}\text{C}$		



No. CNAS L0893

第 145 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	apparatus	Frequency	specification JJF 1874	(1~100)Hz	U=0.6Hz		
		Volume		(10~1000) μ L	U=1.3 μ L		
		Recovery rate		(40~120)%	U=2.0%		
120	*Colony Counter	Colony count	Calibration Specifications for Colony Counter JJF 1751	(20~350)CFU	U _{rel} =3%		
121	*Polymerase Chain Reaction Analyzers	Temperature	Calibration Specification for Polymerase Chain Reaction Analyzers JJF 1527	(10~110)℃	U=(0.20~0.34)℃		
		Concentration		1×10 ² ~1×10 ⁸ copies/μ L	U _{rel} =6.4%		
122	*Gel imaging system	Illuminance	Gel imaging system calibration specifications JJF 1530	(50~3000)lx	U _{rel} =3%		
123	*Thermogravimetric Analyzers	mass	Verification Regulation of Thermogravimetric Analyzers JJG 1135	(0~20) mg	U=0.15 mg		
		temperature		(100~800)℃	U=2 ℃		
X Ionizing radiation measuring instrument							
1	α、β Surface Contamination Monitors	Activity	Verification Regulation of α、β Surface Contamination Monitors JJG 478	α：（2.0×10 ² ~7.84×10 ⁵ ）(min•2πsr) ⁻¹	U _{rel} =10%		
				β：（3.0×10 ³ ~8.15×10 ⁵ ）(min•2πsr) ⁻¹	U _{rel} =10%		
2	Radon Measuring Instruments	Activity	Verification Regulation of Radon Measuring Instruments JJG 825	(100~12000) Bq/m ³	U _{rel} =7.1%		
3	Radioactivity Meter	activity	Verification Regulation of Radioactivity Mete JJG 377	(3.7×10 ⁴ ~3.7×10 ¹⁰)Bq	U _{rel} =4.5%		
4	*Gamma Radio-immunoassay Counters	activity	Verification Regulation of Gamma Radio-immUnoassay CoUnters JJG 969	(500~3300)Bq	U _{rel} =6.0%		



No. CNAS L0893

第 146 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
5	**Germanium Gamma-ray Spectrometers	activity	Calibration Specification for Germanium Gamma-ray Spectrometers JJF 1850	$(10 \sim 4 \times 10^4) \text{Bq}$	$U_{\text{rel}}=5.0\%$		
6	*Low Background Alpha/Beta Measuring Instruments	Activity	Verification Regulation of Low Background Alpha/Beta Measuring Instruments JJG 853	$\alpha: (0.005 \sim 2.46 \times 10^4) (\text{min} \cdot 2\pi)^{-1}$ $\beta: (0.15 \sim 1.45 \times 10^4) (\text{min} \cdot 2\pi)^{-1}$	$U_{\text{rel}}=7.0\%$ $U_{\text{rel}}=7.0\%$		
7	*Medical Diagnostic X-ray Radiation Source for Spiral Computed Tomography (CT)	dose index	Verification Regulation of	$(1.0 \sim 1000) \text{mGy}$	$U_{\text{rel}}=7.5\%$		
		CT number	Medical Diagnostic X-ray Radiation Source for Spiral	$(-4 \sim 4) \text{HU}$	0.1HU		
		slice thickness	Computed Tomography (CT) JJG 961	$(0.5 \sim 15) \text{mm}$	$U=0.04 \text{mm}$		
8	*Medical Diagnostical X-ray RadiationSource	Air kerma	Verification Regulation of Medical Diagnostical X-ray Radiation SoUrce JJG 744	$(0.01 \sim 199) \text{mGy}$	$U_{\text{rel}}=4.2\%$		
9	*Medical Diagnostical X-ray RadiationSource	Air kerma	Verification Regulation of Medical Diagnostic X-ray Source for Dental Panorama JJG 1101	$(0.01 \sim 199) \text{mGy}$	$U_{\text{rel}}=4.2\%$		
10	*Medical Diagnostical X-ray RadiationSource	Air kerma	Verification Regulation of Medical Diagnostic X-ray Radiation Source for Medical Digital Subtraction Angiography JJG 1067	$(0.01 \sim 199) \text{mGy}$	$U_{\text{rel}}=4.2\%$		
11	*Medical Diagnostical X-ray RadiationSource	Air kerma	Verification Regulation of X-ray Radiation Sources for Medical Computed Radiography System and Digital Radiography System JJG 1078	$(0.01 \sim 199) \text{mGy}$	$U_{\text{rel}}=4.2\%$		



No. CNAS L0893

第 147 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
12	*Medical Diagnostical X-ray Radiation Source	Air kerma	Verification Regulation of X-ray Radiation Source for Radiotherapy Simulating Localization JJG 1028	(0.01~199)mGy	$U_{rel}=4.2\%$		
13	*Apparatus for Gamma Radiography	Air kerma rate	Verification Regulation of Apparatus for Gamma Radiography JJG 933	(0.01~10)Gy/min	$U_{rel}=5.0\%$		
14	*X-ray Flaw Detectors	Air kerma rate	Verification Regulation of X-ray Flaw Detectors JJG 40	(0.01~10)Gy/min	$U_{rel}=5.0\%$		
15	*Radiotherapy unit	Absorbed Dose	Verification Regulation of Radiation Source Used in the External Beam Radiotherapy JJG 589	Photon Beam:(0.01~10.00)Gy	$U_{rel}=3.8\%$		
				Electron Beam:(0.01~10.00)Gy	$U_{rel}=4.0\%$		
		Absorbed Dose rate		Photon Beam:(0.01~10.0)Gy/min	$U_{rel}=3.8\%$		
				Electron Beam:(0.01~10.0)Gy/min	$U_{rel}=4.0\%$		
16	Portable Ambient Dose Equivalent (Rate) Meters and Monitors for X and γ Radiations	air kerma	Verification Regulation of Portable Ambient Dose Equivalent (Rate) Meters and Monitors for X and γ Radiations JJG 393	1 μ Gy~1 Gy	$U_{rel}=3.8\%$		
		Air kerma rate		1 μ Gy/h~400 mGy/h	$U_{rel}=3.8\%$		
		dose equivalent		10 μ Sv~1 Sv	$U_{rel}=5.6\%$		
		dose equivalent rate		1 μ Sv/h~700 mSv/h	$U_{rel}=5.6\%$		
17	Personal dose equivalent Hp(10) Monitors for X and γ Radiations	dose equivalent	Verification Regulation of Personal dose equivalent Hp(10) Monitors for X and γ Radiations JJG 1009	10 μ Sv~1 Sv	$U_{rel}=5.6\%$		
		dose equivalent rate		1 μ Sv/h~700 mSv/h	$U_{rel}=5.6\%$		



No. CNAS L0893

第 148 页 共 181

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
18	*X-ray Security Inspection Equipment	Air kerma rate	Calibration Specification for X-ray Security Inspection Equipment JJF 1275	0.01 μ Sv/h~50mSv/h	$U_{rel}=8.7\%$		
19	*Installed Personnel α 、 β Surface Contamination Monitoring Assemblies	activity	Verification Regulation of Installed Personnel α 、 β Surface Contamination Monitoring Assemblies JJG 1102	α : $(10^4 \sim 10^6) / (\min 2 \pi \text{ sr})$ β : $(10^4 \sim 10^6) / (\min 2 \pi \text{ sr})$	$U_{rel}=6.0\%$		
20	Diagnostic dosimeters for X-ray	Air kerma rate	Calibration Specification for Diagnostic Dosimeters	$(6 \times 10^{-5} \sim 1) \text{ Gy/min}$	$U_{rel}=2.0\%$		
		Air kerma lenth	JJF1621,Calibration Specification in Diagnostic Dosimeters for X-ray	$(1 \times 10^{-6} \sim 50) \text{ Gycm}$	$U_{rel}=2.0\%$		
		tube voltage	NIMTT(CM)032,Calibration Specification for Non-	$(20 \sim 50) \text{ kV}$	$U_{rel}=1.0\%$		
			invasive X-ray Tube Voltage Meters Used in Medical Diagnosis JJF1474,Calibration Specification for Medical	$(> 50 \sim 150) \text{ kV}$	$U_{rel}=0.8\%$		
		Exposure time	Non- invasive X-ray Tube Voltage Testers NIMTT(CM) 042,Calibration Specification for Medical Diagnostics X-ray Non- invasive Exposure Time JJF1432	$(10 \sim 6000) \text{ ms}$	$U=0.05 \text{ ms}$		
21	Dose Área Product	Dose area product	Calibration Specification for Dose Area Product Meters JJF1479	$(0.01 \sim 5 \times 10^3) \mu \text{ Gym}^2$	$U_{rel}=2.0\%$		



No. CNAS L0893

第 149 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
22	Monitor of Radiouclide Identification	Dose equivalent rate	Calibration Specification for Hand-held Radiation Monitors for Detection and Identification of Radianuclides JJF1687	$(2 \times 10^{-6} \sim 1 \times 10^{-4})$ Sv/h	$U_{rel}=4.8\%$		
23	γ Ray Spectrometers of Scintillation Detectors	activity	Calibration Specification of γ Ray Spectrometers of Scintillation Detectors JJF1744	$(5 \times 10^2 \sim 1 \times 10^4)$ Bq	$U_{rel}=12\%$		
24	real-time focus meter	focal spot	Calibration Specification for Real-time Focus Meters JJF1688	(0.1 ~ 6.0) mm	$U=0.04\text{mm}$		
25	3D-watertank	position	Calibration Specification for 3D-Watertank NIMTT(CM) 049	(0 ~ 200) mm	$U=0.05\text{mm}$		
26	Half-value Layer Instrument	Half-value Layer	Calibration Specification for Half-value Layer Instrument NIMTT 045	(0.1 ~ 14)mmAl	$U_{rel}=2.0\%$		
27	*Medical X-ray Radiation Sources for Mammographic Equipment	Absorbed dose	Verification Regulation of Medical X-ray Radiation Sources for Mammographic Equipment JJG1145	(0.1 ~ 199)mGy	$U_{rel}=4.2\%$		
XI Special measuring instrument							
1	*Loading Method Automobile Brake Testers	Lifting Height	Loading Method Automobile Brake Testers JJG 1160	(0 ~ 500) mm	$U=0.5\text{mm}$		
		Braking force		(1 ~ 50)kN	$U_{rel}=1.0\%$		
		Slip adhesion coefficient of roller		0.01 ~ 1	$U=0.012$		



No. CNAS L0893

第 150 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Slip ratio of driving motor stop		5%~40%	$U=4.8\%$		
		Mass		1kg~15t	$U_{rel}=0.62\%$		
2	*Vehicle Contour Dimensions Testers	Length	Calibration Specification of Vehicle Contour Dimensions Testers JJF 1749	(1~30)m	$U_{rel}=0.4\%$		
3	Detecting device of vehicle travelling data recorder	Speed	Calibration Specification for Detecting device of vehicle travelling data recorder NIMTT(CM) 029	(20~180) km/h	$U_{rel}=0.6\%$		
		Mileage		(1~5000) m	$U_{rel}=0.6\%$		
		Time		(0~3600) s	$U=0.7s$		
4	*Chassis Dynamometers for Automobile Emissions	torque force		(1N~8kN	$U_{rel}=0.8\%$		
		speed		(0~80)km/h	$U=0.12km/h$		
		dynamometer inertia weight		(1~2000)kg	$U_{rel}=0.62\%$		
		slipping time under fixed load		(1~150)s	$U_{rel}=0.28\%$		
		slipping time under varying load	Calibration Specification for Chassis Dynamometers for Automobile Emissions Testing JJF 1221	(1~150)s	$U_{rel}=0.34\%$		
		load response time		(0~150)s	$U=12ms$		
		mean time between failure		(0~150)s	$U=12ms$		
		internal loss power		(0.1~2.5)kW	$U_{rel}=3.8\%$		




在线扫码获取验证

No. CNAS L0893

第 151 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
5	*Slip plate type automobile side slip tester	indication error of side slip	Verification Regulation of Slip plate type automobile side slip tester JJG 908	(-10~10)m/km	$U=0.07\text{m/km}$		
6	*Tester for wheel deviation of motorcycles	indication error of wheel deviation	Verification Regulation of Slip plate type automobile side slip tester JJG 910	(1~10) mm	$U=(0.07\text{mm})$		
7	*Tester for wheel deviation of motorcycle	Mass	Verification Regulation of Tester for wheel deviation of motorcycles JJG 1014	1kg~15t	$U_{\text{rel}}=0.62\%$		
8	*Roller type speedometer tester	indication error of speed	Verification Regulation of Roller type speedometer tester JJG 909	(1~120) km/h	$U_{\text{rel}}=0.32\%$		
9	*Roller opposite forces type brake tester	Brake force	Verification Regulation of Roller opposite forces type brake tester JJG 906	(1~50)kN	$U_{\text{rel}}=1.0\%$		
		Adhesion Coefficient		0.01~1	$U=0.012$		
		Slip Ratio		5%~40%	$U=4.8\%$		
10	*Head lamp tester for motor vehicle	luminous intensity	Verification Regulation of Headlamp tester for motor vehicle JJG 745	(5~60)kcd	$U_{\text{rel}}=5.0\%$		
		offset of ray axes		up:1°~down:2°, Left:2°~right:2°	$U=5.2'$		
11	*Filter-Type Smoke Meter	Smoke thickness	Verification Regulation of Filter-type smoke meter JJG 847	(1~10)BSU	0.22BSU		
12	*Vehicle exhaust emission measuring instruments	Exhaust emission concentration	Verification Regulation of vehicle exhaust emission measuring instruments JJG 688	$\text{CO}:(1\sim16)\times10^{-2}\text{mol/mol}$	$U_{\text{rel}}=1.4\%$		
				$\text{C}_3\text{H}_8:(1\sim9999)\times10^{-6}\text{mol/mol}$	$U_{\text{rel}}=1.2\%$		
				$\text{CO}_2:(1\sim18)\times10^{-2}\text{mol/mol}$	$U_{\text{rel}}=1.2\%$		



No. CNAS L0893

第 152 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				NO: $(1 \sim 5000) \times 10^{-6}$ mol/mol	$U_{\text{rel}}=1.3\%$		
				O ₂ : $(1 \sim 25) \times 10^{-2}$ mol/mol	$U_{\text{rel}}=1.5\%$		
13	Vehicle travelling data recorder	Velocity	Verification Regulation of Vehicle travelling data recorder JJG(CHUAN)90	(0~180) km/h	$U=0.3\text{km/h}$		
		Mileage		(0~10)km	$U=0.02\text{km}$		
14	Calibrators for headlamp tester of motor vehicle	luminous intensity	Verification Regulation of Calibrators for headlamp tester of motor vehicle JJG 967	(1~60)kcd	$U_{\text{rel}}=1.5\%$		
		light axis		up:3°~down:3°, Left:3°~right:3°	$U=1.9'$		
15	Motor Vehicle Testers for Steering Force and Steering Angle	Steering force	Calibration Specification for Motor Vehicle Testers for Steering Force and Steering Angle JJF 1196	(1~1000)N	$U_{\text{rel}}=0.6\%$		
		Steering angle		(0~1080°)	$U=1.0^\circ$		
16	*Turning Angle Testers for Automobile	angle	Calibration Specification for Turning Angle Testers for Automobile JJF 1141	(-50°~50°)	$U=0.32^\circ$		
17	Non-contact Automotive Speedmeter	Speed	Calibration Specification for Non-contact Automotive Speedmeter JJF 1193	(10~50)km/h	$U=0.30\text{ km/h}$		
				(>50~180) km/h	$U_{\text{rel}}=0.30\%$		
		Distance		(1.0~30)m	$U=0.10\text{m}$		
				(>30~9999.9) m	$U_{\text{rel}}=0.32\%$		
18	*Automotive Suspension Tester	Mass	Calibration Specification for Automotive Suspension Tester JJF 1192	1kg~15t	$U_{\text{rel}}=0.62\%$		
		vibration frequency		(0.83~1333.3)Hz	$U_{\text{rel}}=0.96\%$		
19	*Four-wheel Aligner	angle	Calibration Specification for Four-wheel Aligners JJF 1154	toe: -3°~3°	$U=1.2'$		

No. CNAS L0893

第 153 页 共 181



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				camber: -10° ~10°	U=2.8'		
				caster: -15° ~15°	U=3.6'		
20	*Opacimeter	light absorbance	Verification Regulation of Opacimeters JJG 976	light absorbance N: (0~98.6) % light absorption factor K: (0~9.99)m-1	U=0.68% U=0.016 m-1		
21	portable braking performance tester for vehicles	deceleration	Calibration Specification for portable braking performance tester for vehicles JJF 1168	static: (0~4.9) m/s² static: (>4.9~9.8) m/s² dynamic: (1~9.8) m/s²	U=0.04 m/s² U _{rel} =0.82 % U _{rel} =1.6 %		
22	Automobile fuel consumption meter	mass	Calibration Specification for Oil Consumption Meters of Mass Method JJF 1670,Verification Regulation of Calibration Specification for Four Pistons Linkage Type Consumption Meter JJG (Traffic) 009	(1~160)g	U _{rel} =0.16%		
		volume		(1~500)mL	U _{rel} =0.16%		
				(>500~1500)mL	U _{rel} =0.26%		
23	Calibration Specification for Transmittance Meter of Automobile	transmittance	Calibration Specification for Transmittance Meter of Automobile JJF1225	0~100%	U=0.40%		
24	*Platform Brake Tester	Brake force	Verification Regulation of Platform Brake Tester JJG 1020	(1~50)kN	U _{rel} =0.9%		
		Mass		1kg~15t	U _{rel} =0.62%		
		adhesion coefficient		0~1	U=0.011		



No. CNAS L0893

第 154 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		levelness		(0～50) mm/m	<i>U</i> =0.8mm/m		
25	*Flow Analyzer for Short Transient Loaded Mode of Gasoline Vehicle	Flow	Calibration Specification for Flow Analyzer for Short Transient Loaded Mode of Gasoline Vehicles JJF 1385	(95～180)L/s	<i>U</i> _{rel} =2.0%		
		concentration		5.0%～20.9%	<i>U</i> =1.6%		
26	*Motor Vehicle Engine speed Measuring Instrument	Rotate speed	Calibration Specification for Motor Vehicle Engine Speed Measuring Instruments JJF 1375	(500～6000)r/min	<i>U</i> _{rel} =0.32%		
27	Manipulating force tester for automotive brake	Force value	Calibration Specification for Manipulating force tester for automotive brake JJF 1169	(1～1000)N	<i>U</i> _{rel} =1.2%		
28	*Bump Testing Machines	Acceleration	Shock and Bump Testing Machines JJG 1174	(0.5～1000)m/s ²	<i>U</i> _{rel} =5.0%		
		Time		(1～100) ms	<i>U</i> _{rel} =3.0%		
29	Pyranometer	Sensitivity	Verification Regulation of Pyranometer JJG 458	(7～20) μ V · W ⁻¹ · m ²	<i>U</i> _{rel} =2.4%		
		Irradiance		(1～2000)W · m ⁻²	<i>U</i> _{rel} =2.4%		
30	SphygmomanometersDigital electronic Sphygmomanometers	PressUre	Non-Automated Sphygmomanometers JJG 692	(0～40) kPa	<i>U</i> =0.14kPa		
31	*Multi-Parameter Patient Monitor	heart rate	Verification Regulation for Multifunction Patient Monitoring Instruments JJG1163	(30～300)min ⁻¹	<i>U</i> =1.4min ⁻¹		
		pressure		(0.1～30)kPa	<i>U</i> =0.16kPa		
		oxygen saturation		40%～<75%	<i>U</i> =3.5%		
				75%～100%	<i>U</i> =2.4%		



No. CNAS L0893

第 155 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Respiratory rate		(10~120) min ⁻¹	$U=0.2 \text{ min}^{-1}$		
32	*cardiac defibrillators	tidal volume	Calibration Specification for Cardiac Defibrillators JJF1149	(0~<100)J	$U=3.6J$		
				(100~360)J	$U=5.7J$		
		heart rate		(30~300)min ⁻¹	$U=1.4\text{min}^{-1}$		
33	*Electroencephalograph	Volt	Verification Regulation of Electroencephalograph JJG1043	30 μV ~30V	$U=15 \mu\text{V}$		
34	respiratory rhythm generator	respiratory rate	Calibration Specification for respiratory rhythm generator NIMTT(CM)043	(3~60)min ⁻¹	$U=0.22\text{min}^{-1}$		
35	Defibrillators Analyzer	heart rate	Calibration Specification for Defibrillator Analyzers JJF 1860	(30~300) min ⁻¹	$U_{\text{rel}}=0.08\%$		
		Respiratory		(10~2000) Ω	$U_{\text{rel}}=0.84\%$		
		output energy		(2~360) J	$U_{\text{rel}}=1.4\%$		
		Amplitude		(0.1~3) mV	$U_{\text{rel}}=1.2\%$		
		Square wave signal frequency		(0.125~10)Hz	$U_{\text{rel}}=0.08\%$		
36	*lung ventilators	tidal volume	Calibration Specification for Lung Ventilators JJF1234	(10~1000)mL	$U= (0.023V+1.7) \text{ mL}$		
		frequency		(1~150)min ⁻¹	$U= (0.0024f+0.14) \text{ min}^{-1}$		
		Pressure		(0.1~12)kPa	$U= (0.0035P+0.014) \text{ kPa}$		
37	hemodialysis equipment	conductivity	Calibration Specification for Hemodialysis Equipment JJF1353	(13.5~14.5) mS/cm	$U_{\text{rel}}=1.2\%$		
		temperature		(35~ 40) $^{\circ}\text{C}$	$U_{\text{rel}}=0.60\%$		



No. CNAS L0893

第 156 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		pH		(7.1~7.5)	<i>U</i> _{rel} =0.78%		
		Flow		(450~550) mL/min	<i>U</i> _{rel} =1.5%		
38	*Baby IncuBator	temperature	Calibration Specification for Baby Incubator JJF1260	(25~40) °C	<i>U</i> =0.14℃		
		relative humidity		(25~85) %	<i>U</i> =3.6%		
39	*Electrosurgical Generator	output power	Calibration Specification for Electrosurgical Generator JJF1217	(1~400) W	<i>U</i> = (0.063 <i>P</i> -0.12) W		
		leakage current		(30~300) mA	<i>U</i> _{rel} =3.0%		
40	*the pulmonary function Measuring instrument	VC	Calibration Specification for the Pulmonary Function Measuring Instrument JJF1213	(0.5~<2) L	<i>U</i> =0.026L		
				(2~8)L	<i>U</i> _{rel} =0.74%		
		Flow		(0~<3) L/s	<i>U</i> =0.072 L/s		
				(3~14) L/s	<i>U</i> _{rel} =2.4%		
41	Multiparameter Physiological Simulators	Amplitude of ECG signal	Calibration Specification for Multiparameter Physiological Simulators JJF1470	(0.1~3.0) mV	<i>U</i> _{rel} =1.2%		
		heart rate		(30~300) min ⁻¹	<i>U</i> _{rel} =0.08%		
		Respiratory impedance		(100~3000) Ω	<i>U</i> _{rel} =0.84%		
		Analogue signal of blood pressure		(0.1~200) mV	<i>U</i> _{rel} =0.7%		
		Cardiac output signal		(10~20) kΩ	<i>U</i> _{rel} =0.8%		
		Temperature analog signal		(1~100) kΩ	<i>U</i> _{rel} =0.7%		



No. CNAS L0893

第 157 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Amplitude of Reference waveforms		(0.1~3) mV	$U_{rel}=1.2\%$		
		rate of Reference waveforms		(0.05~150)Hz	$U_{rel}=0.08\%$		
42	Medical Syringe Pump and Infusion Pump Analyzers	flow rate	Verification Regulation of Medical Syringe Pump and Infusion Pump Analyzers JJG1098	(5~<20)mL/h	$U=0.06$ mL/h		
				(20~200) mL/h	$U=0.3$ mL/h		
				(>200~1000) mL/h	$U=0.98$ mL/h		
		occlusion pressure		(1~200) kPa	$U_{rel}=0.8\%$		
43	Precision microinjection pump	flow rate	Calibration Specification for Precision microinjection pump NIMTT(CM) 098	(1~1000) mL/h	$U=0.01$ mL/h		
44	Electrosurgical Generator Analyzers	power	Calibration Specification for Electrosurgical Generator Analyzers NIMTT(CM) 104	(1~300) W	$U_{rel}=1.2\%$		
		leakage current		(5~1000) mA	$U_{rel}=0.42\%$		
		Respiratory		(50~10000) Ω	$U_{rel}=0.84\%$		
45	Spo2 Simulator	oxygen saturation	Calibration Specification for Spo2 Simulator JJF1542	35%~100%	$U=1.4\%$		
		pulse rate value		(30~250) min ⁻¹	$U=1.2$ min ⁻¹		
46	Baby Incubator Detector	temperature	Calibration Specification for Baby Incubator Analyzers NIMTT(CM) 096	(20~50) °C	$U=0.04^{\circ}$ C		
		relative humidity		(10%~90%)	$U=1.3\%$		
47	Hemodialysis Equipment Tester	conductivity	Calibration Specification for Hemodialysis Equipment Tester JJF1541	(12.5~15.5)mS/cm	$U=0.04$ mS/cm		



在线扫码获取验证

No. CNAS L0893

第 158 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		temperature		(10~50)°C	$U=0.03^{\circ}\text{C}$		
		pressure		(-110~150)kPa	$U=0.12\text{kPa}$		
		pH		(0~14)	$U=0.01$		
		Flow		(100~2000) mL/min	$U_{\text{rel}}=0.4\%$		
48	*Medical Magnetic Resonance Image	Magnetic field intensity	Calibration Specification of Medical Magnetic Resonance Image NIMTT(CM) 047	(0.02~2)T	$U_{\text{rel}}=2.5\%$		
49	*Electrocardiogram machine	Volt	Electrocardiograph JJG543	30 μV ~30V	$U=15\mu\text{V}$		
50	Lung Ventilators Analyzer	flow	Calibration Specifications of Lung Ventilator Analyzer NIMTT(CM) 099	(0.5~180) L/min	$U_{\text{rel}}=1.0\%$		
		tidal volume		(5~2000) mL	$U_{\text{rel}}=1.0\%$		
		pressure		(-2~12) kPa	$U=0.012\text{kPa}$		
		Breathing rate		(1~150) min ⁻¹	$U_{\text{rel}}=1.0\%$		
		oxygen concentration		(21%~100%)	$U=1.0\%$		
51	Standard respiratory simulator	VC	Standard respiratory simulator of calibration specifications NIMTT(CM) 040	(0.1~9) L	$U=(1.2+0.67V)\text{ mL}$		
		PEF		(1~100) Nm ³ /h	$U_{\text{rel}}=1.4\%$		
52	NIBP Simulators	Analogue blood pressure indication repeatability	Calibration Specification for NIBP Simulators JJF1626	(1.3~34.0) kPa	$U=0.08\text{ kPa}$		



No. CNAS L0893

第 159 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		Pulse rate		(30~250) min ⁻¹	$U_{rel}=1.0\%$		
		Static pressure		(1~53.3) kPa	$U=0.052$ kPa		
		Volume		(50~3000) mL	$U=0.19\text{mL}\sim 1.7\text{mL}$		
53	Ultrasonic Partial Discharge Tester	Frequency	Calibration Specification for Ultrasonic Partial Discharge Tester NIMTT(CM) 127	(20~500)kHz	$U_{rel}=0.1\%$		
		Amplitude linear		-200dB~200dB	$U=1.8\text{dB}$		
		Sensitivity		-80dB~85dB	$U_{rel}=1\%$		
		Stability		-200dB~200dB	$U=1.8\text{dB}$		
54	electrostaticdischarge simulators	charging voltage	calibration specification for electrostatic discharge simulators JJF1397	$\pm(0.1\sim 30)\text{kV}$	$U_{rel}=3.2\%$		
		discharge current(peak)		$\pm(0.1\sim 30)\text{A}$	$U_{rel}=6\%$		
		rise time		(0.7~1.0)ns	$U_{rel}=13\%$		
		discharge current		$\pm(0.1\sim 30)\text{A}$	$U_{rel}=10\%$		
55	Electric Fast Transient Burst generator	Voltage(Peak)	calibration specification for electrical fast transient/burst simulators JJF1672	(0.25~4)kV	$U_{rel}=5\%$		
		Single pulse rise time		(1~10)ns	$U_{rel}=7\%$		
		Single pulse duration time		(10~150)ns	$U_{rel}=7\%$		
		repetition frequency		(5~100)kHz	$U_{rel}=6\%$		
		Burst Duration		(0.75~300)ms	$U_{rel}=8\%$		
56	Surge generator	Voltage(Peak)	Calibration Specification for Surge Simulators JJF 1741	0.1kV~6kV	$U_{rel}=4\%$		


No. CNAS L0893

第 160 页 共 181



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		Voltage rise time		(0.5~10) μ s	$U_{rel}=4\%$		
		Voltage duration time		(10~1000) μ s	$U_{rel}=4\%$		
		current(peak)		0.01kA~8kA	$U_{rel}=3\%$		
		current rise time		(0.5~10) μ s	$U_{rel}=4\%$		
		current duration time		(10~1000) μ s	$U_{rel}=4\%$		
57	Voltage dips, short interruptions and voltage variations generator	voltage amplitude	calibration specification for voltage dips, short interruptions and voltage variation tests generator JJF1673	0.1V~400V	$U_{rel}=2\%$		
		Time at Reduced Voltage		1ms~10s	$U_{rel}=6\%$		
		Load current		0.1A~40A	$U_{rel}=5\%\sim 7\%$		
58	*Power frequency magnetic field generator	current	Calibration Specification for Power Frequency Magnetic Field Simulators JJF 1737	(0.01~1000)A	$U_{rel}=3\%$		
		Magnetic field intensity		(1~5000)A/m	$U_{rel}=11\%$		
59	*Impulse Voltage generator	Voltage(Peak)	calibration specification for impulse voltage generator NIMTT(CM) 122	1V~7000V	$U_{rel}=4\%$		
		Voltage rise time		1ns~1s	$U_{rel}=4\%$		
		Voltage duration time		0.1 μ s~60s	$U_{rel}=4\%$		
		current (Peak)		0.1A~8000A	$U_{rel}=3\%$		
		current rise time		5ns~1s	$U_{rel}=4\%$		
		current duration time		0.1 μ s~60s	$U_{rel}=4\%$		



在线扫码获取验证

No. CNAS L0893

第 161 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
60	Microwave radiation and leakage measuring instrument(Including Microwave field probe)	Power density	Verification Regulation for Microwave Radiation and Leakage Energy Measuring Instruments JJG776	(1~1000) μ W/cm ² , (915MHz,2450MHz)	U=1.7dB		
				(1000~5000) μ W/cm ² , (915MHz,2450MHz)	U=1.7dB		
61	Current Probe(EMC)	Transfer impedance	calibration specification for Current probe NIMTT(CM) 064	(-40~80)dB Ω , (10kHz~10MHz)	U=0.5dB		
				(-40~80)dB Ω , (10MHz~200MHz)	U=0.6dB		
62	Voltage Probe(EMC)	Insert Loss	calibration specification for voltage probe NIMTT(CM) 065	(0~80)dB, (10kHz~10MHz)	U=2.0dB		
				(0~80)dB, (10MHz~200MHz)	U=2.3dB		
63	Calibration Device for Electrocardiograph and Electroencephalograph	Distortion	Verification Regulation of Verification Instrument for Electrocardiograph and Electroencephalograph JJG749	0.1%~30%, (10Hz~200Hz)	U _{rel} =10%		
		Voltage		0.1mV~20V	U _{rel} =0.7%		
		Frequency		0.01Hz~500Hz	U _{rel} =8×10 ⁻⁵		
64	vehicles transient disturbance/load dump simulators	voltage	calibration specification for vehicles transient disturbance simulators NIMTT(CM) 008	(-700~-1)V, (1~300)V	U _{rel} =4%		
		rise time		1ns~10ms	U _{rel} =7%		
		duration time		1ms~5s	U _{rel} =4%		
65	*Tester for Ductility of Bituminous Materials	Temperature	Verification Regulation of Tester for Ductility of Bituminous Materials JJG (Traffic) 023	(0~50) °C	U=0.12°C		
66	*Apparatus for	geometrical dimensions	Verification Regulation of Apparatus for Softening Point	(0~200) mm	U=0.05mm		



No. CNAS L0893

第 162 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
	of Bitumen	weight	of Bitumen JJG (Traffic) 057	(0~200) g	$U=1.2\text{mg}$		
		heating rate		(0~20) °C/min	$U=0.4\text{°C/min}$		
		Temperature		(0~200)°C	$U=0.07\text{°C}$		
67	*Asphalt Aging Oven	geometrical dimensions	Verification Regulation of Asphalt Aging Oven JJG (Traffic) 056	(0~200) mm	$U=0.06\text{mm}$		
		Temperature		(0~300)°C	$U=0.3\text{°C}$		
68	*Asphalt Mixture's Mixing Machine	Temperature	Verification Regulation of Asphalt Mixture's Mixing Machine JJG(Traffic) 064	(0~250) °C	$U=0.6\text{°C}$		
		heating rate		(0~250) °C/min	$U=0.4\text{°C/min}$		
69	*Verification Regulation of Vibrator for Compacting Mortar Specimen	Frequency	Calibration Specification for Cement Mortar Vibration Compaction Equipment JJF 1867-2020	(5~5000)Hz	$U_{\text{rel}}=1.0\%$		
		Displacement		(10~1000) μ m	$U_{\text{rel}}=5.0\%$		
70	*Vibrating Table for Concrete Test	Frequency	Calibration Specification of Vibrating Table for Concrete Test JJF(chuan) 170	(20~100)Hz	$U=2.0\%$		
		Amplitude		(0.1~10)mm	$U=3.0\%$		
71	Pile Dynamic Measuring Instruments	Acceleration	V. R. of Pile Dynamic Measuring Instruments JJG 930	(0.1~300)m/s ²	$U_{\text{rel}}=3.0\%$		
		Velocity		(0.1~50)cm/s	$U_{\text{rel}}=3.0\%$		
		Dynamic strain		(10~1000)μ ε	$U_{\text{rel}}=5.0\%$		
		Frequency		(10~2000)Hz	$U_{\text{rel}}=1.0\%$		
72	Apparatus for Determining Penetration of	Length	C. S. for Apparatus for Determining Penetration of Bituminous Materials JJF	(0~100)mm	$U=0.1\text{mm}$		
				Φ: (0.01~20)mm	$U=(3+6L) \mu\text{m}, (L: \text{m})$		



No. CNAS L0893

第 163 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
	Bituminous Materials	Surface roughness	1208	$Ra: (0.02 \sim 10) \mu m$	$U_{rel}=4.0\%$		
		Angle		$0^{\circ} \sim 30^{\circ}$	$U=1'$		
		Mass		$(50 \sim 150)g$	$U=0.014g$		
		Temperature		$(0 \sim 50)^{\circ}C$	$U=0.1^{\circ}C$		
73	Rebound Test Hammer	Rebound Value	V. R. of Rebound Test Hammer JJG 817	74、80、83、88	$U_{rel}=3.0\%$		
74	Measuring instruments of paper-thickness	Length	Paper measuring instruments JJG(chuan)(54-75)-1999 JJG (J1) (54-75)	$(0 \sim 2)mm$	$U=1 \mu m$		
75	Standard Gauge Railway	Length	Verification Regulation of Track Gauge for Standard Gauge Railway JJG219	$(1338 \sim 1470)mm$	$U=0.07mm$		
76	Railway Track Gage	Length	Verification Regulation of Calibrator for Railway Track Gage JJG404	$(1410 \sim 1470)mm$	$U=0.01mm$		
77	Anemometer	air speed	Calibration Specification for The pitot static tube anemometer NIMTT(CM) 107, Verification Regulation of portable 3-CUP anemometer JJG431, Verification Regulation of hot ball shaped anemometer JJG(JS)0001	$(0.2 \sim 30)m/s$	$U=0.56m/s+1.4\%V$		
		air flow hood		$(1 \sim 3500)m^3/h$	$U_{rel}=2.0\%$		
78	*Urine Analyzers	pH	Calibration Specification of Urine Analyzers JJF1129	pH: $5.5 \sim 7.5$	$U_{rel}=4\%$		
		PRO		PRO: $(0.01 \sim 3) g/L$	$U_{rel}=4\%$		



No. CNAS L0893

第 164 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		GLU		GLU: (0.1~60) mmol/L	$U_{rel}=2\%$		
		SG		SG: 1.005~1.08	$U=0.003$		
		time		(0~3600) s	$U=0.1$ s		
79	*Blood Cell Analyzers	WBC	Verification Regulation of Blood Cell Analyzers JJG 714	(2~25) ×10 ⁹ /L	$U_{rel}=2.5\%$		
		RBC		(1~8) ×10 ¹² /L	$U_{rel}=2.0\%$		
		HGB		(40~180) g/L	$U_{rel}=2.0\%$		
		PLT		(40~500) ×10 ⁹ /L	$U_{rel}=3.0\%$		
80	*Semiautomatic Clinical Chemistry Analyzers	wavelength	Verification Regulation of Semiautomatic Clinical Chemistry Analyzers JJG 464	(300~800) nm	$U=0.3$ nm		
		absorbance		0.5~1.0	$U=0.002$		
81	*ELISA Analytical Instruments	Wavelength	Verification Regulation of ELISA Analytical Instruments JJG861	(360~800)nm	$U=0.3$ nm		
		Absorbance		0.2~1.5	$U=0.002\sim0.006$		
82	*apparatus of fluidity of cement mortar	Length	Calibration Specification for Flow Table for Determine Cement Mortar Fluidity JJF(chuan) 164	(9~11)mm	$U=0.1$ mm		
		Time		(0~86400)s	$U=0.1$ s		
		Mass		(0~5000)g	0.2g		
83	*cement testing apparatus of the onrmal consistency and setting time	Length	Calibration Specification for Nonmetal Building Materials Plastic Limit Measuring Instruments JJF1090	(0~300)mm	0.03mm		
		Mass		(0~5000)g	0.2g		
		Angle		0° ~320°	$U=0.08^{\circ}$		
Mechanics measuring instrument							
XII Capacity measuring instrument							

No. CNAS L0893

第 165 页 共 181



The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
1	Working Glass Container	Capacity	Working Glass Container JJG 196	(0.001~0.1)ml	$U=0.00006\text{ml}$		
				(0.1~25) ml	$U=0.002\text{ml}$		
				(25~100)ml	$U=0.005\text{ ml}$		
				(100~500)ml	$U=0.03\text{ ml}$		
				(500~2000)ml	$U=0.09\text{ml}$		
				(2000~20000)ml	$U=0.16\text{ ml}$		
2	Standard Capacity Measures(glass)	Capacity	Standard Capacity Measures(glass) JJG 20	(0.001~0.1)ml	$U=0.00006\text{ml}$		
				(0.1~25)ml	$U=0.002\text{ ml}$		
				(25~100)ml	$U=0.005\text{ ml}$		
				(100~500)ml	$U=0.03\text{ ml}$		
				(500~2000)ml	$U=0.09\text{ml}$		
				(2000~20000)ml	$U=0.16\text{ ml}$		
3	Color Tubes	Capacity	Color Tubes JJG 10	(0.001~0.1)ml	$U=0.00006\text{ml}$		
				(0.1~25)ml	$U=0.002\text{ ml}$		
				(25~100)ml	$U=0.005\text{ ml}$		
				(100~500)ml	$U=0.03\text{ ml}$		
				(500~2000)ml	$U=0.09\text{ml}$		
				(2000~20000)ml	$U=0.16\text{ ml}$		
4	Locomotive pipette	Capacity	Locomotive pipette JJG 646	(0.1~20) μL	$U=0.006\text{ }\mu\text{L}$		



No. CNAS L0893

第 166 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
				(20~10000) μ L	$U=0.12 \mu$ L		
				(10~100)ml	$U=0.005$ ml		
				(100~250)ml	$U=0.03$ ml		
5	Standard Glass Hydrometers	Density	Verification Regulation of Standard Glass Hydrometers JJG 86	(650~1500) kg/m^3	$U=0.08 \text{ kg/m}^3$		
		alcohol		q:(0~100)%	$U=q:0.04\%$		
6	Digital Liquid Density Meter-Weighed Method	Density	Verification Regulation of Digital Liquid Density Meter-Weighed Method JJG 999	(650~2000) kg/m^3	$U=(0.08\sim0.20)\text{kg/m}^3$		
7	Laboratory Oscillation-type Liquid Density meters	Density	Verification Regulation of Laboratory Oscillation-type Liquid Density meters JJG 1058	(650~2000) kg/m^3	$U=(0.08\sim0.20)\text{kg/m}^3$		
8	Special capacity	Capacity	Special capacity self-calibration specification NIMTT(CM) 111	(0.0001~0.1)ml	$U=0.00006\text{ml}$		
				(0.1~25)ml	$U=0.002$ ml		
				(25~100)ml	$U=0.005$ ml		
				(100~500)ml	$U=0.03$ ml		
				(500~2000)ml	$U=0.09\text{ml}$		
				(2000~20000)ml	$U=0.16$ ml		
9	Working Glass Hydrometers	Density	Verification Regulation of Working Glass Hydrometers JJG 42	(650~2000) kg/m^3	$U=(0.08\sim0.20)\text{kg/m}^3$		
		alcohol		q:(0~100)%	$U=q: 0.04\%$		
		latex		(1010~1040) kg/m^3	$U=0.10\text{kg/m}^3$		
		Baume		(0~70)Bh	$U=0.04\text{Bh}$		



No. CNAS L0893

第 167 页 共 181

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
		sugar		p:(0~80)%	U=p:0.03%		
10	*Vertical Metal Tank	Capacity (Volume)	Verification Regulation of Vertical Metal Tank Capacity JJG 168	(20~100) m ³	U _{rel} =(2.8×10 ⁻³ ~1.3×10 ⁻³)		
				(100~700) m ³	U _{rel} =(1.3×10 ⁻³ ~9.8×10 ⁻⁴)		
				(700 ~200000) m ³	U _{rel} =(9.8×10 ⁻⁴ ~4.2×10 ⁻⁴)		
11	*Standard Metal Tank	capacity	Verification Regulation of Standard Metal Tank JJG 259	(10~2500)L	U _{rel} =0.86×10 ⁻⁴		
12	*Tank Cars Capacity	capacity	Verification Regulation of Tank Cars Capacity JJG 133	(3~40)m ³	U _{rel} =0.21%		
13	*Horizontal Metallic Tank	Capacity (Volume)	Verification Regulation of the Volume of Horizontal Metallic Tank JJG266	(10 ~200) m ³ ;	U _{rel} =(1.9×10 ⁻³ ~8.7×10 ⁻⁴)		
14	*Spherical metal Tank	Capacity (Volume)	Verification Regulation of Spherical Metal Tank Capacity JJG642	(80~5000) m ³	U _{rel} =(2.6×10 ⁻³ ~6.6×10 ⁻⁴)		
15	Mud Density Meter	Density	Verification Regulation of Mud Density Meter JJG1045	(700~3000)kg/m ³	U=2 kg/m ³		
16	Syringes for Medical Use	Capacity	Syringes for Medical Use JJG 18	(0.25~25)ml	U=0.002ml		
				(25~100)ml	U=0.005 ml		
XIII Flow measuring instrument							
1	*Float Meter	flow	Verification Regulation of Float Meter JJG 257	DN2~DN100, (0.0001~0.3)L/min(Gas)	U _{rel} =0.92%~0.60%		
				DN2~DN100, (0.3~6)L/min(Gas)	U _{rel} =0.60%~0.98%		
				DN2~DN100, (0.2~40)m ³ /h(Gas)	U _{rel} =0.68%~0.54%		
				DN2~DN100, (40~240)m ³ /h(Gas)	U _{rel} =0.54%~0.55%		

No. CNAS L0893

第 168 页 共 181



The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
2	*Ultrasonic Flowmeters	flow	Verification Regulation of Ultrasonic Flowmeters JJG 1030	DN2~DN100, (250~2000)kg/h(Gas)	$U_{rel}=1.5\%$		
				DN2~DN100, (0.004~50)m ³ /h(liquid)	$U_{rel}=0.3\%$		
				DN15~DN100, (0.01~1)m ³ /h(Gas)	$U_{rel}=0.40\% \sim 0.26\%$		
				DN15~DN100, (1~40)m ³ /h(Gas)	$U_{rel}=0.26\% \sim 0.22\%$		
				DN15~DN100, (40~240)m ³ /h(Gas)	$U_{rel}=0.22\% \sim 0.23\%$		
				DN15~DN300, (0.024~0.5)m ³ /h(liquid)	$U_{rel}=0.17\% \sim 0.16\%$		
				DN15~DN300, (0.5~5)m ³ /h(liquid)	$U_{rel}=0.17\% \sim 0.16\%$		
				DN15~DN300, (5~20)m ³ /h(liquid)	$U_{rel}=0.16\%$		
				DN15~DN300, (20~180)m ³ /h(liquid)	$U_{rel}=0.17\%$		
				DN15~DN300, (1~780)m ³ /h(liquid)	$U_{rel}=0.26\% \sim 0.20\%$		
				DN50~DN1600, (2~50000)m ³ /h(liquid) (2~50000)m ³ /h	$U_{rel}=1.2\%$		
3	Water Gas Meters	flow	Calibration Specification for Water Gas Meters JJF 1357	(0.01~10)m ³ /h	$U_{rel}=0.15\%$		
4	Gas Laminar Flow Transducers	flow	Verification Regulation of Gas Laminar Flow Transducers JJG 736	(0.0001~0.3)L/min	$U_{rel}=0.92\% \sim 0.60\%$		
				(0.3~6)L/min	$U_{rel}=0.60\% \sim 0.98\%$		
				(6~60)L/min	$U_{rel}=0.60\% \sim 0.98\%$		



No. CNAS L0893

第 169 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
5	Target Flowmeter	flow	Verification Regulation of Target Flowmeter JJG 461	DN15~DN100,(0.01~5)m ³ /h	$U_{rel}=0.59\% \sim 0.36\%$		
				DN15~DN100,(5~40)m ³ /h	$U_{rel}=0.36\% \sim 0.29\%$		
				DN15~DN100,(40~240)m ³ /h	$U_{rel}=0.29\% \sim 0.25\%$		
6	*Standard Facilities for Liquid Flowrate	volumemass	Verification Regulation of Standard Facilities for Liquid Flowrate JJG 164	(10~25000)L	$U_{rel}=0.05\%$		
				(10~50000)kg	$U_{rel}=0.02\%$		
7	*Vretical Facility for Water meters	volume	Verification Regulation of Vretical Facility for Water meters JJG 1113	(10~10000)L	$U_{rel}=0.1\%$		
8	*Pipe Prover	volume	Verification Regulation of Pipe Prover JJG 209	(10~10000)L	$U_{rel}=0.02\% \sim 0.04\%$		
9	Critical Flow Venturi Nozzle	flow	Verification Regulation of Critical Flow Venturi Nozzle JJG 620	(0.01~10)m ³ /h	$U_{rel}=0.09\%$		
				(0.016~1.6) m ³ /h	$U_{rel}=0.20\%$		
10	*Flow Standard Facilitise by Master Meter Method	flow	Verification Regulation of Flow Standard Facilitise by Master Meter Method JJG 643	(0.01~10)m ³ /h	$U_{rel}=0.23\%$		
				DN2~DN1000, (0.01~50000)m ³ /h	$U_{rel}=0.1\%$		
11	*Weirs and Flumes for Flow Measurement	flow	Verification Regulation of Weirs and Flumes for Flow Measurement JJG 711	(0.1×10 ⁻³ ~1300)m ³ /s	$U_{rel}=(3\sim5)\%$		
12	Vortex Prscession Flowmeters	flow	Verification Regulation of Vortex Precession Flowmeters JJG 1121	DN15~DN300,(1~780)m ³ /h(liquid)	$U_{rel}=0.26\%$		
		flow		DN15~DN100,(0.01~1.2)m ³ /h(Gas)	$U_{rel}=0.40\% \sim 0.25\%$		
				DN15~DN100,(1.2~40)m ³ /h(Gas)	$U_{rel}=0.25\% \sim 0.21\%$		



No. CNAS L0893

第 170 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				DN15~DN100, (40~240)m ³ /h (Gas)	$U_{rel}=0.21\% \sim 0.26\%$		
13	*Flow Integration Meters	flow	Verification Regulation of Flow Integration Meters JJG 1003	(1~999999)Flow engineering unit; frequency: (>0~10)kHz electricity: (>0~20)mA, voltage: (>0~10)V, resistance: (50~1000) Ω	$U_{rel}=0.01\%$ $U_{rel}=0.05\%$; (with pressure temperature feedback)		
14	Diaphragm Gas Meters	flow	Verification Regulation of Diaphragm Gas Meters JJG 577	(0.016~40)m ³ /h (40~160)m ³ /h	$U_{rel}=0.56\% \sim 0.47\%$ $U_{rel}=0.47\% \sim 0.50\%$		
15	*Standard Bell Pipers of Gas Flow	volume	Verification Regulation of Standard Bell Pipers of Gas Flow JJG 165	(10~10000)L	$U_{rel}=0.12\%$		
16	Thermal Mass Gas Flowmeters	flow	Verification Regulation of Thermal Mass Gas Flowmeters JJG 1132	DN2~DN100, (0.01~10)m ³ /h DN2~DN100, (0.0001~0.3)L/min DN2~DN100, (0.3~6)L/min DN2~DN100, (0.2~40)m ³ /h DN2~DN100, (40~240)m ³ /h	$U_{rel}=0.48\%$ $U_{rel}=0.92\% \sim 0.60\%$ $U_{rel}=0.60\% \sim 0.98\%$ $U_{rel}=0.49\% \sim 0.30\%$ $U_{rel}=0.30\% \sim 0.36\%$		
17	*Liquid Positive Displacement Flowmeter	flow	Verification Regulation of Liquid Positive Displacement Flowmeter JJG 667	DN10~DN300, (0.6~20)m ³ /h DN10~DN300, (20~60)m ³ /h	$U_{rel}=0.10\% \sim 0.08\%$ $U_{rel}=0.10\% \sim 0.08\%$		



No. CNAS L0893

第 171 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
				DN10~DN300, (1~780)m ³ /h	$U_{rel}=0.21\% \sim 0.16\%$		
				DN10~DN300, (0.024~0.5)m ³ /h	$U_{rel}=0.076\% \sim 0.070\%$		
				DN10~DN300, (0.5~5)m ³ /h	$U_{rel}=0.076\% \sim 0.072\%$		
				DN10~DN300, (5~20)m ³ /h	$U_{rel}=0.072\% \sim 0.069\%$		
				DN10~DN300, (20~180)m ³ /h	$U_{rel}=0.076\% \sim 0.075\%$		
18	Gas Displacement Meters	flow	Verification Regulation of Gas Displacement Meters JJG 633	DN10~DN100, (0.01~1.2)m ³ /h	$U_{rel}=0.51\% \sim 0.42\%$		
				DN10~DN100, (1.2~40)m ³ /h	$U_{rel}=0.42\% \sim 0.25\%$		
				DN10~DN100, (40~240)m ³ /h	$U_{rel}=0.25\% \sim 0.32\%$		
19	Differential Pressure Flowmeters	flow	Verification Regulation of Differential Pressure Flowmeters JJG 640	DN10~DN100, (0.01~1.2)m ³ /h(Gas)	$U_{rel}=0.85\% \sim 0.64\%$		
				DN10~DN100, (1.2~40)m ³ /h(Gas)	$U_{rel}=0.64\% \sim 0.39\%$		
				DN10~DN100, (40~240)m ³ /h(Gas)	$U_{rel}=0.39\% \sim 0.32\%$		
				DN4~DN300, (1~780)m ³ /h(liquid)	$U_{rel}=0.21\% \sim 0.16\%$		
		Surface roughness		Ra:(0.1~25) μm	$U_{rel}=6\%$		
		Length		d:(12.5~300)mm	$U=(3+L/100) \mu m(L:mm)$		
20	Vortex-shedding Flowmeter	flow	Verification Regulation of Vortex-shedding Flowmeter	DN10~DN100, (0.01~5)m ³ /h(Gas)	$U_{rel}=0.49\% \sim 0.28\%$		



No. CNAS L0893

第 172 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
			JJG 1029	DN10~DN100, (5~40)m ³ /h(Gas)	$U_{rel}=0.28\% \sim 0.22\%$		
				DN10~DN100, (40~240)m ³ /h(Gas)	$U_{rel}=0.22\% \sim 0.29\%$		
				DN15~DN300, (1~780)m ³ /h(liquid)	$U_{rel}=0.21\% \sim 0.16\%$		
21	*Cold Potable Water Meters	flow	Verification Regulation of Cold Potable Water Meters JJG 162	DN15~DN50, (0.004~50)m ³ /h water flow standard device	$U_{rel}=0.58\%$		
				DN15~DN300, (1~780)m ³ /h standard meter method	$U_{rel}=0.60\% \sim 0.47\%$		
				DN50~DN1600, (2~50000)m ³ /h ultrasonic flowmeter	$U_{rel}=1.2\%$		
22	*Electromagnetic Flowmeters	flow	Verification Regulation of Electromagnetic Flowmeters JJG 1033	DN4~DN300, (0.024~0.5)m ³ /h	$U_{rel}=0.13\% \sim 0.12\%$		
				DN4~DN300, (0.5~5)m ³ /h	$U_{rel}=0.13\% \sim 0.12\%$		
				DN4~DN300, (5~20)m ³ /h	$U_{rel}=0.12\%$		
				DN4~DN300, (20~180)m ³ /h	$U_{rel}=0.13\%$		
				DN4~DN300, (1~780)m ³ /h	$U_{rel}=0.24\% \sim 0.17\%$		
				DN50~DN1600, (2~50000)m ³ /h	$U_{rel}=1.2\%$		
23	Hot meters	flow	Verification Regulation of Hot water meters JJG 686	DN15~DN50, (0.004~50)m ³ /h DN(15~50)mm	$U_{rel}=0.62\%$		



No. CNAS L0893

第 173 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
24	*Gas Flow Calibration Facility by Means of Critical Flow Venturi Nozzles	flow	Calibration Specification for Gas Flow Calibration Facility by Means of Critical Flow Venturi Nozzles JJF 1240	(0.01~10)m ³ /h DN2~DN600, (0.01~50000)m ³ /h	$U_{rel}=0.23\%$ $U_{rel}=0.2\%$		
25	*Goriolis Mass Flow meters	flow	Verification Regulation of Goriolis Mass Flow meters JJG 1038	DN2~DN200, (0.024~0.5)t/h	$U_{rel}=0.051\%\sim0.042\%$		
		flow		DN2~DN200, (0.5~5)t/h	$U_{rel}=0.051\%\sim0.046\%$		
		flow		DN2~DN200, (5~20)t/h	$U_{rel}=0.041\%\sim0.046\%$		
		flow		DN2~DN200, (20~180)t/h	$U_{rel}=0.051\%\sim0.050\%$		
		flow		DN2~DN200, (0.0001~6) L/min, (0.00012~7.2) g/min	$U_{rel}=0.6\%\sim0.92\%$		
		flow		DN2~DN200, (0.005~50) L/min, (0.006~60) g/min	$U_{rel}=0.32\%$		
		flow		DN2~DN200, (0.2~60) L/min, (0.22~66) g/min	$U_{rel}=0.15\%\sim0.3\%$		
		flow		DN2~DN200, (0.01~1.2)m ³ /h, (0.011~1.3) kg/h	$U_{rel}=0.51\%\sim0.42\%$		
		flow		DN2~DN200, (0.2~40)m ³ /h, (1.3~44) kg/h	$U_{rel}=0.42\%\sim0.25\%$		
		flow		DN2~DN200, (40~240)m ³ /h, (44~260) kg/h	$U_{rel}=0.25\%\sim0.32\%$		



No. CNAS L0893

第 174 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
		flow		DN2~DN200, (0.5~1000)m ³ /h, (0.55~1100) kg/h	$U_{rel}=0.32\%$		
26	*Verification Regulation of Soap Film Flow meter	flow	Verification Regulation of Soap Film Flow meter JJG 586	(0.0001~6) L/min	$U_{rel}=0.6\%\sim 0.92\%$		
		flow		(0.005~50) L/min	$U_{rel}=0.32\%$		
		flow		(0.2~60) L/min	$U_{rel}=0.15\%\sim 0.3\%$		
27	*Turbine Flowmeter	flow	Verification Regulation of Turbine Flowmeter JJG 1037	DN15~DN100, (0.01~5)m ³ /h(Gas)	$U_{rel}=0.43\%\sim 0.22\%$		
				DN15~DN100, (5~40)m ³ /h(Gas)	$U_{rel}=0.22\%\sim 0.21\%$		
				DN15~DN100, (40~240)m ³ /h(Gas)	$U_{rel}=0.21\%\sim 0.26\%$		
				DN4~DN300, (1~780)m ³ /h(liquid)	$U_{rel}=0.21\%\sim 0.16\%$		
				DN4~DN300, (0.024~0.5)m ³ /h(liquid)	$U_{rel}=0.13\%\sim 0.12\%$		
				DN4~DN300, (0.5~5)m ³ /h(liquid)	$U_{rel}=0.13\%\sim 0.12\%$		
				DN4~DN300, (5~20)m ³ /h(liquid)	$U_{rel}=0.12\%$		
				DN4~DN300, (20~180)m ³ /h(liquid)	$U_{rel}=0.13\%$		
28	*Compressed Natural Gas Dispensers	Flow	Verification Regulation of Compressed Natural Gas Dispensers JJG 996	(1~80) kg/min	$U_{rel}=0.22\%$		
29	*Liquefied Natural Gas Dispensers	Flow	Verification Regulation of Liquefied Natural Gas Dispensers JJG 1114	(1~80) kg/min	$U_{rel}=0.32\%$		



No. CNAS L0893

第 175 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
30	*Compressed Hydrogen Dispensers	Flow	Calibration Specification of Compressed Hydrogen Dispensers NIMTT(CM) 013	(0.1~10) kg/min	$U_{rel}=0.66\%$		
31	Heat Water Meters	flow	Verification Regulation of Heat Meters JJG 225	(0.004~50)m ³ /h, DN(15~50)	$U_{rel}=0.7\%$		
32	Dry piston flow calibrator	flow	Calibration Specification of Dry piston flow calibrator NIMTT(CM) 057	(0.005~50) L/min	$U_{rel}=0.28\% \sim 0.32\%$		
		flow		(0.2~60) L/min	$U_{rel}=0.2\% \sim 0.3\%$		
33	*Electromagnetic Flowmeters	flow	Online Calibration Specification for Liquid Flowmeter JJF (J1) 159	DN50~DN3000, (2~50000) m ³ /h	$U_{rel}=1.2\%$		
34	*p.V.T.t technique	volume	Gas Flow Calibration to p.V.T.t technique JJG 619	(10~30000) L	$U_{rel}=0.03\%$		
		flow		(0.01~4000) kg/h	$U_{rel}=0.05\%$		
35	*measuring systems for natural gas	flow	Calibration Specification of Measuring Systems for Natural Gas NIMTT(CM) 165	(0~9999999) 流量工程单位	$U_{rel}=0.7 \sim 1\%$		
36	Standand Leak	flow	Calibration Specification for Reference Leaks by Soap Film Flowmeter JJF 1627	(0.0001~0.3)L/min	$U_{rel}=(0.92 \sim 0.60)\%$		
				(>0.3~6)L/min	$U_{rel}=(0.60 \sim 0.98)\%$		
37	*Gasoline Vapor Recovery Detectors	flow	Calibration Specification for Gasoline Vapor Recovery Detectors JJF1948	(18~38) L/min	$U_{rel}=0.2 \sim 0.5\%$		
		pressure		(-5.0~5.0) kPa	$U=6\text{Pa}$		
38	*Calibration Specification for Door and window performance detection device	flow	Calibration Specification for Door and window performance detection device NIMTT(CM) 164	(10~600) m ³ /h	$U_{rel}=0.52\%$		
		flow		(100~1000) L/h	$U_{rel}=0.48\%$		
		pressure		(-5000~5000) Pa	$U=6\text{Pa}$		



No. CNAS L0893

第 176 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (<i>k</i> =2)	Note	Effective Date
		Displacement		(1~10) mm	<i>U</i> =7 μ m		
39	*Syringe Pump and Infusion Pump	flow	Calibration Specification for Syringe Pumps and Infusion Pumps JJF 1259	(5~20)mL/h	<i>U</i> = 0.84 mL/h		
				(>20~200) mL/h	<i>U</i> = 0.27 mL/h		
				(>200~1000) mL/h	<i>U</i> =12 mL/h		
		occlusion pressure		(40~106.7) kPa	<i>U</i> = 4.9 kPa		
XIV、Pressure measuring instrument							
1	*Piezoresistive Vacuum Gauge	Pressure	Verification Regulation of Piezoresistive Vacuum Gauge JJG 932	(1×10 ² ~1×10 ⁵)Pa	<i>U</i> _{rel} =0.9%~0.6%		
2	Pressure Transducer	Pressure	V.R. of Pressure Transducer(static) JJG 860	(-0.1~100)MPa	<i>U</i> =0.016%FS		
				(100~250)MPa	<i>U</i> _{rel} =0.027%		
3	Compensated Micro-manometer	Pressure	V.R.of Compensated Micro- manometer JJG 158	-2.5kPa~2.5kPa	<i>U</i> =0.018%FS		
4	Tilting Tube Micro-manometers	Pressure	V.R.of Tilting Tube Micro- manometers JJG 172	(0~2000)Pa	<i>U</i> =0.15%FS		
5	Aneroid Barometer & Aneroid Barograph	Pressure	V.R.of Aneroid Barometer & Aneroid Barograph JJG 272	(500~1060)hPa	<i>U</i> =0.85hPa		
6	Piston Gauge	Pressure	V.R.of Piston Gauge JJG 59	(0.02~250)MPa	<i>U</i> _{rel} =0.012%		
7	Pneumatic Piston Gauge	Pressure	V.R.of Pneumatic Piston Gauge JJG 1086	(0.0015~10)MPa	<i>U</i> _{rel} =0.012%		
8	Precise Pressure Gauge	Pressure	V.R.of Elastic Element Precise Pressure Gauges and Vacuum Gauges JJG 49	(-0.1~0)MPa	<i>U</i> =0.11%FS		
				(0~250)MPa	<i>U</i> =0.14%FS		



在线扫码获取验证

No. CNAS L0893

第 177 页 共 181

The scope of the accreditation in Chinese remains the definitive version.


№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
9	Pressure Gauge	Pressure	V.R.of Elastic Element Precise Pressure Gauges, Pressure-Vacuum Gauges and Vacuum Gauges for General Use JJG 52	$(-0.1 \sim 250)$ MPa	$U=0.6\%FS$		
10	Sphygmomanometer	Pressure	V.R.of Sphygmomanometer JJG 270	$(0 \sim 40)$ kPa	$U=0.6\%FS$		
11	Record Pressure Gauges	Pressure	V.R.of Record Pressure Gauges、Pressure Vacuum Gauges and Vacuum Gauges JJG 926	$(-0.1 \sim 250)$ MPa	$U=0.6\%FS$		
12	Piston Pressure Vacuum Gauge	Pressure	V.R.of Piston Pressure Vacuum Gauges JJG 236	$-0.1\text{MPa} \sim -10\text{kPa}$	$U_{rel}=0.015\%$		
				$10\text{kPa} \sim 0.6\text{MPa}$	$U_{rel}=0.015\%$		
13	Piston Pressure- Vacuum Gauge with Equilibrium Liquid Column	Pressure	V.R.of Piston Pressure- Vacuum Gauge with Equilibrium Liquid Column JJG 51	$-0.1\text{MPa} \sim 0.4\text{MPa}$	$U=0.015\%FS$		
14	Standard Dual Piston Pressure Vacuum Gauge	Pressure	V.R.of Standard Dual Piston Pressure Vacuum Gauge JJG 159	$-0.1\text{MPa} \sim 1\text{MPa}$	$U=0.015\%FS$		
15	Pressure Controllers	Pressure	V.R.of Pressure Controllers JJG 544	$(-0.1 \sim 250)$ MPa	$U=0.8\%FS$		
16	Precision Liquid Manometer for Cistern	Pressure	V.R.of Precision Liquid Manometer for Cistern and U-tube JJG 241	$(-40 \sim 40)$ kPa	$U=0.4\%FS$		



No. CNAS L0893

第 178 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
17	Digital Measuring Device for Optical Interference Methane Detector	Pressure	V.R.of Digital Measuring Device for Optical Interference Methane Detector JJG 1040	(0~100)kPa	$U=0.033\%FS$		
18	Pressure SF6 Gas Density Monitors	Pressure	V.R.of Pressure Type SF6 Gas Density Monitors JJG 1073	(-0.1~0.9)MPa	$U=0.6\%FS$		
19	*Thermal Conductivity Vacuum Gauge	Pressure	Calibration Specification of Working Thermal Conductivity Vacuum Gauge JJF 1050	$(1 \times 10^{-1} \sim 3 \times 10^3)Pa$	$U_{rel}=5.3\% \sim 3.9\%$		
				$(3 \times 10^3 \sim 1 \times 10^5)Pa$	$U_{rel}=3.9\% \sim 4.9\%$		
20	Liquid Manometers for Working	Pressure	V. R. of Liquid Manometers for Working JJG 540	(0~2.5)kPa	$U=0.27\%FS$		
				(2.5~25) kPa	$U_{rel}=0.55\%$		
21	Tyre Pressure Gauges	Pressure	V.R.of Tyre Pressure Gauges JJG 927	(0~6)MPa	$U=0.6\%FS$		
22	Ball Pneumatic Dead Weight Testers	Pressure	V.R.of Ball Pneumatic Dead Weight Testers JJG 942	2kPa~10MPa	$U_{rel}=0.015\%$		
23	Calibration Specification for Pressure Regulators with Bourdon Tube Pressure Gauge	Pressure	Calibration Specification for Pressure Regulators with Bourdon Tube Pressure Gauge JJF 1328	(0~25)Mpa	$U=0.6\%FS$		
24	Liquid Level Gauges	Pressure	V.R of Liquid Level Gauges JJG 971	(0~10)MPa	$U=0.023\%FS$		
25	Digital Pressure Gauges	Pressure	V. R. of Digital Pressure Gauges JJG 875	(-0.1~100)MPa	$U=0.008\%FS$		
				(100~250) MPa	$U_{rel}=0.021\%$		



No. CNAS L0893

第 179 页 共 181

The scope of the accreditation in Chinese remains the definitive version.

№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
26	Digital Barometers	Pressure	V.R. of Digital Barometers JJG 1084	(0~130)kPa abs	$U=0.07\text{hPa}$		
27	Pressure Transmitter	Pressure	V. R. of Pressure Transmitter JJG 882	(-0.1~100)MPa	$U=0.016\%\text{FS}$		
28	*Ionization Vacuum Gauge	Pressure	Calibration Specification of Ionization Vacuum Gauge JJF 1062	(100~250) MPa	$U_{\text{rel}}=0.027\%$		
				$(3 \times 10^{-7} \sim 1)\text{Pa}$	$U_{\text{rel}}=16\% \sim 1.5\%$		
				$(>1 \sim 20)\text{Pa}$	$U_{\text{rel}}=4.2\%$		
29	Mass Spectrometer for Partial Pressure	Pressure	Calibration Specification of Mass Spectrometer for Partial Pressure in the Pressure Range of $(1 \times 10^{-9} \sim 1 \times 10^{-1})\text{Pa}$ JJF (军工) 152	$(1 \times 10^{-6} \sim 1 \times 10^{-1})\text{Pa}$	$U_{\text{rel}}=7.4\% \sim 6.2\%$		
30	*Capacitance Diaphragm Vacuum Gauges	Pressure	Calibration Specification for Capacitance Diaphragm Vacuum Gauges JJF 1503	$(1 \times 10^{-1} \sim 1 \times 10^5)\text{Pa}$	$U_{\text{rel}}=2.6\% \sim 0.02\%$	MPE: $\pm 0.5\%$ for field calibration only	
31	*Reference Leaks for Gases	leak rate	Non Destructive Testing - Leak Testing - Calibration of Reference Leaks for Gases ISO 20486, Calibration Specification for Vacuum Helium Leaks JJF 1833	$(1 \times 10^{-10} \sim <1 \times 10^{-3})\text{Pa} \cdot \text{m}^3/\text{s}$	$U_{\text{rel}}=7.6\% \sim 0.88\%$	Calibration by comparison and calibration by pressure change in a known volume only for ISO 20486	



№	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty ($k=2$)	Note	Effective Date
32	Pressure Reference Leaks	leak rate	Calibration Specification for Reference Leaks by Soap Film Flowmeter JJF 1627, Calibration Specification for Pressure Reference Leaks NIMTT(CM) 012	$(5 \times 10^{-7} \sim 50) \text{ Pa} \cdot \text{m}^3/\text{s}$	$U_{\text{rel}}=6.2\% \sim 0.7\%$		
33	*Mass Spectrometer Leak Detector	leak rate	Calibration Specification of Helium Mass Spectrometer Leak Detector JJF (军工) 186, Calibration Specification of Helium Mass Spectrometer Leak Detector NIMTT (CM) 119	$(1 \times 10^{-10} \sim <1 \times 10^{-5}) \text{ Pa} \cdot \text{m}^3/\text{s}$	$U_{\text{rel}}=14\% \sim 12\%$		
34	*Halogen Leak Detector	leak rate	Calibration Specification for Halogen Leak detector	$(1 \times 10^{-6} \sim 1 \times 10^{-5}) \text{ Pa} \cdot \text{m}^3/\text{s}$	$U_{\text{rel}}=24\% \sim 13\%$		
		Time	JJF1964	$(0 \sim 10) \text{ s}$	$U=1.1 \text{ s}$		
35	Impression Tonometers	mass	V.R.of Impression Tonometers JJG 574	$(0 \sim 100) \text{ g}$	$U=0.012 \text{ g}$		
		length		$(0 \sim 5) \text{ mm}$	$U=0.013 \text{ mm}$		



No. CNAS L0893

第 181 页 共 181

The scope of the accreditation in Chinese remains the definitive version.